A-A-53889
12 June 1989

COMMERCIAL ITEM DESCRIPTION

FLASKS, DENTURE, Upper and Lower, Ejector Type

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-F-36284A, which is cancelled.

Salient characteristics:

This commercial item description covers ejector type flasks suitable for processing upper and lower dentures.

Types. Ejector type flasks shall be of the following types, as specified (See notes).

Type I. Upper

Type II. Lower

Material.

Brass. Brass for forging the flasks shall conform to copper alloy number 377 of ASTM Designation B124 for composition. Composition shall be tested as specified herein.

Steel. Steel for the guide pins shall conform to class 3, austenitic stainless steel, type 303 of ASTM Designation F899 for composition. Composition shall be tested as specifed herein.

Design and dimensions.

Type I, upper flask. The design and dimensions of Type I, upper flask, shall be in accordance with Defense Personnel Support Center (DPSC) Drawing 20485. Type I flask shall be suitable for processing upper dentures.

Type II, lower flask. The design and dimensions of Type II, lower flask, shall be in accordance with Defense Personnel Support Center (DPSC) Drawing 20125. Type II flask shall be suitable for processing lower dentures.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Construction. The denture flasks shall consist of a top, middle and a bottom section. The middle section shall contain two tapered guide pins to maintain registration between the sections. The bottom section shall contain a removable disk.

Finish. Flasks shall have a smooth finish. All matching surfaces shall be machined to insure a close fit and preclude separations between the three sections of the flask. Steel for the guide pins shall be ground and annealed.

Identification marking. Each ejector type flask shall be serialized, and the top, middle and bottom section of the flask shall be permanently stamped with the serial number as shown on the drawings. In addition, as part of the forging operation, the name or registered trademark of the manufacturer and the word "Upper" (for type I flask) or "Lower" (for type II flask) shall be permanently and legibly marked on the top section of each flask.

Workmanship. The flasks shall be free from burrs, scratches and other defects which detract from their appearance or may impair their serviceability.

Unit. Each (EA). One flask, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.5 (percent defective) and an inspection level of II, except for examination of dimensions which shall be conducted with an AQL of 2.0 and an inspection level of S-2. Examination shall include, but shall not be limited to the defects listed below. The unit of product for examination purposes shall be one flask.

Defects.

Type not as specified.
Finish not as specified.
Any component of flask missing or not of design as shown in applicable drawing.
Cracks, breaks, or other workmanship defects seriously affecting proper fit and function of flasks present.
Dimensions not as specified.
Name or trademark of manufacturer and the word "Upper" (for type I) or "Lower" (for type II) not present on top section of flask.
Top, middle, and bottom sections of flask not stamped with serial number.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified below. The unit of product for test purposes shall be one flask.

Testing Criteria.

Characteristic	Inspection level	AQL
Chemical analysis	**	*

*Acceptance number shall be zero for all sample sizes.

**A sufficient quantity of each lot of raw steel or brass used in the fabrication of the instruments shall be selected for chemical analysis.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Chemical analysis. In the absence of certificates of quality, the chemical analysis of the instruments shall be determined in accordance with method 111 or 112 of FED-STD-151. In the event of a conflict, method 111 shall govern.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurances practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquistion Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	Identificatiom Item	Type
6520-00-531-0500	FLASK, DENTURE, Upper, Ejector Type	I
6520-00-559-9935	FLASK, DENTURE, Lower, Ejector Type	II

Ordering data. Intermediate/exterior package quantities and labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2222

LOCATION: NCR Disk No. 7

05421

☆U.S. GOVERNMENT PRINTING OFFICE: 1989 - 604-033/03323

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-BSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53889	FLASKS, DENTURE, Upper and Lower, Ejector Type

3a. NAME OF SUBMITTING ORGANIZATION

b. ADDRESS *(Street, City, State, ZIP Code)*

4. TYPE OF ORGANIZATION *(Mark one)*

- ☐ VENDOR
- ☐ USER
- ☐ MANUFACTURER
- ☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.