A-A-53891
7 April 1989

COMMERCIAL ITEM DESCRIPTION

TOOL KIT, OPTICIAN

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a tool kit for use by an optician during spectacle frame adjustment and repair.

Salient characteristics:

Materials. All tools provided in the tool kit shall be made of corrosion resistant materials. Plastic components shall be chemically inert.

Case. A zippered, vinyl case designed to hold the tools shall be supplied as a part of this kit. The case shall contain straps and/or pockets used to hold each tool securely in place during transportation. Small files and tweezers may share the same strap or pocket.

Tools. The tool kit shall consist of one each of the following tools:

Lens Aligner - With soft rubber circular jaws, recess centers, 7-1/2 inches overall length.

Combination Post & Endpiece Plier - Five (5) inches overall length.

Zyl Temple Angling Plier - Five (5) inches overall length.

Snipe Nose Plier - Five (5) inches overall length.

Wide Jaw Angling Plier - 5-1/4 inches overall length.

End Cutting Nipper - 4-1/2 inches overall length.

Hollow Snipe Plier - Five (5) inches overall length.

AMSC N/A FSC 6540

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Rubber Jaw Plier - 4-3/4 inches overall length, spring opened.

Eyewire Gauge - All metal caliper type, with inch and mm scales

Tweezer - Five (5) inches with fine points and bevelled edges.

Optician Screwdriver - 4-1/2 inches double end blades, one end at 1.55 mm, one end at 2.2 mm.

Automatic Centering Punch - Spring loaded, 5 inches length, adjustable impact force, finish with cup point.

PD Ruler - Stainless steel, 6 inches long, 15/32 inch wide, with inch and millimeter scales.

Pin Vise/Tap Holder - 4 inches long, 5/16 inch diameter, 3/4 inch swivel for holding drills, screwdriver blades.

Pick-up screwdriver - 3-1/4 inch long, with spring tension jaws, 2 mm replaceable blade.

Hex wrench - 3-1/4 inches long, with swivel head and spring ejector, replaceable tip to work on .0885 inch nuts.

Zyl File - Eight (8) inches long, coarse cut, tapered at both ends.

Knife File - 4 inches long, #2 cut, wide back tapering to knife edge.

Narrow Pillar File - 4 inches long, 17/64 inch wide x 7/64 inch thick.

Screw Slotting File - 3 inches long, #4 cut.

Rat Tail File - 3 inches long, #2 cut, tapered to a fine point.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Screws Finishing File - 6 inches long, medium #2 cut for filing zyl screw without marring hinge.

Vertical Hinge Angling Plier - 5-1/4 inches overall length, spring opened, fiber lining on one jaw.

Workmanship. The tool kit shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Set (SE). One set, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

6540-01-145-8776

Ordering data. (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

AGENT:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6540-1196

Location: Disk # 10
CID-TOOL-KIT

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.