A-A-53892
28 April 1989

COMMERCIAL ITEM DESCRIPTION

STIMULATOR, FACIAL NERVE

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a battery powered nerve stimulator suitable for use in the evaluation and prognosis of facial nerve involvement.

Salient characteristics.

The unit shall be a battery powered facial nerve stimulator. The unit shall provide square wave stimulation for measurement of facial nerve response and triangular wave stimulation for measurement of facial muscle response.

Current intensity shall be adjustable from 0 to 10 milliamperes. The unit shall deliver constant current at any given setting by automatically compensating for normal variations in patient circuit resistance. A warning light shall be provided to indicate if the electrical resistance through the patient exceeds the permissible limit for the passage of the selected current.

The unit shall be provided with and operate from an internal rechargeable battery. The battery shall recharge when the unit is connected to a 115 volt, 60 Hz power source. A fully charged battery shall provide for 8 hours of use.

The unit shall be provided with the following accessories:

Probe handle and harness

Ball end probe

Needle end probe

Blunt probe

12 No. 3 disposable return electrodes

AMSC/NA FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

The unit shall be provided with a carrying case. The case shall enclose the stimulator unit and all accessories. The unit shall be provided with support documentation. As a minimum, the documentation shall include an operators manual and a maintenance manual. The maintenance manual shall include calibration instructions, parts lists and wiring diagrams.

Workmanship. The facial nerve stimulator shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA) One stimulator, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defines as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one facial nerve stimulator. Sampling shall be conducted at inspection level S-2 with an AQL of 1.0.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Testing shall include but not be limited to the following.

With the unit connected to an appropriate power source, the unit shall be checked for proper operation. Verify the unit is capable of producing square and triangular wave stimulation. Also, verify current is adjustable from 0 to 10 milliamperes.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standard, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-075-6630	Stimulator, Facial Nerve Battery Powered, Transistorized

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4472

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.