A-A-53893
14 April 1989

COMMERCIAL ITEM DESCRIPTION

POWER SUPPLY (120 Volt, 50/60 Hz, AC)

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers a DC rectifier adapter for alternate operation of a battery powered microhematocrit centrifuge from an AC supply line.

Salient characteristics.

Shall be an electrical power pack for alternate use with a microhematocrit centrifuge, that can also be internally powered by six (6) size C manganese-alkali battery cells.

Shall include a rectifier circuit for converting a 120 volt, 50/60 Hz, AC line voltage input to 6.5 volts DC output.

Shall be rated for at least 3 watts operation, and shall be capable of providing 6.5 volts and 400 milliampere DC outputs when connected to a microhematocrit centrifuge, designed to have an operating speed of 11,500 rpm and a programmed centrifugation time of at least 3 minutes and 20 seconds.

Shall be supplied with an AC input 3-wire line cord that terminates with a non-detachable three-blade male connector for use in receptacles conforming to ANSI C73-11.

Shall be supplied with an output cord with a connector that is suitable for direct connection to a microhematocrit centrifuge.

Workmanship. The power supply shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One power supply, as specified, constitutes one unit.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and testing. Inspection and testing shall be conducted on each power supply to demonstrate its capability for proper operation.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6640-01-070-1500

The following National Stock Numbers covers battery operated centrifuges intended for use with subject power supply:

6640-01-068-9612 (CENTRIFUGE, LABORATORY, Microhematocrit, Battery Powered, 9 volt, DC)

6640-01-205-2422 (CENTRIFUGE, LABORATORY, Microhematocrit, Battery Powered, 9 volt, DC)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6640-1559

Location: Disk No. 16
CID/POWER/SUPPLY/S31

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.