A-A-53895
30 May 1989

COMMERCIAL ITEM DESCRIPTION

SURGICAL PACK, DISPOSABLE

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a sterile, disposable, surgical drape pack fabricated of nonwoven material for creating a sterile field specifically for abdominal surgical procedures.

Salient characteristics.

The pack shall consist of at least the following components:

a. One overwrap/tablecover, 44 by 90 inches. (minimum)
b. One Mayo standard cover with pad, 23 by 55 inches.
c. Two surgical gowns, large or extra-large, with hand towels.
d. Four utility drapes with tape, 14 by 26 inches.
e. One laparotomy drape sheet, 72 by 120 inches (minimum), with bactericidal treatment (optional) fabric reinforcement, an instrument pad, 3 by 12 inch (nominal) fenestration, cord holders, and arm-board covers.
f. One suture bag.

The overwrap/tablecover shall be fan-folded for ease in opening on the surgical pack table. The overwrap shall be impervious to fluids and moisture. The Mayo stand and tray cover shall be made of a plastic material at least 2.75 mils thick and be impervious to fluids. The cover shall be fanfolded with a six inch cuff on the distal end for sterile draping of the Mayo stand. The pad shall be large enough to cover the surface of the Mayo tray.

The gowns shall be large or extra-large size.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

All gowns shall:

- Be soft, comfortable, pliable, quiet, and memory free.
- Have high breathe-ability, facilitating comfort for wearing long periods of time.
- Be fluid repellent.
- Have stockinette cuffs that shall conform to wrists.
- Be wrap-around type to completely cover the back of the individual gowned.
- Have a tie closure that shall not rip or separate when tying. Shall be designed to minimize the possibility of contamination when tying.
- One gown shall be in the overwrap for the scrub nurse.

The utility drapes shall be water repellent.

The T-shaped laparotomy drape shall be 72 by 120 inches (minimum) with nominal 3 by 12 inch fenestration. The sheet shall have a bactericidal treatment (optional) fabric reinforcement around the fenestration. The reinforcement area shall be impervious to fluids yet highly absorbent, eliminating strike-through while preventing pooling of fluids. The sheet shall contain an instrument pad as part of the drape. The sheet shall be cuffed and fan-folded to permit aseptic draping of the operative area.

The pack shall contain a suture bag made of a nonwoven fabric or plastic material.

All nonwoven materials shall be fabricated of a soft material with minimal memory for good drape-ability. They shall resist tearing, pilling, and linting. If the pack is opened but not used, it shall be capable of being sterilized by steam at 275°F or by ethylene oxide.

The pack shall be sterile and individually wrapped in a suitable tear resistant, waterproof, heat sealed package which shall allow aseptic handling and maintain sterility of the contents until opened or damaged. Items shall be packed in sequence of use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

All pack components shall be clearly listed on the exterior wrapper which shall also serve as a dust cover.

Workmanship. The surgical packs shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing four or five sterile surgical laparotomy packs, as specified, constitutes one unit. The surgical pack, shall be individually wrapped in a suitable tear resistant waterproof, heat sealed pack, capable of maintaining sterility of contents unless package is damaged or opened. Items shall be packed in sequence of use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one surgical pack as specified in this description.

Defects

Any component part missing, broken, torn or otherwise not useable.

Pack not free of tears, cuts, holes, or perforations.

Pack not sterilized as required or sterility compromized.

Pack not free of sand, grease, lint, dust, or foreign material.

Uneven or ragged edges.

Raised areas or lumps on front or back.

Any objectional odor.

Sampling for tests. Sampling for tests (dimensional) shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one surgical pack.

First article. Where a first article sample is required, all tests shall be performed on that unit.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Number is covered by this document

6530-00-286-6732

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

A-A-53895

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

Project No. 6530-2042

Location:
CID/SURGICAL/PACK/H15-33

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge (*DO NOT STAPLE*), and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

Fold along this line

Fold along this line

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
PHILADELPHIA, PENNSYLVANIA 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53895	SURGICAL PACK, DISPOSABLE

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION (Mark one)
	☐ VENDOR
	☐ USER
b ADDRESS (Street, City, State, ZIP Code)	☐ MANUFACTURER
	☐ OTHER (Specify)

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER (Last, First, MI – Optional)	b WORK TELEPHONE NUMBER (Include Area Code) – Optional
c MAILING ADDRESS (Street, City, State, ZIP Code) – Optional	8 DATE OF SUBMISSION (YYMMDD)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil - Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.