A-A-53900
2 May 1989

COMMERCIAL ITEM DESCRIPTION

CURETTE, ENDOMETRIAL BIOPSY, Curved, 11-3/4 Inch, Disposable, 12s

The General Services Administration has authorized the use of this commercial item description as a replacement for Military Specification MIL-C-37026A, which is cancelled.

This commercial item description covers a sterile, disposable aspirating device which is used with a vacuum source to collect specimens for detecting endometrial cancer.

Salient characteristics:

General. The endometrial biopsy curette shall consist of an aspirating cannula with attached cap, plastic protector for the cannula, container, filter, captive plug, dispatch lid, plastic bag, and tie tape. The curette shall be supplied with instruction for use. The curette shall be capable of operating from a central vacuum source or vacuum pump at 160 to 260 millimeters (mm) Hg gage pressure.

Material. The slotted filter shall be fabricated from low density polyethylene plastic. The container, attached cannula cup, captive plug, dispatch lid and plastic protector for the cannula tip shall be fabricated from polypropylene plastic. The cannula tube shall be fabricated from Class 3, Type 304 austenitic stainless steel in accordance with ASTM Designation F899 with a finished Rockwell Hardness B Scale (HRB) of 75 minimum, when tested in accordance with ASTM Designation E78 or E92 whichever is applicable. Conversion of hardness values shall be tested in accordance with ASTM Designation E140.

Style, design, and dimensions. The endometrial biopsy curette shall be in accordance with DPSC Drawing No. 24382.

Sterility. The sterility assurance level of the curette shall be 0.0001 when tested for sterility in accordance with USP procedures.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Traceability. Each quantity of product sterilized simultaneously in the same sterilizer shall be termed a sterility lot. Each sterility lot shall be traceable to a particular manufacturing lot.

Corrosion resistance. The stainless steel cannula shall show no signs of corrosion when boil tested in accordance with ASTM Designation F1089.

Certificate of sterility. Sterility testing shall be done by the contractor or by a qualified independent laboratory. Foreign protocols are not permitted. The contractor's sterility certificate verifying sterility testing shall be furnished as objective quality evidence and will consist of a document written by the facility performing the test.

Workmanship. The biopsy curette shall be free from defects which detract from its appearance or impair its serviceability.

Visual and dimensional examination. Curettes shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements.

Unit. Package (PG). One package containing twelve individually packaged sterile curettes, as specified, constitutes one unit. The individual package shall be a sealed peel open type container or pouch capable of maintaining sterility of contents unless package is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:26Z
Check the source to verify that this is the currentversion before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. The curettes shall be examined for compliance with the requirements in this specification. Examination shall be conducted in accordance with MIL-STD-105, at level II and AQL 1.0 (percent defective).

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table I as specified herein. Dimensional testing shall be S-1 and an AQL of 1.0%, sterility testing shall be in accordance with the USP, and boil testing shall be level S-2 and an AQL of 2.5%. The unit of product for test purposes shall be one curette.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conform to the producer's own drawings, specifications, standards and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-149-0332	Curette, Endometrial Biopsy, Curved, 11-3/4 Inch, Disposable, 12s

Ordering data (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4448

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53900	CURETTE, ENDOMETRIAL BIOPSY, Curved, 11-3/4 Inch, Disposable, 12s

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
3b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(FOLD ALONG THIS LINE)

(TO DETACH THIS FORM)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)* and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.