A-A-53907
19 June 1989

COMMERCIAL ITEM DESCRIPTION

APPLICATOR SET, PIN RETENTION, DENTAL, Spiral, Cement

The General Services Administration has authorized the use of this Commercial Item Description.

Salient characteristics:

This commercial item description covers an applicator set suitable for depositing dental cement onto the walls of amalgam reinforcement pin holes in teeth structures, without the entrapment of air.

Material. The shank shall be fabricated of austenitic stainless steel in accordance with class 3, type 302, 303 or 304 of ASTM F899. The operative end shall be fabricated from 0.030 inch diameter, tapered, high carbon steel piano wire. The wire shall be in accordance with ASTM A228M (Steel Wire, Music Spring Quality), except the minimum tensile strength shall be 325,000 pounds per square inch (psi), prior to forming.

Style, design and dimensions. The applicators shall be spiral filling instruments of the Lentulo design, and shall consist of an operative end securely inserted into the shank. Applicators shall be in accordance with Defense Personnel Support Center (DPSC) Drawing 24113 for style, design and dimensions, and shall be furnished as a set of one each of the four sizes shown in the drawing: long large, long fine, short large, and short fine. The large applicators shall be capable of depositing cement upon the walls of amalgam reinforcement pin holes 0.027 inch in diameter; the fine size applicators shall be used for depositing cement upon the walls of pin holes 0.024 inch in diameter during pin-retention techniques, without the entrapment of air.

Identification marking. The spiral applicators shall be permanently and legibly marked with the name or registered trademark of the contractor or manufacturer. Spiral applicators of foreign manufacture other than (U.S.) shall also be marked with the name of the country of origin.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Workmanship. The applicators shall be free from kinks, cracks, or other surface imperfections which detract from their appearance or may impair their function, intended use or serviceability.

Unit. Set (SE). One package containing one set of four applicators, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. Dimensions shall be sampled for examination at inspection level S-2, with an AQL of 1.5 (percent defective). All other sampling for examination shall be conducted at inspection level II, with an AQL of 1.0 (percent defective).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Examination. The applicators shall be examined to determine compliance with all requirements contained in this commercial item description. Any applicator in the sample containing one or more defects shall be rejected, and if the number of defective applicators in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected. Applicators shall be examined for defects including, but not limited to, those listed below. The unit of product for examination purposes shall be one set.

Defects

Identification marking not as specified.
Finish of applicators not as specified.
Style, design and dimensions of applicators not as specified.
Operative ends have kinks, cracks, or other surface imperfections.
Latch end of shank not cleanly machined.
Size assortment of applicators in set not furnished as specified.

Sampling for tests on raw materials. A sufficient quantity from each lot of metal used in the construction of the applicators shall be selected for chemical analysis. A minimum of three specimens from each lot of steel shall be selected for tensile strength determination.

Tests. Tests shall be conducted to determine compliance with specification requirements. The unit of product for test purposes shall be one applicator.

Tensile strength. Tests for tensile strength of the piano wire shall be in accordance with ASTM A228M for testing method only.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pounds units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-00-935-1101	APPLICATOR SET, PIN-RETENTION, DENTAL, Spiral, Cement

Ordering data. Intermediate/exterior package quantities, labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2272
LOCATION: NCR DISK 2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53907	APPLICATOR SET, PIN RETENTION, DENTAL, SPIRAL CEMENT

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*: _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26109:262
Check the source to verify that this is the current version before use.

(FOLD BACK THIS FORM, CUT ALONG THIS LINE.)

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-BSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.