A-A-53908
4 September 1989

COMMERCIAL ITEM DESCRIPTION

SYRINGE, DENTAL IMPRESSION MATERIAL AND ACCESSORY KIT

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-S-36894A, which is cancelled.

This commercial item description covers a dental syringe with accessories designed for injecting light-bodied impression material into prepared teeth.

Salient characteristics:

Types. The syringe and accessory kit shall be of the following types:

Type I - Syringe with accessories
Type II - Accessory kit

Material. The type I syringe and type II accessory kit shall be fabricated from the material listed below:

Component	Material
Syringe body and piston	Rigid white thermoplastic acetal resin
Syringe '0' ring	Synthetic rubber
Cap nut	Rigid acetal resin
Type I syringe curved tip	Translucent, colorless, semi-rigid cellulose acetate butyrate
Type I syringe straight tip	Translucent, colorless, high density polyethylene
Type II accessory kit tip	Translucent, colorless, semi-rigid polyethylene or polypropylene
Adapter	Anodized type 6061-T6 aluminum
Brush bristles	Nylon

AMSC N/A FSC 6520
DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Component	Material
Brush wire	Corrosion resistant metal
Lubricant	USP grade white petrolatum

Style and design. The type I syringe shall include the following components: one syringe (body, piston, '0' ring and cap nut), one straight tip, one curved tip, one cleaning brush, one extra '0' ring and one tube of lubricant. The tube shall contain a minimum of 1/8 ounce of lubricant. The style and design of the type I syringe shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 24278.

The type II accessory kit shall contain 100 tips. For the alternate tip design, the kit shall consist of 100 tips and two adapters. The style and design of the type II accessory kit shall be in accordance with DPSC Drawing No. 24600.

Accessory kit. For the alternate tip design, the adapters shall securely hold the tips in the impression material syringe.

Performance. There shall be no leakage or clogging and the flow of impression material shall be uniform when tested as specified herein.

Finish. Surfaces of the type I syringe and accessories and the type II accessory kit shall be smoothly finished. Components of syringe shall be free of flash and burrs.

Marking. The type I syringe body shall be legibly and permanently marked with the name or registered trademark of the manufacturer. In addition, syringes of foreign manufacture shall be marked with the name of the country of origin.

Instructions. Adequate and accurate instructions for assembly, use, disinfection, and cleaning of the syringe shall be included with each type I syringe. Instructions detailing the assembly and proper use of the type II accessory kit shall be provided with each kit. For the alternate tip design, the instructions shall include the statement that one adapter shall fit a syringe having 14 threads per inch and the second adapter shall fit a syringe having 20 threads per inch.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Workmanship. The type I syringe and type II accessory kit shall be free from defects which may detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing one syringe with accessories, type I, or one accessory kit, type II, as specified, constitutes one unit.

Unit package. Each type I syringe and Type II accessory kit shall be packaged in a container designed to prevent damage to any component during shipment or storage.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an inspection level of II. The unit of product for examination purposes shall be one type I syringe and one type II accessory kit, except that a minimum of 3 tips, or for the alternate design: 3 tips and 2 adapters per accessory kit selected. Examination shall include but shall not be limited for the following defects.

Defects

Style and design not as specified.
Finish not as specified.
Instructions incomplete or missing.
Marking not as specified.
Components of syringe do not fit properly.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, with an inspection level of S-2. The unit of product for test purposes shall be one type I syringe and one type II accessory kit.

Performance.

Type I syringe. The syringe shall be assembled and then partially filled with light-bodied polysulfide rubber based impression material (See Note 1). The syringe shall then be slowly emptied of impression material.

Type II accessory kit. The accessory kit shall be attached to the type I impression material syringe and the syringe performance test repeated.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pounds units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

National Stock Numbers	Item Identification
6520-00-469-3497	Syringe, Dental Impression Material
6520-01-105-4173	Accessory Kit, Dental Impression Material Syringe

NOTE 1: This specification refers to National Stock Number 6520-00-764-2262, Impression Material, Dental Polysulfide Paste, light-bodied.

Ordering data. Intermediate/exterior package quantities, labeling and marking must be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

A-A-53908

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2269
LOCATION: NCR DISK 2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.