A-A-53911
10 October 1989

COMMERCIAL ITEM DESCRIPTION

CARVER, DENTAL (Amalgam and Wax)

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Federal Specification GG-C-1265C, dated 19 November 1976, which is cancelled.

Salient Characteristics:

This commercial item description covers carvers suitable for carving amalgam restorations and wax patterns. Carvers shall be of the following designations.

Wall No. 3
Tanner No. 5T
Tanner No. 6T
Black No. 89-92
Hollenback No. 1/2
Hollenback No. 3

Materials.

One-piece. One-piece instruments shall be fabricated entirely of Class 4, type 440A martensitic stainless steel as specified in ASTM F899, for composition only.

Three-piece.

Handle. The handles of the three-piece instruments shall be fabricated of Class 6, type 430F ferritic stainless steel as specified in ASTM F899, for composition only or C26000 brass alloy, light annealed, chromium over nickel plated as specified in ASTM B135 for composition only.

Operative end. The operative ends of the three piece instruments shall be fabricated of Class 4, type 440A, 440B or 440C martensitic stainless steel as specified in ASTM F899, for composition only.

AMSC N/A FSC 6520
Distribution Statement A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Hardness. Operative ends shall have a Rockwell hardness of C52 to C56 (type 440A), C54 to C58 (type 440B); C56 to C60 (type 440C) after proper stress relieving. Handles and shanks shall be properly tempered and shall have a hardness suitable for the purpose intended. Hardness shall be determined as specified herein.

Design and dimensions. Carvers shall be double-ended. Carvers shall be suitable for carving amalgam restorations and wax patterns. The design and dimensions of the carvers shall conform to Defense Personnel Support Center (DPSC) Drawings listed below.

Drawing numbers

Name	Drawing
Hollenback No. 1/2	24377
Hollenback No. 3	24377
Wall No. 3	24378
Black No. 89-92	24378
Tanner No. 5T	24379
Tanner No. 6T	24379

For three-piece carvers, alternate handles 1,2,3 or 4 shall be furnished.

Finish.

Operative ends and shanks. The operative ends of the carvers shall be finished to provide a sharp edge suitable for carving amalgam and wax patterns. The finish of the operative ends of the carvers shall be as indicated below. The shanks shall be smooth and shall have a bright polished finish.

Finish of Operative Ends

Carver	End shape	Finish
Wall No. 3	Spoon end, Outer face	Bright
	Spoon end, Inner face	Fine hone
	Chisel end	Fine hone
Tanner No. 5T	Spoon end, Outer face	Bright
	Spoon end, Inner face	Fine hone

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Finish of Operative Ends (continued)

Tanner No. 6T	Side faces of blade Top face of blade	Fine polished Bright or fine polished
Black No. 89-92	Cleoid and Discoid ends Outer face Cleoid and Discoid ends Inner face	Bright Fine hone
Hollenback No. 1/2 and Hollenback No. 3	(See applicable drawing)	Fine hone

Handles.

Octagonal. Octagonal handles of one-piece carvers shall have a bright polished finish. Handles shall have 40 ± 10 fillet or notch-type serrations per inch on seven faces and shall have one smooth face, as shown in the applicable DPSC Drawing. Depth of serrations shall be 0.002 to 0.004 inch.

Round. Round handles of the three-piece carvers shall have a bright polished finish, except for the finger grip portion which shall be finely serrated or fine diamond knurled.

Corrosion resistance. The carvers shall show no evidence of corrosion when tested as specified herein.

Marking. The carvers shall be permanently and legibly marked as specified herein. The instrument number shall be placed on the handle no less than 5 mm nor more than 12 mm from the junction of the handle and shank of the applicable operative end. The carvers shall be marked with the name or registered trademark of the manufacturer or contractor and the word stainless. Instruments of foreign manufacture shall be marked with the name of the country of manufacture. As an alternate to the word "stainless", the use of a registered trade name identifying the steel as stainless shall be acceptable, provided prior approval is received from the cognizant contracting officer.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Workmanship. The carvers shall be free from coarse grind marks, cracks, burrs, pits, or feathers or other defects which may detract from their appearance or impair their serviceability, proper functioning, or intended use.

Unit. Each (EA). One carver, as specified, constitutes one unit.

Unit package. The carvers shall be wrapped in neutral tissue or other anti-corrosive material and shall be enclosed in a close fitting container. The carvers shall be so packaged to prevent damage to the operative end.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality for raw material, supplied by the manufacturer of the steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection, examination and test characteristics. The unit of product for inspection, examination and test purposes shall be one carver. Unless otherwise specified, sampling for examination shall be conducted at inpsection level II, except for dimensions which shall be conducted at inspection level S-1, and sampling for tests (when applicable) shall be conducted at inspection level S-2 and acceptance number shall be zero (0) for all sample sizes.

Examination. Examination shall be as specified herein. Examination shall include, but shall not be limited to the following:

Design not as specified for the applicable style of instrument.
Finish of operative end and shank not as specified.
Finish of handle not as specified.
Carving edges not properly sharpened.
Dimensions not as specified on applicable drawing.
Identification marking not present, not properly located, illegible, or complete.

Examination under magnification. The instruments shall be examined under four power (4x) magnification for the following defects:

Defects

Carving edges not properly sharpened.
Operative end not free from coarse grind marks, cracks, burrs, pits, feathers and other imperfections.

Tests.

Chemical analysis. In the absence of certificates of quality, the chemical analysis of the instrument shall be determined in accordance with method 111 or 112 of FED-STD-151. In the event of a conflict, method 111 shall govern.

Hardness. Hardness shall be determined in accordance with ASTM E18 (Rockwell) or E92 (Vickers or Diamond Pyramid) and E140.

Corrosion resistance. The carvers shall be subjected to the boil and copper sulfate tests in accordance with ASTM F1089.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine of the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document:

NATIONAL STOCK NUMBERS	ITEM IDENTIFICATION
6520-00-511-5450	CARVER, DENTAL, Hollenback, No. 3, Double-ended
6520-00-584-2565	CARVER, DENTAL, Walls, No. 3, Double-ended
6520-00-935-7171	CARVER, DENTAL, Hollenback, No. 1/2, Double-ended
6520-00-935-7252	CARVER, DENTAL, Tanner, No. 5T, Double-ended
6520-00-935-7154	CARVER, DENTAL, Cleoid, Discoid, No. 89-92, Double-ended
6520-01-016-3690	CARVER, DENTAL, Tanner, No. 6T, Double-ended

Ordering data. Intermediate/exterior package quantities and labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6520-2235
Location- NCR Disk 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53911	CARVER, DENTAL (Amalgam and Wax)

3a. NAME OF SUBMITTING ORGANIZATION

b. ADDRESS *(Street, City, State, ZIP Code)*

4 TYPE OF ORGANIZATION *(Mark one)*

- ☐ VENDOR
- ☐ USER
- ☐ MANUFACTURER
- ☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:26Z
Check the source to verify that this is the current version before use.