A-A-53912
13 June 1990

COMMERCIAL ITEM DESCRIPTION

TRIMMERS, GINGIVAL MARGIN, DENTAL
(Corrosion-Resisting Steel)

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Federal Specification GG-T-1368B, which is cancelled.

Salient characteristics:

This commercial item description covers corrosion-resisting steel dental gingival margin trimmers for use in cavity preparation during dental operative procedures.

Classification. Trimmers covered by this item description shall be of the following styles and numbers as specified in "Identification marking" and "Note" section.

Style 1 - No. 26
Style 2 - No. 27
Style 3 - No. 28
Style 4 - No. 29

Requirements. The trimmers shall be in accordance with the applicable requirements as specified herein.

Material. The trimmers shall be fabricated of Class 4, type 440A martensitic stainless steel, as specified in ASTM F899 for composition only.

Hardness. The operative ends shall have a Rockwell hardness of C52 to C56 after proper stress relieving. The handles and shanks shall be properly tempered and shall have a hardness suitable for the purpose intended. Hardness shall be determined as specified herein.

Style, design and dimensions. The trimmers shall be double-ended and of one-piece construction. The trimmers shall be of Dr. Black's design modified by Dr. Ferrier. The style, design and dimensions shall be in accordance with Defense Personnel Support Center (DPSC) Drawing 24182. The trimmers shall conform to Black's formula as specified in the drawing and as specified in Table I, for the applicable instrument number. The trimmers shall be suitable for planing enamel and cutting dentin in cavity preparation during dental operative procedures.

AMSC N/A FSC 6520
DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

TABLE I. Instrument Number and Applicable Black's Formula

Style	Instrument Number	Black's Formula
1	26	13-95-8-14
2	27	13-80-8-14
3	28	10-95-7-14
4	29	10-80-7-14

Drawings. The drawings forming a part of this commercial item description are not production drawings. Drawings establish limitations and control design and configuration. The supplier shall be responsible for preparing his own shop drawings. The supplier shall select dimensions and tolerances within the limitations prescribed on the drawings to provide a product with proper design, configuration, and function. No deviation from the dimensional limitations prescribed on the drawings is permissible without prior approval of the contracting officer.

Finish.

Operative ends. The operative ends (blades) of the trimmers, including faces of beveled portion forming sharpened edges, shall have a fine, honed finish. The trimmer blades shall have sharp cutting (trimming) edges and shall be free of defects, such as rounded edges, coarse grind marks, cracks, burrs, feathers, and deep pits when examined under magnification, as specified herein.

Handles and shanks. The handles shall be straight and flat. The handles and shanks of the trimmers shall have a bright polished finish. The handles shall have fillet or notch type serrations on seven faces and shall have one smooth face for identification marking.

Corrosion resistance. The trimmers shall meet the copper sulfate and the boil test requirements as specified herein.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Identification marking. The trimmers shall be permanently and legibly marked on the smooth face of the handle. Each trimmer shall be marked with the applicable instrument number and Black's formula as specified in Table I. Trimmers of foreign manufacture shall bear the name of the country of origin. The word "stainless" shall also be marked on instruments of foreign manufacture. As an alternate, to the word "stainless", the use of a registered trade name identifying the steel as stainless shall be acceptable, provided prior approval is received from the cognizant contracting officer. The use of a score line on the handle of the instrument to distinguish the left operative end from the right operative end is optional.

Workmanship. The operative end, shank and handle of the trimmers shall be free from defects such as coarse grind marks, cracks, or burrs which may detract from their appearance or impair their serviceability, proper functioning or intended use.

Unit. Each (EA). One trimmer, as specified, constitutes one unit.

Unit packaging. Each trimmer shall be wrapped in neutral tissue paper or other suitable non-corrosive material. The cutting edge shall be protected in such a manner so as to prevent damage during shipping and handling. As an alternate, each trimmer may be packaged in a rigid plastic tube having a metal or plastic screw cap or a plastic plug-type closure. The trimmer shall be immobilized within the container and the operative end protected with cork or other suitable material.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one trimmer. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, except for operative end dimensions which shall be conducted at inspection level S-1, and sampling for tests (when applicable) shall be conducted at inspection level S-2. For tests, acceptance number shall be zero (0) for all sample sizes.

Examination. Examination shall be as specified herein. Examination shall include, but shall not be limited to the following:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Defects

Style and design for the applicable type, and shape trimmers not as specified.
Cutting edge of operative end rounded or not sharp.
Cutting edge not free from nicks, burrs, feathers, indentations, or other irregularities that affect cutting performance.
Operative end finish not as specified.
Operative end not free of corrosion.
Operative end dimensions not as specified.
Handle and shank finish not as specified.
Handle not straight and flat.
Identification marking not present, not complete, not properly located, or illegible.
Trimmers not free from dirt, polishing compound, or foreign matter.
Handle serrations not uniform or not complete.
Handle serration dimensions not as specified.

Examination under magnification. The instruments shall be examined under four power (4X) magnification for the following defects:

Defects

Cutting edge of operative end rounded or not sharp.
Cutting edge not free from nicks, burrs, feathers, indentations, or other irregularities that affect cutting performance.
Operative end finish not as specified.

Tests.

Chemical analysis. In the absence of certificates of quality, the chemical analysis of the instrument shall be determined in accordance with method 111 or 112 of FED-STD-151. In the event of a conflict, method 111 shall govern.

Hardness. The hardness shall be determined in accordance with ASTM E18 or E92.

Corrosion resistance. The boil test and copper sulfate test shall be determined in accordance with ASTM F1089.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pounds units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document:

National Stock Numbers	Item Identification
6520-00-935-7185	TRIMMER, GINGIVAL MARGIN, Black's Formula 13, 95, 8, 14; No. 26
6520-00-935-7186	TRIMMER, GINGIVAL MARGIN, Black's Formula 13, 80, 8, 14; No. 27
6520-00-935-7187	TRIMMER, GINGIVAL MARGIN, Black's Formula 10, 95, 7, 14; No. 28
6520-00-935-7188	TRIMMER, GINGIVAL MARGIN, Black's Formula 10, 80, 7, 14; No. 29

Ordering data. Intermediate/exterior package quantities, labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2273
LOCATION: NCR DISK 2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
-A-53912	Trimmers, Gingival Margin, Dental (Corrosion-Resisting Steel)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53912	Trimmers, Gingival Margin, Dental (Corrosion-Resisting Steel)

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil – Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.