A-A-53917
19 June 1989

COMMERCIAL ITEM DESCRIPTION

CURETTES, ALVEOLAR, MOLT

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Federal Specification GG-C-1282C which is cancelled.

This commercial item description covers single-ended alveolar curettes of Molt's design suitable for use in dental surgical procedures.

Salient characteristics:

The curettes shall be of the following types and styles as specified below:

Type I. - Straight

Style A. - Molt No. 1
Style B. - Molt No. 2
Style C. - Molt No. 4

Type II. - Angular

Style D. - Molt No. 5L
Style E. - Molt No. 6R
Style F. - Molt No. 9L
Style G. - Molt No. 10R

Material.

Operative ends. The operative ends and shanks of the curettes shall be class 4, type 440A martensitic stainless steel in accordance with American Society for Testing and Materials (ASTM) Standard F899.

Handles. Handles of the curettes shall be stainless steel or brass, as specified below for chemical composition only. Stainless steel shall be class 3, type 304, except with a chromium content of 17 to 20 percent, or class 4, type 416 in accordance with ASTM F899. Brass shall conform to copper alloy number C26000 or C33000 of ASTM B135, or shall have a composition of 62 to 71 percent copper; 0.5 percent lead, maximum; 0.2 percent iron, maximum; 0.2 percent other elements, maximum; and the remainder zinc. Brass handles shall be nickel plated or chromium-over-nickel plated.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Hardness. After proper stress relieving, the operative ends shall have a Rockwell hardness of C52 to C56 when tested as specified in ASTM E18 or E92. Handles and shanks shall be of proper temper and shall have a hardness suitable for the purpose intended.

Style, design and dimensions. The curettes shall be single-ended curettes of Molt design. The design and dimensions of the curettes shall conform to Defense Personnel Support Center (DPSC) Drawing No. 20282 for the applicable Molt number. Curettes shall be alveolar bone dental surgical curettes suitable for use in oral surgical procedures. Curettes shall be suitable for use as exploratory instruments and to remove tooth particles or debris from sockets, to enucleate cysts, dental granulomas and cystic neoplasms, and to remove small sequestra from sockets.

Construction. Curettes shall be of two piece construction. Curettes shall consist of an operative end (blade), shank, and handle. The shanks shall be permanently secured in the handle by brazing. The interface of the shank and handle shall be smoothly blended with no crevices evident. There shall be no movement, loosening, or separation between the operative end and the handle when the operative end is subjected to an 85 pound (minimum) pull out force.

Finish. The operative ends and shanks shall have a bright finish except for the inside surface and curetting edge of the spoon which shall have a fine honed finish. The operative ends shall have a sharp edge around the periphery of the blade and shall be free of defects such as coarse grind marks, cracks, burrs, deep pits, feathers, turned over edges, and other imperfections when viewed as specified herein. Brass handles shall have a longitudinal scratch, sandblast, or satin finish. Stainless steel handles shall have a satin finish.

Corrosion resistance. Corrosion resistance tests in accordance with ASTM F1089 shall be applicable. The entire instrument shall meet the boil test requirement. Stainless steel parts shall meet the copper sulfate test requirement.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Identification marking. Curettes shall be permanently and legibly marked on the shank of the instrument (area near the junction with the handle). Marking shall include the name or registered trademark of the manufacturer and the word "stainless". Curettes of foreign manufacture shall bear the name of the country of origin. Curettes shall also be marked with the applicable Molt number.

Workmanship. The curettes shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One curette, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one curette. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, except for dimensions, which shall be conducted at inspection level S-1, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1. For tests, the acceptance number shall be zero (0) for all sample sizes.

Examination. The curettes shall be examined for defects including, but not limited to, those listed below.

Defects

Style, design and dimensions not as specified for the applicable type, style, and number of curette.
Construction of curette not as specified.
Finish of operative end not as specified.
Operative end edge not sharp or has turned over edges when viewed under four power (4X) magnification.
Operative end and shank not free from coarse grind marks, cracks, burrs, and deep pits when viewed under four power (4X) magnification.
Shank and handle finish not as specified.
Edges and corners of handle not free of burrs and sharpness.
Identification marking not present, not properly located, illegible, or incomplete.
Instrument not free from dirt, oil, polishing compound, or foreign matter.
Plating of handle (brass handle only) not uniform in appearance.
Plating of handle (brass handle only) not free of blisters, visible porosity, or other defects.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pounds units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document:

National Stock Number	Item Identification
6520-00-519-6550	Curette, Alveolar, Molt, No. 1
6520-00-519-6600	Curette, Alveolar, Molt, No. 2
6520-00-519-6700	Curette, Alveolar, Molt, No. 4
6520-00-519-6740	Curette, Alveolar, Molt, No. 5L
6520-00-519-6770	Curette, Alveolar, Molt, No. 6R
6520-00-519-6800	Curette, Alveolar, Molt, No. 9L
6520-00-519-6900	Curette, Alveolar, Molt, No. 10R

Ordering data. Intermediate/exterior package quantities, labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2278
LOCATION: NCR DISK 2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53917	CURETTES, ALVEOLAR, MOLT

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-BSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.