
FED. SUP CLASS
6640



PART NUMBER	CAPACITY		GRADUATION INTERVAL		NUMBERING INTERVAL		INITIAL NUMBERING	
	A	B	A	B	A	B	A	B
MS35956-1	25 mL	8 dr	5 mL	1.0 dr	5 mL	2.0 dr	5 mL	2.0 dr
MS35956-2	50 mL	2 oz	5 mL	0.25 oz	10 mL	0.5 oz	10 mL	0.5 oz
MS35956-3	100 mL	4 oz	10 mL	0.5 oz	20 mL	1.0 oz	20 mL	1.0 oz
MS35956-4	250 mL	8 oz	25 mL	0.5 oz	50 mL	1.0 oz	50 mL	2.0 oz
MS35956-5	500 mL	16 oz	25 mL	1.0 oz	50 mL	2.0 oz	100 mL	4.0 oz
MS35956-6	1000 mL	32 oz	50 mL	2.0 oz	100 mL	4.0 oz	200 mL	8.0 oz
MS35956-7	5 mL	60 min	1 mL	5 min	1 mL	10 min	1 mL	10 min
MS35956-8	10 mL	120 min	2 mL	†	2 mL	†	2 mL	10 min

† 10-40 in 5 min and 40-120 in 10 min

MATERIAL: Graduates shall be of borosilicate glass in accordance with ASTM E 438, Type I, Class A.

DESIGN: Graduates shall be the pharmaceutical type, conical shape with pourout spout, hex base and double scale; metric on one side (A) and apothecaries on the reverse side (B). Marking shall be etched or engraved as specified in Federal Specification NNN-C-692.

NOTES:

1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.
2. For design feature purposes, this standard takes precedence over procurement document referenced herein.
3. Referenced documents shall be of the issue in effect on date of invitations for bid.
4. FSCM: 11273, 14674, 75278, 84116.

APPROVED 25 Sep 1959 REVISED (A) 26 Mar 1969 (B) 4 Dec 1985

(B) ENTIRE STANDARD REVISED

P.A.	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
EA		GRADUATE, LIQUID LABORATORY GLASS, CONICAL	MS 35956
Other Cust MS 99			
Procurement Specifications NNN-G-692	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED)

PROJECT NUMBER 6640-1308

PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

USER ACTIVITIES: ARMY - ME, PS

REVIEWER ACTIVITIES: ARMY - MI
NAVY - YD
AIR FORCE - 82

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:31Z
Check the source to verify that this is the current version before use.