INCH-POUND
MIL-B-36558B
24 January 1992
SUPERSEDING
MIL-B-36558A
14 September 1978

MILITARY SPECIFICATION

BEDS, ADJUSTABLE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers two types of lightweight, folding, adjustable hospital beds for use by medical personnel in field or fixed installations and as indicated in 6.6.

2. APPLICABLE DOCUMENTS

2.1 Government Documents.

2.1.1 Specifications, standards, and handbooks. The following specifications, standards, and handbooks form a part of this document to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specifications and Standards (DoDISS) and supplement thereto, cited in the solicitation (see 6.2).

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-MST, 2800 South 20th Street, Philadelphia, PA 19145, by using the self-addressed Standardization Document Improvement Proposal, (DD Form 1426) appearing at the end of this document or by letter.

AMSC/NA FSC 6530
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

Federal

PPP-B-601	-	Boxes, Wood, Cleated Plywood
PPP-B-621	-	Box, Wood, Nailed and Lock-corner
PPP-B-636	-	Boxes, Shipping, Fiberboard

Military

MIL-P-116	-	Preservation, Methods of
MIL-M-7298	-	Manual, Technical, Commercial Equipment
MIL-L-10547	-	Liner, Case and Sheet, Overwrap, Water-Vaporproof, or Waterproof, Flexible
MIL-C-37734		Cover, Mattress Pad
MIL-P-37735	-	Pad, Mattress, Hospital Bed
MIL-S-37826	-	Stand, Bed Service
MIL-C-37827	-	Cover, Bed Frame
MIL-R-37828	-	Rod, Insect Bar
MIL-B-37829	-	Bag, Personal Effects

STANDARDS

Federal

FED-STD-151	-	Metal, Test Methods
FED-STD-595	-	Colors

Military

MIL-STD-105	-	Sampling Procedures and Tables for Inspection by Attributes
MIL-STD-129	-	Marking for Shipment and Storage
MIL-STD-147	-	Palletized Unit Loads
MIL-STD-794	-	Parts and Equipment, Procedures for Packaging of

(Unless otherwise indicated, copies of federal and military specifications, standards, and handbooks are available from:

Standardization Documents Order Desk, Building 4D, 700 Robbins Avenue, Philadelphia PA 19111-5094).

2.1.2 Other Government documents, drawings and publications. The following other Government documents, drawings, and publications form a part of this document to the extent specified herein. Unless otherwise specified, the issues are those cited in the solicitation.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

DRAWINGS

Defense Personnel Support Center (DPSC)

25024 - Bed, Folding, Hospital, Field
25520 - Bed, Folding, Hospital, Field, Modified

2.2 Non-Government publications. The following documents form a part of this document to the extent specified herein. Unless otherwise specified, the issues of the documents which are DoD adopted are those listed in the issue of the DoDISS cited in the solicitation. Unless otherwise specified, the issues of documents not listed in the DoDISS are the issues of the documents cited in the solicitation (see 6.2).

ASTM

ASTM D-522 - Test Method for Elongation of Attached Organic Coatings
ASTM B-117 - Method of Salt Spray (Fog) Testing

(Application for Copies should be addressed to ASTM, 1916 Race Street, Philadelphia, PA 19103-1187).

(Non-Government standards and other publications are normally available from the organizations that prepare or distribute the documents. These documents also may be available in or through libraries or other informational services.)

2.3 Order of precedence. In the event of a conflict between the text of this document and the references cited herein (except for related associated detail specifications, specification sheets, or MS standards), the text of this document takes precedence. Nothing in this document, however, supersedes applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3.1 First article. When specified in the procurement document, the contractor shall furnish sample unit(s) for first article inspection and approval (see 4.1.3.1 and 6.3).

3.2 Material. Except as hereinafter specified, materials shall conform to the applicable portions of the specifications, standards, and drawings listed herein. Materials shall be subjected to the tests and meet the requirements specified in

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

these documents. The metals shall conform to the chemical and physical requirements of the applicable specifications. Materials, not specifically designated herein, but used in the fabrication of the bed, shall be suitable for the purpose intended.

3.3 Design and construction. The design and construction of the folding, adjustable, hospital bed shall be in conformance with DPSC Drawing No. 25024 for type I beds and Drawing No. 25520 for type II beds. All of the components listed below shall fit into their intended locations. All adjustments shall function smoothly throughout their design ranges. All components shall fit into the configuration of the bed folded 180 degrees onto itself for storage and shipment. When the bed is fully erected, the outer frame shall be 82 plus or minus 1 inches long by 30 plus or minus 1/2 inches wide, and the top of the outer frame shall be 28 plus or minus 1/2 inches high. Tolerances indicated on the DPSC drawings shall not be cumulative so as to exceed the tolerances listed above in overall dimensions. The bed shall consist of the components listed in 3.3.1 through 3.3.7 (also see 6.7).

3.3.1 Inner and outer frame. The frame shall be complete with cables, legs, locking assemblies, and elevating arcs.

3.3.2 Bed frame cover. The bed frame cover shall be in conformance with MIL-C-37827. The cover shall be nylon with a velcro type fastener, sewed on the patient side of the cover, that shall attach to the mattress cover to prevent the mattress from sliding when a bed end is elevated.

3.3.3 Mattress pad. The mattress pad shall be in conformance with MIL-P-37735. The pad shall be closed cell vinyl foam, 3/8 inch thick. Overall pad dimensions shall be 76-1/2 plus or minus 2 inches in length and 27-1/2 plus or minus 1/2 inches in width.

3.3.4 Mattress pad cover. The mattress pad cover shall be in conformance with MIL-C-37734. The pad cover shall be nylon duck dimensioned to fit the foam pad and with a velcro type closure sewed on the patient side of the cover.

3.3.5 Personal effects bag. The personal effects bag shall be in conformance with MIL-B-37829. The personal effects bag shall be olive drab nylon duck, 16 inches in width, 12 inches in height, and with three pockets each 10 inches deep.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

3.3.6 Bed service stand. The aluminum bed service stand shall be in conformance with MIL-S-37826. The overall dimensions of the stand shall be 11-5/8 inches in width by 24 inches in length with one 1-1/2 inch flange along the 24 inches edge.

3.3.7 Insect-bar rods. The two aluminum insect-bar rods shall be in conformance with MIL-R-37828.

3.4 Performance.

3.4.1 Assembly. When assembled and tested as specified in 4.5.2, the bed shall have the correct configuration, all accessories shall fit properly in all intended locations, and all adjustments shall function as specified.

3.4.2 End elevation. Maximum elevation angle, for each end of the elevating frame, shall be 65 degrees plus or minus 5 degrees. The elevating frame shall not slip when tested as specified in 4.5.3.

3.4.3 Strength of elevating frame. The bed shall exhibit no permanent set, slippage, or other failure when tested as specified in 4.5.4.1.

3.4.4 Load bearing capacity. There shall be no failure of any component part of the bed frame when tested as specified in 4.5.4.2.

3.4.5 Stability. The plumb bob shall not oscillate outside the 3-inch circle when the bed is tested as specified in 4.5.5.

3.4.6 Elevating-frame locking cam. The locking cam shall release the elevating frame when a pressure of not less than 20 pounds, nor more than 35 pounds, is applied, when tested as specified in 4.5.6.

3.5 Finish.

3.5.1 Powder Coating. An epoxy powder formulated for electrostatic spray equipment application shall be used as the finish on metal surfaces. The epoxy coating shall be fast curing, and shall provide superior adhesion with a tough mar-resistant smooth finish that withstands high impact without chipping or cracking. See 6.5 for availability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

3.6 Marking.

3.6.1 Caution. The top of the end rail to the right of the elevating frame locking cam, at each end, shall be marked "CAUTION, OPERATE RELEASE WITH HAND UNDERNEATH THIS RAIL." Marking shall be clearly visible in conspicuously colored letters at least 1/4 inch high, except for "CAUTION", which shall be 1/2 inch high. Lettering shall be permanent and legible. Marking may be made directly onto the finished surface, or by means of an attached metal plate.

3.6.2 Identification. The bed shall be marked, permanently and legibly, with the contract or purchase order number, the National Stock Number, the item identification, and the name or registered trademark of the manufacturer. Marking shall be on the underside of the bed frame cover, separate and distinct from the cover's own identification marking. The bed components shall be marked in accordance with the specification indicated in the appropriate paragraph therein, 3.3.2 through 3.3.7.

3.7 Service data. The contractor shall furnish, with each unit, two copies each of a technical manual which contains complete instructions for assembly, disassembly, operation, maintenance, and lists of component parts. Assembly and disassembly shall be fully illustrated and explained comprehensively in simple steps. The disassembly instructions required as part of the service data shall indicate the step-by-step procedure for reclosing the bed prior to storage. The procedure shall be described as follows:

1. Disassemble the bed by reversing the assembly procedure.

2. Unbuckle the elevating frame cover side support flaps and fold them inside the cover.

3. Lay the pad assembly and its cover inside one half of the bed.

4. Arrange all bed accessories neatly on the pad, including the two bed legs.

5. Fold both ends of the pad over the accessories, approximately 12 inches on one end and 20 to 24 inches on the other end.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

6. Fold the main frame, leaving the elevating section open.

7. Tuck the remaining two legs into the pad.

8. Fold the elevating section sheet and strap into place.

The above is intended to cushion the accessories and bed components, thereby preventing possible damage to the elevating frame cover.

The type II bed legs are hinged and not separate loose pieces as on the type I bed. Therefore, steps 4 and 7 shall not apply for the type II bed as far as the legs are concerned, but the legs shall be closed after the wrapped accessories are in place. The manual shall comply with the requirements of MIL-M-7298. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the Contracting Officer two preliminary copies of his proposed technical manual for review and comment. The preliminary manuals shall be submitted sufficiently early to permit adequate review (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract or purchase order. In addition, the contractor shall furnish 15 copies of the approved technical manual to the Contracting Officer for distribution. Should the contractor have furnished acceptable manuals for the identical item within two years preceding the date of the contract and propose to furnish manuals identical to those previously accepted, then the requirements for review of drafts and submission of 15 additional copies may be waived upon request to the Contracting Officer. Waiver of submission of drafts and submission of 15 additional copies shall be at the option of the Government.

3.8 Workmanship. The beds shall be free from defects which detract from their appearance or impair their serviceability. All burrs and sharp edges shall be removed from the beds.

3.8.1 Welding. Welding shall be durable, sound in every detail and in accordance with the latest practices recommended by the American Welding Society. All welds shall be produced by qualified welding operators with satisfactory welding equipment using satisfactory filler metal or electrode where applicable to provide the best possible weld. Surfaces to be welded shall be free from grease, oil, oxides or any other foreign matter. Any cleaning method which will remove these substances and leave the

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

surface clean is satisfactory. All weld joints shall blend into the adjacent metal in gradual, smooth curves. Beads shall be smooth and free of slag, excessive undercut, or excessive splatter. Sufficient metal shall be added to provide a suitable fillet or reinforcement unless otherwise specified. Welds shall not be oxidized. All welds shall be continuous with no visible voids, pin holes, or cracks.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements (examinations and tests) as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable or the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification where such inspections are deemed necessary to ensure supplies and services conform to prescribed requirements.

4.1.1 Responsibility for compliance. All items shall meet all requirements of sections 3 and 5. The inspection set forth in this specification shall become a part of the contractor's overall inspection system or quality program. The absence of any inspection requirements in the specification shall not relieve the contractor of the responsibility of ensuring that all products or supplies submitted to the Government for acceptance comply with all requirements of the contract. Sampling inspection, as part of manufacturing operations, is an acceptable practice to ascertain conformance to requirements, however this does not authorize submission of known defective material, either indicated or actual, nor does it commit the Government to accept defective material.

4.1.2 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

4.1.3 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3.1 Classification of inspection. The inspection requirements specified herein are classified as follows:

a. First article inspection.
b. Quality conformance inspection.

4.1.3.1.1 First article inspection. First article inspection shall include examination of 4.4 and tests of 4.5.

4.1.3.1.2 Quality conformance inspection. Quality conformance inspection shall include the examination of 4.4 and tests of 4.5.

4.1.4 Certificates of quality. Certificates of quality, supplied by the manufacturer of the metals may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government, and provided also that the test methods and the results obtained comply with those described in this specification. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Inspection lot.

4.2.1 Raw materials. An inspection lot for raw material inspection shall be all material produced under the same conditions and controlled to the degree that uniformity of product within the designated lot can be assured. Lot identification of the raw material by the manufacturer is one method of control, as cited above, which is considered acceptable.

4.2.2 End product. An inspection lot for inspection of end product shall consist of all beds of the same type produced under essentially the same conditions and offered for inspection at one time.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

4.3 Sampling.

4.3.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and table I. The unit of product for sampling purposes shall be one bed or component, as ordered.

TABLE I. Sampling for examination

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Total defects (Major and Minor combined)	II	4.0
For dimensional examination	S-1	2.5

4.3.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II, with an inspection level of S-4 and an acceptance level of zero.

TABLE II. Sampling for tests

Component Unit of product)	Characteristic	Requirement	Test procedure
Metal	Chemical composition	3.2	4.5.1
Bed	Assembly	3.4.1	4.5.2
	End elevation	3.4.2	4.5.3
	Load (elevating frame)	3.4.3	4.5.4.1
	Load (complete bed)	3.4.4	4.5.4.2
	Stability	3.4.5	4.5.5
	Elevating catch mechanism	3.4.6	4.5.6

4.3.3 Raw materials. A sufficient quantity of each lot of each metal used in the manufacture of the beds shall be selected for chemical analysis.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

4.4 Examination. The bed shall be examined to determine compliance with all requirements contained in this specification.

4.4.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table III. Examination shall not be restricted to the classified possible defects listed below.

TABLE III. Classification of defects

Categories	Defects
Major	
101	Any component part missing, broken, or otherwise unusable.
102	Exposed edge not free of burrs or sharpness.
103	Components not properly assembled or aligned.
104	Nylon cover not taut (when bed is opened).
105	Bed cannot be opened and erected as required.
106	Bed ends do not elevate and hold in position as specified.
Minor	
201	Identification marking not complete, permanent, legible, and waterproof.
202	Finish not free of chipped areas, cracks, scratches, or roughness.
203	Style, design and construction not as shown on drawing.
204	Bed does not rest evenly on a flat surface.

4.4.2 Dimensional examination. The bed shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4.3 Inspection of packaging. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.5 Tests. Tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

4.5.1 Chemical composition. Four ounces from each lot of metal shall be tested in accordance with FED-STD No. 151.

4.5.2 Assembly. The bed shall be erected in accordance with the directions supplied. All accessory items shall be fitted into all designed locations to check for proper fit.

4.5.3 End elevation. The bed shall be fully assembled, except for the service stand and insect-bar rods. Each end shall be elevated to 10, 20, 30, 40 degrees, and the maximum setting (all measured from the horizontal). The bed ends shall be examined for slippage at each position.

4.5.4 Load bearing capability.

4.5.4.1 Elevating frame. With the bed assembled as specified in 4.5.2, both ends shall be elevated to an angle of 45 degrees plus or minus 2 degrees. A uniformly distributed load of 60 pounds shall be placed along the end rail of the elevating frame.

4.5.4.2 Complete bed. With the bed assembled as specified in 4.5.2, and the ends horizontal, uniformly distributed load of 900 pounds shall be placed on the load-bearing surface and allowed to remain for 30 minutes.

4.5.5 Stability. With the bed surface in a horizontal position, the legs shall be blocked to prevent slippage. A plumb bob shall be fastened by a cord at the center of a transverse outer frame member. The point of the plumb bob shall be located slightly above, and centered over a 3-inch diameter circle on the floor. An individual weighing approximately 160 pounds shall lie down on the bed, roll over, move about, and sit up in the bed. The individual shall sit on the side of the bed and move along the length of the bed. The movements of the individual shall approximate, as closely as possible, those movements of a typical hospital patient. During all movements of the individual, the plumb bob shall be observed to determine whether the bob oscillates outside the 3-inch diameter circle.

4.5.6 Elevating catch mechanism. One end of the bed shall be elevated to an angle of 25 degrees. A uniformly distributed load of 40 pounds shall be placed along the end rail of the elevating frame. A spring balance shall be attached, perpendicular to the release trigger, and the pull necessary to release the cam shall be recorded.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C, as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Cleaning and drying. The bed shall be cleaned and dried in accordance with MIL-P-116.

5.1.1.2 Preservative applications. Preservatives shall not be used.

5.1.1.3 Unit package. Each bed along with components shall be clean and dry. The I.V. and insect-bar rods, elevating arcs, bed side tray, vinyl foam pad, pad cover and the patient's effects bag shall be neatly folded and individually wrapped in kraft paper or other suitable material. The legs shall be detached from their sockets on the bed frame but not removed from the support cables, and wrapped in kraft paper or other suitable material. All wrapping shall be secured with pressure sensitive or kraft gummed tape. The bed frame with the canvas elevating cover in place shall be folded. The wrapped components shall be neatly inserted between the folded halves of the bed frame. The frame shall be secured in the folded position by buckling of the two end straps. Closure shall be as specified in the applicable box specification.

5.1.2 Level C.

5.1.2.1 Unit package. Preservation for the bed shall conform to the requirements of MIL-STD-794 for level C.

5.2 Packing. Packing shall be level A, B or C, as specified (see 6.2).

5.2.1 Level A. Each bed packaged as specified in 5.1.1.3 shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, overseas type, PPP-B-621, class 2, or PPP-B-636 weather-resistant. Closure and strapping shall be as specified in the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in PPP-B-636.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

5.2.2 Level B. Each bed packaged as specified in 5.1.1.3 shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic type, PPP-B-621, class 1, or PPP-B-636, class domestic. Closure shall be as specified in the applicable box specification.

5.2.3 Level C. Each bed shall be packed in accordance with the requirements of MIL-STD-794 for level C.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination exceed 250 pounds or 20 cubic feet. Loads shall be unitized on pallets conforming to MIL-STD-147. Quantities less than 20 cubic feet need not be unitized. Pallet loads including the pallet shall not exceed 43 inches in length 52 inches in width, and 54 inches in height. Pallet shall conform to type IV or type V.

5.2.5.1 Levels A and B. The bed packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets as specified in 5.2.5.

5.2.5.2 Level C. The bed packed as specified in 5.2.3 shall be unitized as specified in MIL-STD-794.

5.3 Marking.

5.3.1 Levels A, B and C. Each unit pack, intermediate package, exterior container and unitized load shall be marked as specified in MIL-STD-129.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.4.3.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

6. NOTES

(This section contains information of a general or explanatory nature that may be helpful, but is not mandatory.)

6.1 Intended use. The bed is intended for use in field or fixed military medical facilities.

6.2 Acquisition requirements. Purchasers should select the preferred options permitted herein and include the following information in procurement documents.

a. Title, number and date of this specification.

b. Issue of DODISS to be cited in the solicitation and if required, the specific issue of individual documents referenced (see 2.1).

c. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

d. When a first article is required (see 3.1, 4.1.3.1 and 6.3).

6.3 First article. When a first article is required, it shall be tested and approved under the appropriate provisions of the Defense Acquisition Regulation/Armed Services Procurement Regulation. The contract or purchase order specifies the quantity, procedure for delivery and other specific instructions for inspection and approval of the first article.

6.4 Changes from previous issue. Asterisks are not used in this revision to identify changes with respect to the previous issue, due to the extensiveness of the changes.

6.5 Powder coating. A suitable epoxy coating powder is Corvel ECA-1555-GP2, available from the Polymer Corporation, Reading, PA, or equal

6.6 Bed components. The bed components listed in 3.3 are stock listed as follows:

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

National Stock Number	Item Identification
7210-00-118-0008	COVER, BED FRAME, Cotton Duck 26-3/4 by 77-5/8 inches
7210-00-118-0009	PAD, MATTRESS, Hospital Bed, Vinyl Foam, 27-1/2 by 76-1/2 inches
7210-00-118-0010	COVER, MATTRESS PAD, Cotton Duck, 28 by 77 inches
8465-00-115-6029	BAG, PERSONAL EFFECTS, Olive-Drab, Cotton
6530-00-115-6028	STAND, BED SERVICE, Aluminum, 24 by 11-5/8 by 11-5/8 inches
7210-00-118-0011	ROD, INSECT BAR, Aluminum

6.7 Additional stability. The following product may be ordered separately to be used on the type I bed frame when additional stability is required: Brace, Assembly, Adjustable Bed, 2s, NSN: 6530-01-328-5194.

6.8 Subject term (key word) listing.

Bag, Personal Effects
Brace, Legs
Canvas/Duck Support
Cover, Bed
Elevating Ends
Field Use
Folding Bed
Frame, Bed
Mattress, Inflatable
Pillow, Inflatable
Portable Bed
Stand, Service, Bed

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

6.9 Item identification. This specification covers the following items:

National Stock Number	Item Identification
6530-01-015-5189 Type I	BED, ADJUSTABLE, Hospital, Folding, Field.
6530-01-325-6590 Type II	BED, ADJUSTABLE, Hospital, Folding, Field, Modified.

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing activity:

DoD - MB

Review activities:

Army - TM
Navy - YD
Air Force - 50

Agent:

DLA - DM

Civil Agency Coordinating Activities:

FDA - MPQAS
USPHS
VA - OSS

Project No. 6530-2311

Location: BEDSM11.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER MIL-B-36558B	2. DOCUMENT DATE (YYMMDD) 24 January 1992

3. DOCUMENT TITLE
BEDS, ADJUSTABLE

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 198290

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:48Z
Check the source to verify that this is the current version before use.