MIL-G-36592
Amendment-3
1 June 1973
SUPERSEDING
Amendment-2
5 December 1969 and
Amendment-1
31 October 1969

MILITARY SPECIFICATION

GLOVE, PATIENT EXAMINING AND TREATMENT, RUBBER LATEX, NON-STERIL.

This Amendment forms a part of Military Specification MIL-G-36592, dated 11 September 1969, and is approved for use by all Departments and Agencies of the Department of Defense.

Page 3, Paragpraih 3.2.1 add the letter "s" to "Dimension". Table I, Dimensions, change as follows:

Overall length, delete "10" and substitute "9-1/2".

Flat width of forefinger, add "at base", delete "1-13/16" and substitute "1-3/16".

Flat width of middle finger, add "at base", delete "1-7/16 and substitute: "1-1/4".

Page 8:

5.2.1.1 Exterior container. At end of paragraph add: "In addition, fiberboard boxes shall conform to the special requirements as specified in Table II of PPP-B-636."

Page 9:

5.2.2.1 Exterior container. At end of paragraph add: "Closure of fiberboard boxes shall conform to method II as specified in PPP-B-636. In addition, fiberboard boxes shall conform to the special requirements as specified in Table I of PPP-B-636."

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:52Z
Check the source to verify that this is the current version before use.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1735

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:52Z
Check the source to verify that this is the current version before use.