MIL-G-36593
AMENDMENT-1
8 June 1972

MILITARY SPECIFICATION

GLOVE, PATIENT EXAMINING AND TREATMENT, PLASTIC, NON-STERILE

This Amendment forms a part of Military Specification MIL-G-36593, dated 12 September 1969, and is approved for use by all Departments and Agencies of the Department of Defense.

Page 3:

3.2.1 Dimensions. Table I, after "Flat width of base of index, middle and ring fingers" delete "1 to 1-3/16" and insert "1 to 1-3/8".

Page 9:

5.1.2 Packaging quantities. In Table VI, in column headed "Exterior container", delete "12 units" and substitute "20 units".

Custodians:

Army - MD
Navy - MS
Air Force - 03

Review Activites:

Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1652

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:52Z
Check the source to verify that this is the current version before use.