MIL-H-36575A
31 October 1972
SUPERSEDING
MIL-H-36575
10 September 1969

MILITARY SPECIFICATION

HOLDER, RUBBER DAM, DENTAL, "U" SHAPED, AND STRAP

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers rubber dam holders and a holder strap suitable for use during operative procedures.

* 1.2 Classification. The rubber dam holders and the holder strap shall be of the following types and sizes (see 6.1):

Type I. Holder, Rubber Dam, "U" Shaped (Complete).

Size 1. Adult
Size 2. Child

Type II. Strap, Rubber Dam Holder (Replacement).

2. APPLICABLE DOCUMENTS

* 2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Setup.

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap: Water-Vaporproof or Waterproof, Flexible.

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:50Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

24126 Holder, Rubber Dam, Dental.

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material.

* 3.1.1 Holder. The rubber dam holders shall be fabricated from half hard brass rod.

* 3.1.2 Strap. The elastic strap shall be fabricated from cotton, rayon and natural rubber.

* 3.1.2.1 Lock. The strap lock shall be fabricated from suitable carbon steel or brass.

* 3.1.2.2 Loops. The strap loops shall be fabricated from free-cutting brass.

3.2 Style, design, and dimensions.

3.2.1 Type I, holders. The type I holder shall consist of one rubber dam holder and two straps (see 3.1.2 and 3.2.2,as applicable). The type I, size 1, rubber dam holder shall be suitable for use during adult dental operative procedures, and the type 1, size 2, rubber dam holder shall be suitable for use during pedodontic operative procedures. Both the size 1 and the size 2 rubber dam holders shall be suitable for use in supporting and stabilizing a six-inch-square sheet of rubber dam on the face of the patient during dental operative procedures. The style, design, and dimensions of the holders shall be in conformance with Defense Personnel Support Center (DPSC) Drawing 24126 for the applicable type and size. The outer construction of the "U" shall have seven permanently attached pins which shall properly engage and retain the rubber dam. The two vertical projections of the "U" shall each be provided with an offset ball-shaped protective end or knob to insure optimal protection to the eyes of the patient. The outer portions of the vertical projections shall each contain a blunt-headed hook suitable for engaging and retaining the adjustable, elastic, rubber dam holder strap. The straps supplied with each holder shall be in conformance with the requirements of 3.2.2.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:50Z
Check the source to verify that this is the current version before use.

* 3.2.2 Type II, straps. Type II straps shall be adjustable, elastic neck straps suitable for use with type I holders and with six-inch-square rubber dams in providing tension for the establishment of required cheek retraction during dental operative procedures (see 3.1.2). Type II straps shall also serve as replacements for the straps initially furnished with the type I holder. The style, design, and dimensions of the straps shall be in conformance with the applicable portions of DPSC Drawing 24126. The strap shall have a metal loop at each end keeping the strap flat against the patient's face while furnishing an eyelet for attachment of the strap to the rubber dam holder. The strap shall also be provided with a readily manipulated lock to permit ease of tension adjustment by increasing or decreasing the length of the strap. The webbing shall consist of 16 ± 2 ends of natural rubber covered with cotton, a cotton warp, and a rayon filling. For type II, six straps shall be furnished as replacement parts (see 1.2).

3.3 Finish.

* 3.3.1 Type I. The holder shall be chromium-over-nickel plated and shall have a bright polished finish. All soldered joints on the holder (such as the juncture of the rod with the knobs, hooks, and pins) shall be cleanly and properly finished. The two straps supplied with each holder shall be finished in accordance with the requirements of 3.3.2.

* 3.3.2 Type II. The loops of the strap shall be chromium-over-nickel plated and shall have a bright polished finish. The lock of the strap shall be carbon steel, coated with a dark blue high-gloss enamel, or nickel plated carbon steel, or nickel plated brass. The webbing shall be dyed blue, or black.

3.4 Resistance to boil. When subjected to the boil test specified in 4.4.1, all surfaces of the type I holder shall be free from corrosion.

3.5 Elongation and rupture. When tested as specified in 4.4.2, the webbing shall have an elongation of no less than 90 percent and the rubber strands shall show no evidence of rupturing.

3.6 Identification marking. Holder shall be legibly and permanently marked with either the name or registered trademark of the manufacturer.

3.7 Instructions. Accurate and adequate instructions for use shall accompany each holder.

3.8 Workmanship. Workmanship shall be first class throughout. The holders and straps shall be free from defects which may detract from their appearance or impair their serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:50Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification, where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

* 4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of supplies to which such records relate.

* 4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

* 4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the component or material, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the Item Identification, the name of the component or material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in tables I and III (see 4.2.1 and 4.3) and table II (see 4.2.2).

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQLs) shall be as indicated in table I. The unit of product for examination purposes shall be one complete type I holder or one individual type II strap.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:50Z
Check the source to verify that this is the current version before use.

Table I. Sampling for examination

Type of defect	Inspection level	AQL (Percent defective)
Major A	II	1.0
Major B	S-2	2.5
Minor	II	4.0

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II, except that the acceptance number shall be zero for all sample sizes. Unit of product for test purposes shall be one individual holder or one strap.

Table II. Sampling for tests

Component	Characteristic	Requirement	Test procedure
Holder	Resistance to boil	3.4	4.4.1
Webbing	Elongation and rupture	3.5	4.4.2

4.2.3 Sampling of raw materials. A sufficient quantity of each lot of raw material used in the construction of the holders shall be selected for chemical analysis.

4.3 Examination. The holders and straps shall be examined for defects including, but not limited to, those listed in table III.

Table III. Classification of defects

Category	Defect
Major A:	
101	Type or size not as specified.
102	Finish not as specified.
103	Burrs, sharp edges, crevices, rough areas or cracks present in finished metal components.
104	Soldered joints not cleanly and properly finished.
105	Pins on holder do not allow proper engagement of the rubber dam.
106	Hooks on holder unsuitable for proper engagement and retention of strap.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:50Z
Check the source to verify that this is the current version before use.

Major B:

151	Dimensions not as specified.

Minor:

201	Identification marking on holder missing, incorrect, illegible, or incomplete.
202	Instructions for use of holder neither accurate nor adequate.
203	Lock on strap not readily manipulated.
204	Knobs on holder do not provide optimal eye protection.
205	Hooks on holder not blunt-headed.
206	Packaging not as specified.
207	Packing not as specified.
208	Marking on packaging and pack not as specified.

4.4 Test procedures.

4.4.1 Resistance to boil. Before testing, the holders shall be scrubbed with soap and water, rinsed thoroughly in distilled water, and dried. After being cleaned, the holders shall be placed in a glass container in which distilled water is added, boiled for 30 minutes, and allowed to remain submerged for an additional 24 hours.

4.4.2 Elongation and rupture. The test for elongation shall be determined by manually stretching 2.0 inches of webbing to a minimum of 3.8 inches. Any rupture of rubber strands shall be determined both during and after the test.

5. PREPARATION FOR DELIVERY

* 5.1 Packaging. Packaging shall be Level A or C as specified (see 6.1).

5.1.1 Unit of issue. One holder with two straps, or one package (six straps), as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:50Z
Check the source to verify that this is the current version before use.

Table IV. Packaging quantities

Item	Unit package	Intermediate package	Exterior container
Holder	1 unit	12 units	144 units
Strap	1 unit	12 units	288 units

5.1.2.1 Variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Unit package shall be a box of appropriate size and design constructed in accordance with either PPP-B-566 or PPP-B-676, except that commercial colors will be acceptable. Closure shall be adequate to prevent accidental opening under normal handling.

* 5.1.3.2 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-676, except that commercial colors shall be acceptable. Closure shall be adequate to prevent accidental opening.

5.1.4 Level C. Units shall be packaged in standard commercial containers, of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

* 5.2 Packing. Packing shall be Level A, B or C as specified (see 6.1).

5.2.1 Level A.

* 5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table III, using type I, class 1 plywood; PPP-B-621, table III, or PPP-B-636, type CF, class weather-resistant. Closure and strapping shall be in accordance with appendix of applicable box specification. Special requirements of applicable Table shall apply for PPP-B-636 box.

* 5.2.1.2 Case liner. Each Level A wood box shall be lined with a water-proof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.2 Level B.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:50Z
Check the source to verify that this is the current version before use.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table II, using type III, class 1 plywood or shall be designed for type 2 load and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic. Closure shall be in accordance with appendix of applicable box specification. Fiberboard box shall be closed by method II closure of PPP-B-636. Special requirements of applicable Table shall apply for PPP-B-636 box.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers, of the type, size, and kind commonly used for the purpose, and shall be constructed so as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES.

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.

(b) Type and size of item required (see 1.2).

(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 Limitations. This specification does not cover all types or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet Military requirements.

6.3 Specification changes. The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:50Z
Check the source to verify that this is the current version before use.

Custodians
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Proj. No. 6520-1337

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:50Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-H-36575A

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:50Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

Defense Supply Agency

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:50Z
Check the source to verify that this is the current version before use.