MILITARY SPECIFICATION

HOLDER, COTTON ROLL, DENTAL

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers cotton roll holders suitable for maintaining a dry oral cavity during dental operative procedures.

1.2 Classification. Cotton roll holders shall be of the following styles, as specified (see 6.1):

Style 1. Adult Size.
Style 2. Child Size.

2. APPLICABLE DOCUMENTS

2.1 Specifications, standards, and drawing. The following documents, of the issue in effect on date of invitations for bid or requests for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
MMM-A-250	Adhesive, Water-Resistant, (For Sealing Fiberboard Boxes).

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:52Z
Check the source to verify that this is the currentversion before use.

PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, (For Carton Sealing).

Military

MIL-P-116	Methods of Preservation.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

24198	Holders, Cotton Roll, Dental and Saliva Ejector, Right and Left, Adult and Child Sizes.

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 General requirements. The cotton roll holders shall be in accordance with applicable requirements of GG-I-526 and as specified herein.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:52Z
Check the source to verify that this is the current version before use.

3.2 Material. The components of holders shall be fabricated of materials conforming to GG-I-526 as shown in Table I.

Table I. Materials

Component	Composition
Cotton holder section and chin clamp	Type 420 /1
Tubing and hose ferrule	Type 304
Bolt for wingnut and screws for tube holder block	Types 303 or 304
Tube holder block	Type 304
Wingnut	Type 304

/1 Carbon content shall be 0.33 to 0.38 percent.

3.3 Hardness. The cotton holder section (body) and chin clamp shall be hardened, tempered, electro-polished, and stress-relieved to a Rockwell hardness of C46 to C52.

3.4 Style, design, and dimensions. The cotton roll holder set shall consist of two separate holders with saliva ejector attachments: one for the left side, and one for the right side of the mouth. The holders shall be suitable for holding cotton rolls in the oral cavity during dental operative procedures, and for aspirating fluids from the oral cavity. The style, design, and dimensions shall conform to Defense Personnel Support Center (DPSC) Drawing 24198 for Styles 1 and 2, as specified (see 6.1).

3.4.1 Style 1. Style 1 holders shall be the adult size set, Ivory-Darby type saliva ejector and tongue guard cotton roll holders, with two 4-pronged cotton retention sections, as shown in the drawing.

3.4.2 Style 2. Style 2 holders shall be the child size set of the Ivory-Darby type saliva ejector and tongue guard cotton roll holders, with two 4-pronged cotton retention sections, as shown in the drawing.

3.5 Finish. Each set of cotton roll holders shall have a mirror finish, except for wingnuts, bolts, and screws, which shall have a satin or polished finish.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:52Z
Check the source to verify that this is the current version before use.

3.6 Corrosion resistance. The units shall be properly passivated and shall pass the copper sulfate and boil tests specified in GG-I-526.

3.7 Instructions. Adequate and accurate operating instructions shall accompany each set of cotton roll holders.

3.8 Identification marking. Each holder shall be legibly and permanently marked with the name or registered trademark of manufacturer (and contractor, if other than manufacturer). The words "Right" and "Left" shall be marked on the respective right and left holder sections. In addition, Style 2 holders shall be marked with the word "Child". The lettering shall be die-stamped in letters 1/32 inch high. Letters shall be located on the upper surface of the cotton holder sections midway between the retaining portion of the holder and the hinge joint. All marking shall be legible and permanent.

3.9 Workmanship. Workmanship shall be first class throughout. The cotton roll holders shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:52Z
Check the source to verify that this is the current version before use.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in tables II and IV (see 4.2.1 and 4.3) and table III (see 4.2.2).

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQL's) shall be as indicated in table II. Unit of product for examination purposes shall be one complete cotton roll holder set.

Table II. Examination Criteria

Type Defect	Inspection level	AQL (percent defective)
Major A	II	1.0
Major B	S-2	1.5
Minor A	II	2.5
Minor B	S-2	4.0

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in Table III, except that the acceptance number shall be zero for all sample sizes. Unit of product for test purposes shall be one set.

Table III. Tests

Characteristic	Requirement	Test procedure	Inspection level
Composition	3.2	GG-I-526	See 4.2.3
Hardness	3.3	GG-I-526	S-2
Boil Test	3.6	GG-I-526	S-2
Copper Sulfate Test	3.6	GG-I-526	S-2

4.2.3 A sufficient quantity of each lot of metal used in the construction of the holders shall be selected for chemical analysis.

Source: https://assist.dla.mil -- Downloaded: 20260-21T08:52Z
Check the source to verify that this is the current version before use.

4.2.4 Sample unit usage. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. Holders shall be examined for defects including, but not limited to, those listed in table IV.

Table IV. Classification of Defects

Category	Defect
Major A	
101	Holders not style (adult or child) or design (right and left) specified.
102	Edges sharp or not free from burrs.
103	Surfaces not smooth and free from defects which may injure oral tissue.
104	Tubing partially or completely blocked.
105	Soldered joints not sound, clean and free from excess flux.
106	Wingnut does not hold chin clamp properly; excessive force required.
Major B	
151	Thickness of body and chin clamp not as specified.
Minor A	
201	Packaging and packing not as specified.
202	Identification marking not as specified.
203	Block screws not secure.
204	Finish not as specified.
Minor B	
251	Dimensions, other than 151, not as specified.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One holder set, as specified, constitutes one unit of issue.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:52Z
Check the source to verify that this is the current version before use.

5.1.2 Packaging quantities. The number of units of issue indicated in table V shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table V. Packaging Quantities

Unit Package	Intermediate Package	Exterior container
1 unit	12 units	144 units

5.1.2.1 Variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Cleaning and drying. Each unit shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

5.1.3.2 Preservative. Not required.

5.1.3.3 Intimate wrap. Each unit shall be wrapped in neutral tissue or other suitable anticorrosive materials.

5.1.3.4 Unit package. Each unit shall be packaged in accordance with MIL-P-116, method IA-15.

5.1.3.5 Intermediate package. Intermediate package shall be a box of appropriate size and design, constructed in accordance with PPP-B-636, type CF, class weather - resistant. Boxes shall be closed by taping the top and bottom flaps with 2-inch wide tape conforming to PPP-T-76. Closing of bottom flaps with either adhesive (conforming to MMM-A-250) or metal stitches is permitted.

5.1.4 Level C. Units shall be packaged in standard commercial containers, of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:52Z
Check the source to verify that this is the current version before use.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table III, using Type 1, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant. Closure and strapping shall be in accordance with appendix of applicable box specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table II, using type III, class 1 plywood; or shall be designed for type 2 load and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic. Closure shall be in accordance with appendix of applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers, of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:52Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Style required (see 1.2).
(c) Selection of applicable levels of packaging and packing (See 5.1 and 5.2).

6.2 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet Military requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:52Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:52Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements .

SPECIFICATION

MIL-H-36601 FSN

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD , FORM , JAN 66 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:52Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
MAR 68

EDITION AUG 66, WILL BE USED UNTIL EXHAUSTED

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:52Z
Check the source to verify that this is the current version before use.