# MILITARY SPECIFICATION

## PELVIMETER, MARTIN

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

## 1. SCOPE

1.1 Scope. This specification covers one type of calibrated pelvimeter, Martin design.

## 2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

| | |
|---|---|
| GG-I-526 | Instruments, Dental and Surgical; General Specifications. |
| PPP-B-585 | Boxes, Wood, Wirebound. |
| PPP-B-601 | Boxes, Wood, Cleated-Plywood. |
| PPP-B-621 | Boxes, Wood, Nailed and Lock-Corner. |
| PPP-B-636 | Box, Fiberboard. |

Military

| | |
|---|---|
| MIL-B-116 | Perservation, Methods of. |

STANDARDS

Military

| | |
|---|---|
| MIL-STD-105 | Sampling Procedures and Tables for Inspection by Attributes. |
| MIL-STD-129 | Marking for Shipment and Storage. |



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:54Z
Check the source to verify that this is the current version before use.

DRAWING

Defense Personnel Support Center (DPSC)

10065 Pelvimeter, Martin

(Copies of specifications, standards and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer).

3. REQUIREMENTS

3.1 General requirements. The pelvimeter shall be in accordance with the applicable requirements of GG-I-526, and as specified herein.

3.2 Material. The pelvimeter shall be fabricated from class 4, Type 420, or class 3, Type 303, except for the scale which is Type 304 in accordance with GG-I-526.

3.3 Hardness. After proper stress relieving, the Rockwell hardness for the class 4, type 420 shall be C 40 to C 45 and for the class 3, type 303 shall be B 75 to B 85.

3.4 Style, design and dimensions. The style, design, and dimensions of the pelvimeter shall conform to the Defense Personnel Support Center (DPSC) Drawing No. 10065 , Revision D dated 3 December 1969.

3.5 Finish. Shall be mirror, or mirror and satin in accordance with GG-I-526.

3.6 Corrosion resistance. Shall be in accordance with GG-I-526.

3.7 Identification marking. Shall be in accordance with GG-I-526. Instruments of foreign manufacture shall be marked "stainless."

3.8 Workmanship. Shall be in accordance with GG-I-526.



Source: https://assist.dla.mil -- Downloaded: 202610-21T08:54Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Shall be in accordance with GG-I-526.

4.1.1 Inspection. Shall be in accordance with GG-I-526.

4.1.2 Certificates of quality. Certificates of quality supplied by the manufacturer of the stainless steel may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor or supplier, the FSN, the item identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in Table I. The unit of product for examination shall be one pelvimeter.

Table I. Sampling for Examination

| Type of Defect | Inspection level | AQL (Percent defective) |
|---|---|---|
| Major | II | 1.0 |
| Minor | II | 2.5 |

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in Table II.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:54Z
Check the source to verify that this is the current version before use.

Table II. Sampling for tests

| Component (Unit of product) | Characteristic | Requirement | Test procedure | Inspection level | AQL (percent defective) |
|---|---|---|---|---|---|
| 1-Pelvimeter | Hardness | 3.3 and GG-I-526 | GG-I-526 | S-2 | * |
| 1-Pelvimeter | Copper sulfate test ** | GG-I-526 | GG-I-526 | S-2 | * |
| 1-Pelvimeter | Boil test | GG-I-526 | GG-I-526 | S-2 | * |
| 1-Pelvimeter | Dimensions | 3.4 | -------- | S-1 | * |

* Acceptance number shall be zero for all sample sizes.
** Applicable to martensitic steel only.

4.2.3 Raw materials. Four ounces of each lot of each metal used in construction of the pelvimeter shall be selected for chemical analysis.

4.2.4 Sample unit usage. Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The pelvimeter shall be examined to determine compliance with all specification requirements. Examination shall be conducted in accordance with MIL-STD-105 and the following classification of defects. Examination shall include, but shall not be limited to the defects listed in Table III.

Table III. Classification of Defects

| Categories | Defects |
|---|---|
| Critical | None defined. |
| Major | |
| 101 | Style and design not as specified. |
| 102 | Pelvimeter not complete. |
| 103 | Pelvimeter not free of sharp edges, rough areas, burrs, cracks, grind marks, or overlaps. |
| 104 | Tips not properly aligned. |
| Minor | |
| 201 | Identification marking not present, correct, complete or legible. |
| 202 | Finish not as specified. |



Source: https://assist.dla.mil -- Downloaded: 202610-21T08:54Z
Check the source to verify that this is the current version before use.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements in accordance with GG-I-526.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One pelvimeter, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in the following table shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging Quantities

| Unit package | Intermediate package | Exterior container |
|---|---|---|
| 1 unit | NOT REQUIRED | 24 units |

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Cleaning and drying. Each unit shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

5.1.3.2 Preservative. Not required.

5.1.3.3 Intimate wrap. Each unit shall be wrapped in neutral tissue or other suitable anticorrosive materials.

5.1.3.4 Unit package. Each unit shall be packaged in accordance with MIL-P-116, method IA-15.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:54Z
Check the source to verify that this is the current version before use.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:54Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Intended use. The pelvimeter is intended for calculating and measuring the pelvis to determine size of outlet both internally and externally.

6.2 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Selection of applicable levels of packaging and packing. (see 5.1 and 5.2).
(c) Federal stock number.

6.3 This specification is intended to cover the following item appearing in the Federal Supply Catalog: FSN 6515-354-9800, Pelvimeter, Martin. It does not include all types, sizes or grades as indicated by the title or all types which are commercially available.

6.4 Supersession data. This specification includes the requirements of Purchase Description No. 3 dated 21 November 1961.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preapring activity:
Defense Supply Agency - DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.



Source: https://assist.dla.mil -- Downloaded: 202610-21T08:54Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:54Z
Check the source to verify that this is the current version before use.

# SPECIFICATION ANALYSIS SHEET

Form Approved
Budget, Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements.

SPECIFICATION

MIL-P-36618 FSN:

ORGANIZATION

CITY AND STATE | CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)* | DATE

DD FORM 1 JAN 88 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:54Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

---

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

---

Fold

DPSC FORM 2418
MAR 68

EDITION AUG 66, WILL BE USED UNTIL EXHAUSTED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:54Z
Check the source to verify that this is the current version before use.