MIL-S-36588A
26 June 1974
SUPERSEDING
MIL-S-36588
28 July 1969

MILITARY SPECIFICATION

STETHOSCOPE, COMBINATION TYPE, LIGHTWEIGHT

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers a lightweight, bell-diaphragm, combination stethoscope for diagnostic analysis.

1.2 Classification. Stethoscopes shall be of the following sizes:

Size 1 - Adult
Size 2 - Child

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
QQ-C-320	Chromium Plating.
QQ-N-290	Nickel Plating (Electrodeposited).
PPP-B-585	Boxes, Wood, Wirebound.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:51Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS (cont'd)

Federal

PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.

Military

MIL-P-116	Preservation, Methods of.
MIL-L-10547	Liners, Case and Sheet Overwrap; Water-Vaporproof, Waterproof, Flexible.
MIL-A-36495	Audiometer, General, Diagnostic.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

Drawing No. 22186	EARTIP, MUSHROOM TYPE STETHOSCOPE, LIGHTWEIGHT.

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material.

* 3.1.1 Metal parts. Metal parts shall be fabricated of brass, aluminum alloy, zinc alloy, corrosion-resistant steel, or plain carbon steel suitable for the purpose intended. Matals, other than corrosion-resistant steel, shall be rendered corrosion-resistant by chromium plating over nickel plating.

* 3.1.1.1 Brass. Brass shall be suitable for the purpose intended and comply with the plating and corrosion-resistance requirements of GG-I-526, class 8, types 260, 270, 330 and 360.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:51Z
Check the source to verify that this is the current version before use.

* 3.1.1.2 Aluminum and zinc. Shall be nickel-chromium plated and comply with QQ-C-320, class 1 and type 1.

3.1.1.3 Corrosion-resistant steel. The corrosion-resistant steel shall meet the copper sulfate test requirement as specified in 4.4.4.1 of GG-I-526.

* 3.1.1.4 Carbon steel. The chromium-nickel plated steel shall be of best quality, heat-treated to spring temper, and meet the salt-spray corrosion-resistance requirement of 4.4.4.2 of GG-I-526. As an alternate, the steel spring, if enclosed in the plastic tubing need not be nickel-chromium plated. However, it shall be protectively finished and withstand the same salt-spray corrosion-resistance requirement as nickel-chromium plated steel.

3.1.2 Plastic and rubber parts.

3.1.2.1 Tubing. Tubing shall be made of polyvinyl chloride plastic, and be of one-piece construction, flexible, smooth, seamless, dry, and non-tacky to the touch.

* 3.1.2.2 Ear tips. Plastic for ear tips shall be non-shattering and suitable for the purpose intended. As an alternate, eartips shall be fabricated from soft, oil-resistant rubber or plastic if of the mushroom design (see 3.2.1.2). The plastic or rubber shall be non-toxic and free from objectionable odor.

* 3.1.2.3 Diaphragm. Diaphragm shall be made of plastic or hard rubber and be suitable for the intended function. The diaphragm in the chest piece as well as the two spare diaphragms shall be identical.

* 3.2 Style, design, and dimensions. The stethoscope shall be lightweight, approximately 24 inches in length and weigh 4.6 ounces maximum. The chest-piece shall be a combination type, embodying a bell and diaphragm unit, and shall operate with a two position valve for optional choice of either side as the functional unit. Each stethoscope shall be fully assembled and supplied with two pairs of ear tips, one pair 9/16 inch and one pair 1/2 inch in diameter, and two extra diaphragms.

3.2.1 Binaural section. The binaural section shall consist of two aural tubes, a steel spring and two ear tips.

* 3.2.1.1 Aural tubes. The aural tubes shall be fabricated of austenitic stainless steel or chromium over nickel plated brass tubing. The aural end of each tube shall have external threads for not less than 3/16 inch to receive and securely hold the ear tips specified in 3.2.1.2. The distal ends of the tubes shall have suitable means for securely retaining the plastic tubing specified in 3.2.1.4. The aural tubes shall be shaped so when assembled with the steel spring specified in 3.2.1.3, the assembled tubes shall be comfortable in use. The tubes shall be shaped so that there are no sharp bends or obstructions of the tubes' lumens. The aural tubes shall be rotationally adjustable. The aural tubes shall be secured by a ring at the distal end of the tubes or by a groove near the distal end and with a corresponding depression in the spring to prevent the tubes from slipping out.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:51Z
Check the source to verify that this is the current version before use.

* 3.2.1.2 Ear tips. The bulbular ear tips shall be fabricated of plastic, be a minimum of 1/2 inch long, and be provided with a lumen for the entire length. The lumen shall have well defined internal threads and shall fit securely over the threads of the aural tubes. As an alternate, ear tips shall be of the mushroom design in accordance with DPSC Drawing No. 22186.

3.2.1.3 Spring. The steel spring shall be of corrosion-resistant steel or carbon steel, U-shaped, and between 3 to 6 inches long. Each end of the spring shall be properly shaped for secure attachment to the aural tubes.

3.2.1.4 Tubing. The plastic vinyl tubing shall be Y-shaped, the vertical portion of which shall be at least 13 inches long. All three ends shall make leakproof connections with their attached parts.

3.2.2 Combination chest piece (effectiveness of selector valve). The combination chest piece shall have a two position valve for selecting the bell or diaphragm function and shall substantially reduce all sounds for the chest-piece not selected so that a minimum increase of 25 decibels is required for audible sound to pass through the closed bell when tested as specified in 4.5.2. At least one side of the combination chest piece shall have a mark to indicate to the operator the function selected. As an alternate, an upward curved valve extension may be present in place of a marking. When the valve extension is up, the lower function is operative.

3.2.3 Disphragm chest-piece assembly. The diaphragm chest-piece assembly shall consist of the body, retaining ring and diaphragm.

* 3.2.3.1 Body. The chest-piece body shall be of the combination type, fabricated of brass, aluminum or zinc alloy with a mirror finish chromium over nickel plating. The bell and diaphragm sides shall both be circular with the diaphragm side suitably threaded on the outside to receive the retaining ring specified in 3.2.3.2. The body shall be fitted with the adapter specified in 3.2.2.

3.2.3.2 Retaining ring. The retaining ring shall be suitably threaded on the inside surface to fit the diaphragm side of the body. The ring shall not bind but be easily removed. The diaphragm, as specified in 3.2.3.3, shall be held in place by the inner circumference of the ring.

3.2.3.3 Diaphragm and bell. The diaphragm shall be such as to allow the selective passage of sounds of a frequency of 1,000 to 3,000 cycles per second. (See Note 6.2). The thickness of the diaphragm shall be 0.016 inch maximum. The diameters of the diaphragms and bells shall be as follows:

	Bells	Diaphragms
Size 1	1 inch minimum	1-1/2 to 2 inches
Size 2	1-3/32 inches maximum	1 to 1-1/2 inches

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:51Z
Check the source to verify that this is the currentversion before use.

* 3.2.4 Sound transmission. The sound transmission shall be such that the corrected threshold values in decibels (lowest audiometer volume for audibility) shall be the following maxima when tested as specified in 4.5 and 4.5.1.

Frequency	125	500	1,000	2,000	3,000 cycles/second
Bell	0	5	10	-*	-
Diaphragm	=	=	10	15	35

*Blank spaces indicate absence of requirement.

3.3 Finish. All brass and aluminum components shall be highly polished and be chromium over nickel plated. Carbon steel components shall be suitably protected to comply with test requirements specified herein.

3.4 Identification marking. The chest-piece shall be permanently and legibly marked with the manufacturer's name or registered trademark.

3.5 Workmanship. Workmanship shall be first class throughout. Stethoscopes shall be free from defects which detract from their appearance or may impair their serviceability. Parts which have been internally machined shall be free from burrs.

4. QUALITY ASSURANCE PROVISIONS

* 4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

* 4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:51Z
Check the source to verify that this is the current version before use.

4.2 Sampling.

4.2.1 Unit of product. The unit of product, for inspection purposes only, shall be one stethoscope.

4.2.2 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

The inspection level and acceptable quality level (AQL) shall be as indicated in Table I. Unit of product shall be indicated in 4.2.1.

Table I Sampling for Examination

Component	Type of Defect	Inspection Level	AQL (percent defective)*
Stethoscope	Major	S-1	1.0
	Minor	S-2	2.5

* If acceptance number is not exceeded, lot shall be accepted.

4.2.3 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and Table II. Unit of product shall be as indicated in 4.2.1.

Table II. Sampling for Other Tests

Component	Characteristic	Req.	Test Procedure	Insp. level	AQL
Stethoscope	Sound transmission	3.2.4	4.5 & 4.5.1	S-1	*
	Effectiveness of selector valve	3.2.2	4.5 & 4.5.2	S-1	*
	Copper Sulfate	3.7.1 of GG-I-526b	4.4.4.1 of GGG-I-526b	S-1	*
	Boil Test	3.7.3 of GG-I-526b	4.4.3 of GG-I-526b	S-1	*

* Acceptance number shall be zero for all sample sizes.

4.2.4 Raw materials. Four ounces of each lot of each metal used in construction of the stethoscope shall be selected for chemical analysis.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:51Z
Check the source to verify that this is the current version before use.

4.2.5 Sample unit usage. Where feasible, the same sample unit shall be used for the determination of two, or more test characteristics.

4.3 Inspection procedure. Examination and testing shall be conducted in accordance with Table III, Classification of Defects, in accordance with MIL-STD-105 and with the tests in section 4, to determine compliance with all requirements contained in this specification. Inspection is not restricted to the defects listed below. Any sample unit containing one or more defects shall be rejected (defective) and if the number of defective units in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

Table III Classification of Defects

Categories	Defects
Critical	None defined
Major	
101	Sound transmission inadequate.
102	Selector valve fails to eliminate substantially sounds from the chest-piece not in use.
103	Selector valve not marked or designed to indicate chest-piece selected.
104	Ear tips cracked or damaged.
105	Lumen of ear tips not clear and free from obstructions.
106	Two pairs of ear tips not included.
107	Aural tubes fail to accommodate threaded ear tips for 1/4 inch.
108	Plastic tubing not securely fitted to the binaurals and adapter valve.
109	Diaphragms cracked, damaged or have holes.
110	Material not as specified.
111	Plated surfaces show evidence of flaking, pitting or other imperfections.
112	Retaining ring not easily removed from chest-piece.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:51Z
Check the source to verify that this is the current version before use.

Table III Classification of Defects (cont'd)

Major		
	113	Two extra diaphragms not included.
	114	Aural tubes not dimensioned and shaped as specified.
	115	Burrs present on internally machined surfaces.
Minor		
	151	Chest-piece not permanently and legibly marked.
	152	U-shaped spring not dimensioned as specified.
	153	Ear tips not a minimum 1/2 inch long.
	154	Vertical portion of Y-shaped tubing not at least 13 inches long.
	155	Total length of unit not approximately 24 inches.
	156	Weight not 4.6 ounces or less.

4.4 Tests. All plated parts of the stethoscope shall be tested for plating and corrosion-resistance as follows:

Brass	- GG-I-526, class 8, type 8, 260, 270, or 360
Aluminum	- QQ-C-320, class 1, type 1
Carbon Steel	- GG-I-526, class 1
Stainless Steel	- GG-I-526, class 4

* 4.5 Sound transmission. All sound transmission tests on the stethoscopes shall be conducted in a room with a minimum and constant level of background noise. An audiometer complying with the below-listed requirements of MIL-A-36495 shall be the test instrument:

Receivers	Frequency
Attenuation	Air conduction receivers
Wave form	Tone interrupter

The audiometer is allowed to warm up for 15 minutes and the hearing of the tester shall be established as being normal by comparing his audiogram with audiograms of six other people whose hearing is considered normal. Normal hearing is defined as being no more than 15 decibels (db) above audiometric zero for a

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:51Z
Check the source to verify that this is the currentversion before use.

given frequency. The frequencies at which the tester's hearing is screened are at 125, 500, 1,000, 2,000, and 3,000 cycles per second (cps). Audibility thresholds are determined for each frequency, that is, the minimum db level at which sound is detectable. These threshold values for the tester are run each time just prior to the evaluation of stethoscopes. Audiogram values for the tester shall be run each half-hour in order to compensate primarily for variations in the background noise level. The maximum audiogram values at each frequency, whether obtained with one ear or the other during the testing, shall be used as corrections. Although both earphones are placed over the ears of the tester, only one earphone shall transmit the sound to the tester. The other earphone is kept soundless. The audiogram values are subtracted from the corresponding frequency values obtained on the stethoscopes to give corrected readings.

* 4.5.1 Frequency response characteristics. The operative side of the chest-piece is placed in the center of the output earphone of the audiometer. A pressure on the chest piece just adequate to maintain a hermetic seal is maintained during the testing. The binaurals of the stethoscopes are adjusted for optimum reception by the tester. When the chest piece, especially the bell of the pediatric size, is smaller than the center hole of the foam rubber pad of the audiometer earphone, it will be necessary to use a soft, one-eighth inch layer of foam rubber or plastic with a hole in it just slightly smaller than the bell being tested. Such an adapter gasket provides the necessary hermetic coupling of the bell to the audiometer earphone. Starting with a 125 cps frequency setting the volume of sound is varied until the point is reached at which the sound from the audiometer is just detectable. This is the threshold value. The same procedure is repeated for 500, 1,000, 2,000 and 3,000 cps. The audiogram value for each frequency is subtracted from the value for the corresponding frequency obtained on the stethoscope being tested. The resultant corrected readings are compared with the requirements stated in 3.2.4.

4.5.2 Effectiveness of selector valve. This test shall be run at 125 cps only. The bell side of the chest-piece is shut off. The stethoscope is fitted to the ears of the tester and the binaurals adjusted for optimum hearing. The bell is placed concentrically over the output earphone and an airtight junction established. The diaphragm (on top side) is masked with a sponge or the palm of the hand. The volume of sound is increased until the threshold value is obtained. There shall be a minimum increase of 25 db required to reach the threshold value with the bell shut compared to the threshold value at 125 cps with the bell open. Since this determination is based on the difference in readings, no corrections should be applied. The greater the increase in dbs required to reach the audibility threshold, the greater the effectiveness of the selector valve.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:51Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.4).

5.1.1 Unit of issue. One stethoscope of the type specified, constitutes one unit of issue.

5.1.2 Packaging Quantities. The number of units of issue indicated in Table IV, shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	6 units	48 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Cleaning and drying. Each unit shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

* 5.1.3.2 Preservative. Preservative shall not be applied.

5.1.3.3 Intimate wrap. Each unit shall be wrapped in neutral tissue or other suitable anticorrosive materials.

* 5.1.3.4 Unit package. Each unit shall be packaged in accordance with MIL-P-116, method IC-2. Tubing shall be arranged in such a manner that it will not kink or be permanently deformed.

* 5.1.3.5 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class domestic. Closure shall be adequate to prevent accidental opening.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purposes.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:51Z
Check the source to verify that this is the current version before use.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.4).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PP-B-636, class weather-resistant, grade V3c. Closure and strapping shall be as specified in the appendix of the applicable box specification. Special requirements of applicable table shall apply for PPP-B-636 box.

* 5.2.1.2 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof in accordance with paragraph 30.4 of PPP-B-636.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Fiberboard box shall be closed by method II of PPP-B-636. Special requirements of applicable table shall apply for PPP-B-636 box.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 This specification is intended to cover stethoscopes which are normally procured to meet military requirements. It does not include all types, sizes, etc., as indicated by the title, nor all types which are commercially available.

6.2 Stethoscope application force. The force most commonly applied to the stethoscope chest-piece by the examiner during auscultation is in the range of 40 to 60 grams.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:51Z
Check the source to verify that this is the current version before use.

6.3 This specification covers the following items appearing in the Federal Supply Catalog:

National Stock No.	Size	Item Identification
6515-00-935-4088	1	STETHOSCOPE, Combination Type, Lightweight, Adult
6515-00-935-4089	2	STETHOSCOPE, Combination Type, Lightweight, Child

6.4 Procurement documents should specify the following:

a) Title, number and date of this specification.
b) Size required (see 1.2).
c) Applicable levels of packaging and packing (see 5.1 and 5.2).

* 6.5 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1847

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:51Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
hiladelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:51Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:51Z
Check the source to verify that this is the current version before use.