MIL-T-36606
12 October 1969

MILITARY SPECIFICATION

TABLE, OBSTETRICAL AND GYNECOLOGICAL

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE.

1.1 This specification covers one type of obstetrical and gynecological table suitable for obstetrical delivery and operative work in medical treatment facilities.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-1055	Barrier Material, Waterproofed, Flexible.

Military

MIL-B-26195	Box, Wood-Cleated, Skidded, Load-Bearing Base.
MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

/FSC 6530/

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:53Z
Check the source to verify that this is the current version before use.

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publication. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

National Fire Protection Association (NFPA)

56 Flammable Anesthetics Code

(Application for copies should be addressed to the National Fire Protection Association, 60 Batterymarch Street, Boston, Massachusetts 02110.)

3. REQUIREMENTS

3.1 Materials. Materials shall be as specified herein, and suitable in all respects for the purpose intended.

3.1.1 Conductive rubber sheeting. Conductive rubber sheeting shall be double coated material meeting the requirements of article 2541 of NFPA 56.

3.2 Construction. Table shall consist of a mobile base, pedestal, two-section narrow top, the necessary mechanisms for positioning and control, and the accessories specified herein. The table shall be designed for ease of maintenance and cleaning, and to prevent seepage of oil to the table exterior, or the entrance of liquids or dirt into any of the internal mechanisms. Table shall be electrically conductive in accordance with Article 253 of NFPA 56.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:53Z
Check the source to verify that this is the current version before use.

3.2.1 Base. The base shall be made of 30,000 P. S. I. minimum cast iron, adequately ribbed and so designed to withstand any strain incident to ordinary operation. There shall be located within the base, and at each corner, four ball bearing, swivel, conductive casters not less that 2" in diameter by 1-1/4" wide. The casters shall be placed in mobile position simultaneously by the action of a single foot pedal at the head end of the table. When the table is immobilized, the table shall stand on four rigid conductive elastomer pads, and shall rest evenly and securely on the floor in any position. Casters and pads shall conform to the requirements of Article 253 of NFPA 56.

When supported on casters, the table shall be capable of being moved easily and smoothly. The base shall be so designed and balanced that the table will not tip, regardless of the position of the patient on the table, and shall be firm and rigid in all desired positions. The base shall be covered with a corrosion resisting steel drawn cover to which is attached a steel foot rail located at the foot end of the base. Base cover shall not have any slots, crevices, or other assemblies on top surface that will allow drainage or bacteria to penetrate into base assembly.

Base casting shall incorporate an integral reservoir designed so that the complete hydraulic system can be sealed from any access to dirt or foreign material that could cause contamination of hydraulic fluid thereby resulting in scoring of elevating cylinder or malfunction of valves.

Table will be shipped with the proper charge of hydraulic fluid in reservoir.

Hydraulic components shall be designed so that individual assemblies such as elevating cylinder, pump, and valve mechanism can be easily removed and repaired in field.

The caster pedal and elevation pedals shall be located so that the pedal bars protrude from under the base cover.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:53Z
Check the source to verify that this is the currentversion before use.

3.2.1.1 Elevating mechanism. The pedestal shall consist of a piston type hydraulic cylinder and pump mechanism capable of adjusting the table height from 32 to 43 inches, excluding the mattress pad, with not more than 43 strokes of the foot pedal located toward the head end of the table. The table shall be lowered by fully depressing the pump pedal. When tested with a 300 pound weight as specified in 4.4.2.1, the table shall not settle more than 1/8 inch in eight hours.

The hydraulic elevating cylinder piston and reservoir shall be machined from high quality cast iron free from porosity. The cast iron cylinder shall be machined and honed to provide a smooth surface in order to eliminate binding and wear on the hydraulic polyurethane "U" cup seal. The cast iron pistons shall be machined and smoothly ground to an exact diameter in order to make a precision fit with the I. D. of the cylinder and control the diametral clearance to maintain the hydraulic seal. The piston shall be keyed to the cylinder with hardened steel keys, accurately fitted to prevent rotation of the table top in relation to the base. The hydraulic pump and valve assembly shall be fabricated from aluminum bar stock and located in head end of base. There shall be a simple adjustment available for the length of the pumping stroke and rate of descent. The hydraulic pump, cylinder and actuating mechanism shall be completely concealed by two telescoping cylindrical corrosion resisting steel shields.

3.2.2 Table top. The table top shall be securely mounted on the pedestal and shall be divided into two main sections, an articulating body section and a retractable leg section. The leg section shall retract easily and completely beneath the body section so that the patient's perineum can be positioned precisely at theperineal end of the body section. With foot section extended the table shall measure not less than 73 inches in length.

3.2.2.1 Body section. The body section shall be 22-1/2 inches wide and 45 inches long, designed to accomodate the retractable leg section. The table top shall be capable of being articulated a minimum of 30° Trendelenburg and 25° reverse Trendelenburg by means of folding hand-cranks located on both sides of the table near the head end.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:53Z
Check the source to verify that this is the currentversion before use.

The body section shall consist of an aluminum cast frame accurately machined for pivot bearings, shafts, mounting holes and other elements, and a top formed from 18 United States Standard (U.S.S.) gauge corrosion resisting steel with a 1-1/2 inch hemmed flange on all sides. The center of the top is reinforced with two 18 U.S.S. gauge formed channels spotwelded to the underside. The cushion shall be held in place by four snap fasteners for ease of attachment and removal. The tilting mechanism shall consist of a shaft with a 1 inch -5 rolled Acme screw thread and a hand fitted accurately machined Acme nut fabricated from bearing bronze. The Acme screw shall be driven by a gear train consisting of machined steel bevel gears attached to the Trendelenburg cranks. The gear shaft and Acme screw shall be mounted on anti-friction bearings. Cam follower type ball bearings shall be mounted on the frame casting to provide anti-friction support rollers for the retracting leg section. All gears, shafts, slides, etc. shall be located inside the body section casting. The sides and end of the casting shall be concealed by two formed side panels of polished corrosion resisting steel. There shall be two 3/8 inch x 1-1/8 inch chrome plated steel side rails mounted on the sides of the body section to accomodate the sliding accessories. Shall be designed to prevent the mattress from slipping throughout the table's complete range of operation.

3.2.2.2 Retractable Leg Section: The retractable leg section shall be 16 inches wide by 27-1/2 inches long. It shall be approximately level with the body section when in the extended position and shall be completely concealed in the body section when in the retracted position. The labor position shall be obtained, with one motion; by releasing a safety lock and pulling the leg section out of the body section to the fully extended locked position, level with the top. The delivery postion shall be obtained quickly by releasing the same locking device and pushing the section into the retracted position inside the body section. The leg section shall be capable of being used as a shelf in intermediate positions which are locked in position by a positive stop.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:53Z
Check the source to verify that this is the current version before use.

The leg section shall be formed from 18 U.S.S. gauge corrosion resisting steel with a 1-1/4" hemmed flange on all sides. The center of the top shall be reinforced with an 18 guage channel spotwelded to the underside of the top. The top shall be screwed to a welded frame assembly fabricated from 1-1/4 x 1-1/4 x 3/16 inch steel angle braced in two places. The leg section top shall be guided in sliding carrier frame by four ball bearings. The sliding carrier frame shall consist of two 11 U.S.S. gauge nickel plated die formed plates with a profiled cam slot, that directs the leg section top to drop 4-1/2 inches when the leg section is retracted. The plates are spaced apart by a machined aluminum casting. The leg section assembly shall be mounted on two 11 U.S.S. gauge nickel plated formed channels 44 inches long which engage the four cam follower bearings mounted in the body section casting. The protruding ends of the channels shall be spanned by an aluminum casting. The sides of the leg section shall be concealed by corrosion resisting steel panels. Shall be designed to prevent mattress pad from slipping throughout the table's complete range of operation.

3.2.4 Controls. Controls for retracting and extending the leg section, and extending and retracting the utility shelf shall be located at the foot end of the table and easily operable by personnel standing or seated at that end of the table. The foot pedal control for raising and adjusting the table top, and the one for lowering and retracting floor locks shall be located at the head end of the table within easy reach of personnel seated at that location. A crank handle to place table top in Trendelenburg and reverse Trendelenburg positions shall be located on each side of the head end of the table to permit operation from the head end or either side of the table.

3.3 Maneuverability. When tested as specified in 4.4.2.2, the table shall be capable of being adjusted without perceptible binding or jerking.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:53Z
Check the source to verify that this is the current version before use.

3.4 Rigidity. When table is locked to floor by means of the four corner floor locks, the table shall rest evenly and securely without tilting or tipping. When set in any operating position, the table shall stand solid and remain rigid without evidence of tipping, vibration, slipping, or settling of any component and without noticeable slack or play between the parts or in the articulations. The total side play in the pedestal shall not exceed 1/8 inch, and the total rotational play in the pedestal shall not exceed 1 inch. During the positioning of any part of the table by means of the crank handles or foot controls, such position shall be attained freely and easily without any perceptible jerks, jars, or rotations. Such rigidity shall be maintained while an approximate 300 pound patient is placed on or removed from the table. It shall be tested as specified in 4.4.3.

3.5 Standard accessories. Each table shall be provided with the following accessories which shall be standard with the manufacturer:

(a) Mattresses: The table shall be equipped with a nominal 22-1/2 inch wide, 45" long and 2-1/2 inch thick mattress for the body section and a nominal 22 inch wide, 28" long, 2-1/2 inch thick mattress for the leg section. The mattresses shall be constructed of pin core medium density foam rubber covered with conductive rubber sheeting.

(b) Crutch Sockets: Multiple, fast-acting clamps, combined in a single universal-type crutch socket, simultaneously released or securely locked by one control lever operating in a 90° downward movement, shall be mounted at the end corners of the body carriage to allow for any combination of vertical, rotational, lateral (inward and outward), and longitudinal adjustments of each leg holder post used in obstetrical posturing. Lateral and longitudinal angulation shall be indicated by two permanent, readily legible angulation scales. For easier insertion and removal of crutch rod, socket aperture shall lock at maximum opening when control lever is fully raised.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:53Z
Check the source to verify that this is the current version before use.

(c) Delivery Crutches & Footstep: The knee-rests and foot rests shall be designed to eliminate cutting pressure, proximal and distal to the popliteal area, and to provide the full range of leg posture from extreme high lithotomy to extended Walcher with variable hip-joint flexion. The knee rests and foot rests shall be mounted on hardened steel posts to permit a vertical adjustment of not less than 18 inches. Posts shall be suitably hardened to prevent indentation and marring caused by the locking system. Posts shall be provided with permanent, readily legible height scales. The foot rests shall be foot conforming with curved sides, encasing the instep and shall be adjustable longitudinally and laterally. Three canvas straps with bracelets shall be furnished on knee rests and foot rests.

(d) Wristlets: The table shall be equipped with one pair of wristlets constructed with inside lined with soft buckskin and padded with soft foam rubber. The wristlets shall have a minimum 18 inch long leather strap.

(e) Shoulder Rests: The table shall be equipped with one pair of padded shoulder rests and mounting bars for attachment to side rail sockets on the table top. The shoulder rests shall be form-fitting and made of high strength aluminum alloy, adjustably mounted on chrome plated steel brackets supported in sliding sockets on side rails of body section. Padding should be covered with conductive rubber sheeting. The sliding sockets shall also support the chrome plated anesthesia screen which shall be furnished.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:53Z
Check the source to verify that this is the current version before use.

(f) Traction Handles: The table shall be equipped with one pair of folding and adjustable form-fitting traction handles. The handles shall be cast aluminum, mounted in brass sockets which shall slide on the body section side rails. Slots for the wristlets shall be cast integral with the hand grip assembly. Handles shall be capable of bein lowered below mattress level to allow patient to be transferred to an from stretcher without interference.

3.6 Optional accessories. Table shall be capable of accommodating the following optional accessories, which must be standard with the manufacturer.

(a) A padded intravenous armboard covered with conductive rubber sheeting and suitable for easy attachment to the siderails. Shall adjust through a horizontal arc of not less than 160 degrees by pulling spring loaded finger latch and lock in increments of not greater than 10 degrees. Arm board shall have a minimum one inch thick pad and 7 inch vertical adjusting rod.

(b) A pair of lighweight polished metal lithotomy type leg holders, vertically adjustable, and designed to fit the universally adjustable crutch sockets mounted on the perineal end corners of the body section. Shall be equipped with a snap swivel and strap for supporting patient's foot.

3.7 FINISH:

Chrome Plate: All chrome plated steel parts shall be plated with a minimum of .0006 inch thickness of nickel followed immediately by a minimum of .00001 inch thickness of chrome plate. The finish shall be satin chrome.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:53Z
Check the source to verify that this is the current version before use.

Painted Surfaces:

Fabricated Metal Parts: All exposed fabricated steel parts and assemblies shall be automatically cleaned, degreased, phosphatized, spray primed and baked at high temperature. Final heavy coat of enamel shall be sprayed and baked at high temperature.

Exposed Cast Surfaces: All exposed cast surfaces shall be ground, automatically cleaned, degreased, phosphatized, and spray primed, baked at high temperature, glazed, baked and sanded to a smooth contour. Final heavy coat enamel shall be baked at high temperature.

Stainless Steel: All stainless steel shall be Commercial Standard No. 4 finish.

Aluminum: All exposed aluminum components shall be anodized for white appearance and durability.

3.8 Identification marking. Each table shall be fitted, in a conspicuous location, with a nameplate that shall contain the following data:

Name of manufacturer or registered trade-mark.
Contract or purchase order number.
Federal stock number.
Letters U.S.

Marking secured by adhesives shall not be used.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:53Z
Check the source to verify that this is the current version before use.

3.9 Service data. The contractor shall enclose, with each unit, two copies each of one or more booklets which contain complete instruction for installation, operation, maintenance, and a list of component parts. Wiring diagrams shall be included for electrical items or components. Each part of the unit other than common fasteners, shall be identified in the service data and parts list by the manufacturer's part number. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the contracting officer, via the cognizant Government inspector, two preliminary copies of his proposed service data, for review and comment by the Defense Personnel Support Center. The preliminary service data, shall be submitted sufficiently early to permit adequate review by the Defense Personnel Support Center (based on a maximum period of 30 days for such review)and to allow incorporation by the contractor of required revisions and corrections in the final service data without delaying delivery under the contract or purchase order. In addition, the contractor shall mail 15 copies of the approved service data to the contracting officer for distribution.

3.10 Workmanship. Workmanship shall be first class throughout. Table shall be free from defects which detract from its appearance or may impair its serviceability.

3.10.1 Welding. Welding shall be durable, well fused, sound in every detail and free from blisters or holes. Welding shall be accomplished without any resultant distortion and shall be free from twists or strain. Seam and fillet welds shall be continuous, even and smooth. Spot welds shall be accomplished so as to insure sound welds of sufficient strength to tear the spot out of the parent metal when pulled apart. All exposed welds shall be ground and polished smooth to eliminate traces of weld marks and to blend with adjacent surfaces. All welds shall be cleaned to remove scale and impurities.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:53Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification is defined as both examination (such as visual, or auditory, investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

Table I.

	Inspection level	AQL (percent defective)
For visual examination		
Major defects	II	1.0
Total (major and minor combined) defects	II	4.0
For dimensional examination	S-3	2.5

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:53Z
Check the source to verify that this is the current version before use.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and Table I.

Table II.

Characteristic	Requirement	Test procedure	Inspection level*
Electrical conductivity	3.2	4.4.1	100%
Settling	3.2.1.1	4.4.2.1	S-4
Maneuverability	3.3	4.4.2.2	S-4
Rigidity	3.4	4.4.3	S-4

*Regardless of sample size, acceptance number shall be zero (0).

4.3 Examination. The table shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects:

Categories and defects*

Major

101. Components not properly aligned and fitted.
102. Caster or pad cracked, broken, not securely attached.
103. Metal parts not free of cracks, tears, ruptures, and deformed surfaces.
104. Any weld missing, cracked, broken, or not sound.
105. Any exposed or accessible edge, corner, or surface not free of burrs, sharpness and sharp projections.
106. Table not complete with all required parts.

Minor

201. Finish not uniform; not free of coarse grid marks and scratched areas.
202. Any weld not continuous, not clean, not free of pitting, splatter, and discoloration.
203. Any exposed weld not ground smooth to blend with adjacent surfaces.
204. Unit not clean.
205. Identification marking missing, incomplete, illegible, or not permanently attached.

*Examination shall not be restricted to the classified possible defects listed above.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:53Z
Check the source to verify that this is the current version before use.

4.3.2 Dimensional examination. The tables shall be examined for defects in dimensions. Any dimension not within the tolerance specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Electrical conductivity. The table shall be tested, as specified in Article 253 of NFPA 56, for electrical conductivity twice, once for the casters and once for the floor locks.

4.4.2 Load tests. The table, immobilized on the floor locks, shall be subjected to load tests as follows:

4.4.2.1 Settling. With the table top in the horizontal position and with an evenly distributed 300 pound load, hydraulically raise and lower the table top from 32 to 43 inches. Elevate the table to a maximum elevation and allow to stand for 8 hours. The amount of settling shall then be measured from the the work surface of the table top (excluding mattress) to the floor.

4.4.2.2 Maneuverability. With the table top in the horizontal position, immobilized on floor locks and approximately 37 inches above the floor, supporting an evenly distributed 300 pound load, tilt the table to a minimum of 30 degrees trendelenburg and 25 degrees reverse trendelenburg.

4.4.3 Rigidity. During, and at the completion of the load test in 4.4.2, the table shall be examined for compliance with 3.4. Table shall show no evidence of wear as a result of this test. The table top shall be raised to its maximum height. With the base of the table stationary, the table top shall be pushed to one side as far as it will go, and shall then be displaced to the opposite side. The total displacement shall be measured. With the table top still at maximum elevation, the table top shall be rotated in one direction as far as it will go, and then shall be displaced in the opposite direction. The total displacement shall be measured.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:53Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Level A.

5.1.2 Unit of issue. One (1) table, as specified, constitutes one unit of issue.

5.1.3 Unit preparation. Each unit shall be disassembled as far as is practicable. All exterior finished surfaces shall be protected by wrapping with kraft paper and cushioning with excelsior pads or flexible corrugated fiberboard or other suitable materials. Wrapping and cushioning material shall be secured by tying with cotton cord or tape, or by sealing with kraft gummed paper tape. Work shall be accomplished in such a manner as to insure against marring, scratching, or damage to the unit. The mattress pads shall be protected by wrapping kraft paper or packaged in a carton of suitable size and design. Disassembled parts and accessories shall be adequately protected and packaged in one or more cartons of suitable size and design. Closure shall be adequate to prevent accidental opening.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A. Each unit shall be over boxed in an exterior container designed for a type 3 load and constructed in accordance with MIL-B-26195, type II; PPP-B-601, overseas type; or PPP-B-621, class 2. The unit shall be adequately cushioned and secured within the exterior container by means of bolts, wood braces or other interior fittings. The top, ends and sides of the container shall be lined with a waterproofed barrier conforming to PPP-B-1055, class L-2. Closure and strapping shall be as specified in the appendix of applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:53Z
Check the source to verify that this is the current version before use.

5.2.2 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Interior packages shall be marked to identify contents and quantity contained.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify (a) Title, number and date of this specification, (b) Levels of packaging and packing.

6.2 This specification does not include all types, classes, grades, sizes, of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, which are normally procured to meet Military requirements.

6.3 Table is covered by FSN 6530-462-0837

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:53Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements .

SPECIFICATION MIL-T-36606 FSN

ORGANIZATION

CITY AND STATE | CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)* | DATE

DD FORM 1 JAN 66 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:53Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
MAR 68

EDITION AUG 66, WILL BE USED UNTIL EXHAUSTED

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:53Z
Check the source to verify that this is the current version before use.