MIL-T-36612
AMENDMENT-2
6 April 1974
SUPERSEDING
Amendment-1
6 April 1970

MILITARY SPECIFICATION

TUBE, STOMACH, SURGICAL, PLASTIC RADIOPAQUE, LEVIN

This amendment forms a part of Military Specification MIL-T-36612, dated 29 October 1969, and is approved for use by all Departments and Agencies of the Department of Defense.

Page 4:

Add the following new paragraph:

"3.9 Radiopacity. The tube shall throw a definite, uniform shadow. The degree of opaqueness shall be such that the shadow of a small foreign body in the front or rear of the tube shall be visible for diagnosis."

Delete 4.1 and substitute the following:

* "4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements."

Add the following new paragraph:

* "4.1.3 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate."

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:54Z
Check the source to verify that this is the current version before use.

Page 5:

4.2.2 For tests. Add as the last item in Table III the following information under columns as indicated:

Component (unit of product)	Characteristic	Requirement	Test procedure	Insp. level	AQL (percent defective)
1-Tube	Radiopacity	3.9	4.4.5	S-2	*

* 4.2.3 Raw material. Delete entirely.

Page 7:

Add the following new paragraph:

"4.4.5 Radiopacity test. Each sample tube shall be radiographed as follows: 85 KVP, 10 milliampere seconds, 25 inch distance, 24 centimeters U.S.P. paraffin superimposed over tubes, and using a standard make of cassette with screens, films, and processing solutions. The foreign body used for test purposes shall be a small fragment of sheet lead 1/2 millimeter in thickness."

Page 8:

Add the following new paragraphs:

* "5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraph of appendix thereto. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed in accordance with 30.4 of PPP-B-636.

"NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution."

* 5.2.2.1 Exterior container. At end of paragraph add: "Closure of fiberboard boxes shall conform to method I or II of PPP-B-636. In addition, fiberboard boxes shall conform to the special requirements as specified in Table I of PPP-B-636.

Source: https://assist.dla.mil - Downloaded: 2026-10-21T08:54Z
Check the source to verify that this is the current version before use.

Page 9:

Add the following new paragraph:

"6.4 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue."

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DH

Project No. 6515-1017

Review information is current as of the date of this docume For future coordination of ch to this document, draft circu should be based on the inform in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:54Z
Check the source to verify that this is the current version before use.

A

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 314

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:54Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-T-36512 Amendment-1

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

TYPED NAME OF ACTIVITY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1426 JAN 72

REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:54Z
Check the source to verify that this is the current version before use.