A-A-53925
13 June 1989

COMMERCIAL ITEM DESCRIPTION

SURGICAL SYSTEM, ADVANCED INTRA-ARTICULAR

The General Services Administration has authorized the use of this commercial item description.

The commercial item description covers a Surgical System, advanced arthoscopic suitable for resecting tissue.

Salient characteristics:

Shall be an advanced intra-articular surgical system suitable for use during arthroscopic surgery for resecting tissue within joint spaces. The system shall consist of the following components:

Component	Qty.
Battery Power Unit	1
Battery Charger	1
Footswitch, Two-way	1
Motor Drive Unit	1
4.5 mm cannula	1
4.5 mm trocar	1
4.5 mm obturator	1
3.5 mm cutter blade assembly	3
4.5 mm cutter blade assembly	3
3.5 mm trimmer blade assembly	3
4.5 mm trimmer blade assembly	3
Sterilization case	1
Infusion adapter	1
Cleaning brush	1
Shipping/storage case	1

Battery power unit. The battery power unit shall supply the power to drive the motor unit at a speed of 0-3500 rpm with a minimum torque of 40 ounce-inches.

The power shall be supplied by two in-series 12 volt gel-type lead acid batteries. The batteries shall be capable of being recharged by using a battery charger and be designed to provide a minimum of two years of continuous service when re-charged on a daily basis or a minimum of at least 700 charges.

The batteries shall be encased in a unit measuring 10.25 ± 0.50 inches wide by 5.25 ± 0.50 inches high by 8.55 ± 0.50 inches deep. The complete unit shall weigh less than 20 pounds.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

The front panel of the case shall have a three position function control switch for operating the unit in the ON, OFF, and CHARGING modes. The front panel shall have a meter calibrated for continuous battery condition monitoring and to indicate when the batteries have approximately one hour of normal use remaining. The meter shall also indicate charging condition when the function selector switch is in the charging mode and the batteries are being charged. The front panel shall have a speed adjustment control knob for adjusting motor speeds. The speed adjustment control knob shall offer a wide range of operating speeds starting from 0 to 3500 rpm. The unit shall have a footswitch connector, battery charger connector, and an output jack connector for accepting and securely connecting the motor drive power cable.

Battery charger unit. The battery charger unit shall be small in size, weigh approximetally 4 pounds, and be capable of operating on 120 or 230 VAC, 50/60 Hz and have a leakage current not exceeding 16 or 24 mirco-amperes respectively. It shall have a jack for direct connection into the battery power unit, and a grounded three wire hospital grade power cord of 8 feet minimum length, equipped with a three blade hospital grade connector in accordance with American National Standards Institute (ANSI) Standard C73.11. The connector which connects the charger to the power unit shall be designed such that the other jack plugs on the power unit shall not be physically occluded by the charger unit when attached. The charger unit shall be designed to re-charge the batteries of the power unit in 8 hours or less and shall have an indicator light which illuminates when the charger unit is charging the batteries. The charger unit shall be protected through a slow-blow circuit breaker which is capable of being reset externally. The unit shall be Underwriters Lab (UL) listed.

Footswitch. The footswitch shall have two pedals labeled "forward" and "reverse" which control power delivery to the motor and turn the motor in the clockwise or counter-clockwise directions respectively. The footswitch shall have a cord 6 feet (minimum) in length, with an integral adapter which connects securely to the output jack of the power supply unit. A spare power cord is to be provided with the system.

Motor drive unit. The motor drive unit shall be 1.0 to 1.25 inches in diameter, 6-10 inches in length, and shall weigh less than 2.5 pounds. The external motor drive housing shall have a sandy finish or have several bands of a cross-knurled surface for easy gripping. The housing shall be manufactured of austenitic, corrosion-resistant steel which has been properly hardened and passivated.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

The motor drive unit shall have a detachable power cord, 8 feet (minimum) in length, with an integral adapter which connects securely to the output jack of the power supply unit. A spare power cord is to be provided with the system.

The motor drive unit shall operate at a speed between 0 and 3500 rpm, as selected on either the power supply unit or footswitch. The motor drive unit shall operate in the forward or reverse direction, as selected by the footswitch. The motor drive unit shall deliver 40 ounce-inches (minimum) of torque to the cutting edge of the accessory blade.

The distal end of the motor drive unit shall have a collet mechanism which makes a mechanically secure connection with the blade assemblies. The release mechanism for the collet shall be a sleeve capable of one-handed operation. The collet mechanism shall hold the sheath of the blade assemble stationary and firmly in place.

The motor drive unit with the power cord shall be capable of standing repeated sterilization via or steam at temperatures up to 275°F or long term (10 hour) soaking in a cold sterilization solution.

Safety. All electrical components shall be in compliance with the requirements of Underwriters Laboratories (UL) Standard 544.

Cannulas, trocar, and obturator. The system shall be supplied with two 4.5 mm metal cannulas having an overall length of less than 9 inches with a slightly angulated distal end and a Luer twist type connector on the proximal end for secure cannula attachment. The cannulas shall be designed to accommodate a 3.5 mm and a 4.5 mm cutter and trimmer blade assemblies.

The system shall contain a 4.5 mm metal trocar having an overall length of less than 9 inches that is beveled to form a sharp distal tip for use with the 4.5 mm intra-articular surgery cannula.

The trocar shall fit smoothly into the compatible cannula and shall have the millimeter size or company part number imprinted on it.

The system shall contain a 4.5 mm metal obturator having an overall length of less than 9 inches that is blunt and conical at the distal tip for use with the 4.5 mm intra-articular surgery cannula.

The obturator shall fit smoothly into the compatible cannula and shall have the millimeter size or company part number imprinted on it.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Cutter blade assemblies. The 3.5 mm and 4.5 mm intra-articular cutter blade assemblies shall be in accordance with specification requirements for 3.5 mm and 4.5 mm intra-articular cutter blade assembles respectively (See Note 2). The cutter blade assemblies shall also be a component part of and compatible with an intra-articular arthroscopic surgical system (See Note 2).

Trimmer blade assemblies. The 3.5 mm and 4.5 mm intra-articular trimmer blade assemblies shall be in accordance with specification requirements for 3.5 mm and 4.5 mm athroscopic trimmer blade assembles respectively (See Note 3). The trimmer blade assemblies shall also be a component part of and compatible with an intra-articular arthoscopic surgical system (See Note 3).

Infusion adapter. The infusion adapter shall be in accordance with specification requirements for an infusion adapter (See Note 4).

Cleaning brushes. The system shall include cleaning brusnes approximately 9 inches in length and in sufficient diameters to clean each of the cannulas, sheaths, and blades.

Sterilization tray. The system shall contain a sterilization tray that is fabricated of plastic or a corrosion-resisting steel that is properly hardened and passivated and is suitable for steam sterilization at 275°F or long term (10 hour) soaking in a cold sterilization solution. The tray shall have multiple interior compartments, shaped to the contour of the system components, and shall have handles on each of the outside panels. The top and bottom tray surfaces shall be flat panels that are preforated for maximum sterilization efficiency. The bottom of the tray shall have four integral V-notched metal strips capable of holding in place 12 instruments. The tray shall have a detachable cover with dual spring type locks for securing itself and the V-notched metal strip to the case during handling and sterilization.

Shipping and storage case. The system shall contain a case having multiple padded compartments of sufficient size capable of accomodating the component parts of the intra-articular surgery system. The case shall have a handle for easy transportation, and dual flip-type locks on the front to safeguard the contents.

Reusable container. The system shall be supplied with a reusable container.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Finish. The unit shall be fabricated from materials which are inherently resistant to corrosion or which are treated to render them corrosion-resistant. Where painting or plating is used, appropriate cleaning and priming shall be performed prior to finishing. Unless otherwise specified, finish shall be equivalent or superior to that of comparable commercial equipment and shall be free of gaps, scratches, chips, and other imperfections.

Workmanship. The Advanced Intra-Articular Surgical System shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA) One Advanced Intra-Articular Surgical System unit, as specified, constitutes one unit.

Packaging and labeling. Packaging and labeling shall meet the requirements of the Federal Food, Drug and Cosmetic Act and shall be that usually furnished by the supplier as in commercial practice.

Provisioning. The system shall be provisioned in accordance with applicable requirements.

Electrical characteristics. The electrical characteristics shall be consistent with applicable requirements.

Quality Assurance Provisions.

Responsibility for Inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one Advanced Intra-Articular Surgical System.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE 1: The following National Stock Number is covered by this document:

NSN	ITEM DESCRIPTION
6515-01-208-5978	SURGICAL SYSTEM, ADVANCED INTRA-ARTICULAR

NOTE 2: The Advanced Intra-Articular Surgical System covered by this specification shall consist of a 3.5 mm Intra-Articular Cutter Blade (NSN 6515-01-167-3787) and a 4.5 mm Intra-Articular Cutter Blade (NSN 6515-01-168-3870) which are components of and compatible to an Intra-Articular Arthoscopic Surgical System (NSN 6515-01-168-2603).

NOTE 3: The Advanced Intra-Articular Surgical System covered by this specification shall consist of a 3.5 mm Arthoscopic Trimmer Blade Assembly (NSN 6515-01-168-3779) and a 4.5 mm Arthroscopic Trimmer Blade Assembly (NSN 6515-01-167-3786) which are components of and compatible to an Intra-Articular Arthroscopic Surgical System (NSN 6515-01-168-2603).

NOTE 4: The Advanced Intra-Articular Surgical System covered by this specification shall consist of an Infusion Adapter (NSN 6515-01-165-6711).

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and or order .

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

A-A-53925

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4602

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53925	SURGICAL SYSTEM, ADVANCED INTRA-ARTICULAR

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.