A-A-54080
30 January 1990

COMMERCIAL ITEM DESCRIPTION

CENTRIFUGE, LABORATORY, MICRO-HEMATOCRIT

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers one type of high-speed centrifuge with direct reading capabilities, suitable for use in the micro method of determining hematocrits.

Salient characteristics.

The high-speed centrifuge shall be operable from 120 volts, 50-60 Hz AC supply lines.

The high-speed laboratory centrifuge shall include a housing with hinged safety shield, motor assembly with automatic braking system, two part head assembly, timer, starter switch and power cord. The centrifuge shall be designed to accept at least 24 standard hematocrit capillary tubes up to 75 millimeters in length. The motor driven head shall operate at a fixed speed of 11,700 rpm ± 10 percent and a relative centrifugal force (RCF) of 13,700 G ± 200 G at the rated speed. The head shall incorporate a 24 or more place tray and a hematocrit scale plate. The hematocrit scale plate shall allow for direct reading of all hematocrits after centrifuging without removing either the capillary tubes or the tray from the centrifuge.

The centrifuge shall be equipped with a built-in electric timer capable of timed operation from 1 to 15 minutes in 1 minute intervals. The electric power to the motor shall be controlled by a safety switch and the timer. The safety switch shall prevent the motor from operating when the cover is open. When the motor is operating, a safety latch shall prevent the cover from being opened until centrifuging is completed. An automatic motor brake shall be actuated after completion of the selected spin time and shall stop the rotor within 90 seconds.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

The centrifuge shall be supplied with a power cord terminating with a 3-blade male plug conforming to the applicable requirements of American National Standards Institute (ANSI) Standard C73.11.

Workmanship. The centrifuge shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One laboratory centrifuge, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted on the laboratory centrifuge in accordance with commercial practice. Sampling for inspections and tests shall be in accordance with commercial practice.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

6640-01-098-6638 CENTRIFUGE, LABORATORY, MICRO-HEMATOCRIT

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6640-1581

Location: CID/CENTRIFUGE/S3D/17

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.