A-A-54090
12 February 1990

COMMERCIAL ITEM DESCRIPTION

CATHETER, SPECIMEN COLLECTING UNIT, ENDOTRACHEAL (Delee)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a catheter-specimen collecting unit, of the Delee design, suitable for removing mucus from the upper respiratory tracts of neonates and infants.

Salient characteristics:

General. Shall be a sterile, disposable, 10 Fr suction catheter with a whistle tip. The catheter shall be attached to a 20 cc mucus trap and specimen container. The unit shall be of the Delee design. The specimen container shall be made of transparent plastic and shall be graduated in 2 cc increments from 2 cc to 20 cc.

The unit shall consist of the following components:

Container
Catheter tube
Suction tube
Cap

Material.

Catheter tube. The catheter tubing shall be fabricated from clear flexible plastic and have a matte finish. Lead content shall not exceed 5 ppm.

Suction tube. The suction tubing shall be fabricated from clear, flexible plastic and shall have a matte finish.

Specimen container. The specimen container shall be fabricated from rigid plastic.

Cap. The cap shall be fabricated from clear, flexible or rigid plastic.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Style, design and dimensions. The catheter-specimen collecting unit shall be a graduated plastic specimen container covered by a snap-on type plastic cap at the top end. The plastic cap shall have two circular holes. Into one hole is inserted a plastic, 10 ± 1 Fr whistle tip catheter tube. Into the other hole is inserted a plastic 13 ± 1 Fr suction tube, dilated at the distal end to permit rapid attachment to standard suction devices. The catheter-specimen collecting unit shall be in accordance with DPSC Drawing No. 23429.

Performance. The catheter tube and suction tube shall have a minimum breaking strength of 10 lbs., (20 inches, minimum) when tested as specified herein. The catheter tube and suction tube shall have ultimate elongations of 140% and 175%, minimum, respectively, when tested as specified herein. The catheter-specimen unit shall be shown to be free of foreign matter in vial and lumen tubes when tested as specified herein.

Sterility. The catheter-specimen collecting unit shall be sterile in accordance with the USP or AAMI requirements. Sterility assurance level shall be 1×10^{-6}.

Breaking strength. The breaking strength of the catheter and suction tubing shall be in accordance with American Society for Testing and Materials (ASTM) Designation D412.

Ultimate elongation. The ultimate elongation of the catheter and suction tubing shall be in accordance with ASTM Designation D412.

Class IV plastics. The plastic catheter tubing of the unit shall be in accordance with the United States Pharmacopeia (USP) for Class IV plastics.

Safety and freedom from pyrogens. Samples of the finished catheter-specimen collecting units shall be safe and free from pyrogens in accordance with the USP.

Freedom from foreign matter. The inside of the vial and lumen of tubes the catheter-specimen collecting unit shall be viewed from a distance of 14 ± 1 inches without magnification. The illumination for the units shall be twin, 15 watt, cool white fluorescent tubes.

Workmanship. The catheter-specimen collecting unit shall be processed in such a manner as to be uniform and free from defects which affect life, performance, utility, or appearance.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Package (PG). One package containing 50 individually packaged catheter-specimen collecting units, as specified, constitutes one unit. The individual catheter-specimen collecting unit shall be a sealed, peel open type container capable of maintaining sterility of contents unless package is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one catheter-specimen collecting unit. Defects shall be those listed below.

Defects

Surface of tubes is tacky.
Outside surfaces of catheter tube not smooth and free from burrs or surface imperfections.
Hole in catheter tube at distal end not free from burrs, slivers, loose particles or other foreign matter.
Tubes not straight and free from kinks.
Specimen collecting container cracked, chipped, broken, or crazed.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one catheter-specimen collecting unit. AQL shall be 1.0 (percent defective) with an inspection level of S-2.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-062-7986	CATHETER-SPECIMEN COLLECTING UNIT, ENDOTRACHEAL, Plastic, Disposable, Delee, 50's

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS.

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 47005 Project No. 6515-4267

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54090	CATHETER, SPECIMEN COLLECTING UNIT, ENDOTRACHEAL (DeLee)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use submitting comments and suggestions for improvements. All users of military standardization documents are invited to prov suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request warvers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to warve any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTC
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:28Z
Check the source to verify that this is the current version before use.