A-A-54092
16 February 1990

COMMERCIAL ITEM DESCRIPTION

TUBE, BLOOD COLLECTING (Vacuum with Separator Gel)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers blood collecting tubes with separator gel.

Salient characteristics.

Shall be prelabeled sterile blood collecting tubes with separator gel, having the following aspiration capacities and tube sizes:

Aspiration capacities: 9.5, 13 milliliters (ml)

Tube sizes (diameter X length): 16 X 100, 16 X 125 millimeters (mm)

A commercially designated 10 ml capacity tube size may be offered as an alternate to the specification designated 9.5 ml aspiration capacity tube. Commercially designated 13.5 ml and 15 ml capacity tubes may be offered as alternates to the specification designated 13 ml aspiration capacity tube. A tube size of 16 X 127 mm may be offered as an alternate to the specification designated 16 X 125 mm size.

Shall be suitable for collecting and centrifuging whole blood. The serum separator gel shall form a barrier between the serum and blood cells after centrifuging.

Shall consist of serum separator gel in evacuated transparent glass tubes with diaphragm type self sealing rubber stoppers that form airtight seals and do not come off during centrifuging. The stoppers shall be color coded red or red/grey. The glass shall have sufficient optical clarity to visually inspect a collected blood sample or centrifuged serum separation.

Shall have each glass tube affixed with a label that includes a blank space for use by hospital personnel.

AMSC N/A FSC 6630

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Shall have minimum draws of 8.5 ml and 11.5 ml for the specification designated 9.5 ml and 13 ml aspiration capacity tubes respectively.

Shall comply with the applicable requirements of NCCLS (National Committee for Clinical Standards) Approved Standard ASH-1 (Standard for Evacuated Tubes for Blood Specimen Collection).

Shall be suitable for use with a standard double-ended blood-collecting needle/holder, described in the ASH-1 standard, that contains a bevel edge needle in its open end. When the stopper ends of the tubes are inserted into the needle/holder to the depth even with the needle/holder's reference mark, the needle point shall be embedded in the stoppers without puncturing the stoppers.

Workmanship. The blood collecting tubes shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing one thousand tubes, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Inspection and testing. Inspection and testing shall be conducted in accordance with commercial practice on the blood collecting tubes. Sampling for inspections and testing shall be in accordance with commercial practice.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Other publications. The following documents form a part of this specification to the extent that is specified herein. Unless otherwise indicated, the effect on the date of invitation for bids or request for proposals shall apply.

NATIONAL COMMITTEE FOR CLINICAL STANDARDS (NCCLS)

ASH-1 Standard for Evacuated Tubes for Blood Specimen Collection

(Application for copies should be addressed to National Committee for Clinical Laboratory Standards, Inc., 771 E. Lancaster Ave., Villanova, PA 19085-1519).

NOTE: The following National Stock Numbers are covered by this document:

National Stock Number	Item Identification	Tube Size	Aspiration Capacity
6630-01-066-0416	TUBE, BLOOD COLLECTING With Serum Separator Gel, 16 X 100 MM, 9.5 Ml, 1000's	16 X 100 mm	9.5 ml
6630-01-066-0418	TUBE, BLOOD COLLECTING With Serum Separator Gel, 16 X 125 MM, 13 Ml, 1000's	16 X 125 mm	13 ml

Ordering data. (Expiration dating period required, intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6630-0502

Location: CID/TUBE/BLOOD/S31/17

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.