A-A-54109
26 February 1990

COMMERCIAL ITEM DESCRIPTION

TAG, TISSUE SPECIMEN

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a data tag for use with tissue specimen bags.

Salient characteristics.

Shall be a tear-proof, water-proof and ink smear-proof data tag for use with tissue specimen bags. Shall include a pressure sensitive tape on the backside for attachment to a tissue specimen bag.

Shall be perforated to allow removal of a file copy containing an applied identification number with clinical information while leaving a taped stub portion with an applied matching specimen number on a tissue specimen bag. The perforation shall divide the the data tag into 3 X 5 inch side data filling card with a detachable stub. The backside of the stub shall have a pressure sensitive tape for attachment to a tissue specimen bag. The pressure sensitive tape shall be supplied with a removable protective liner.

Shall include provisions and appropriate legends for a designated specimen number to be marked on both the stub and file card portions of the data tag. The file card shall include provisions and appropriate legends for the following type of data to be marked on the file card:

Patient name
Surgeon name
Specimen/culture site
Diagnosis
Clinical history
Tests requested

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Workmanship. The tissue specimen tag shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing two hundred and fifty tags, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and testing. Inspections and testing shall be conducted on the tissue specimen tag in accordance with commercial practice. Sampling for inspection and testing shall be in accordance with commercial practice.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6640-01-304-6446

The following National Stock Numbers cover the tissue specimen bags intended for use with the subject tissue specimen tag:

National Stock Number	Item Identification
6640-01-086-7020	BAG, TISSUE SPECIMEN, 6 Inches Long X 3 Inches Wide, 250's

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

National Stock Number	Item Identification
6640-01-086-7021	BAG, TISSUE SPECIMEN, 6 Inches Long X 6 Inches Wide, 250's
6640-01-086-7022	BAG, TISSUE SPECIMEN, 8 Inches Long X 7 Inches Wide, 250's
6640-01-086-7023	BAG, TISSUE SPECIMEN, 12 Inches Long X 9 Inches Wide, 250's
6640-01-086-7024	BAG, TISSUE SPECIMEN, 12 Inches Long X 12 Inches Wide, 250's

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6640-1607

Location: CID/TAG/S31/12

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.