A-A-54126
9 April 1990

COMMERCIAL ITEM DESCRIPTION

PIPET-DILUENT, BLOOD, LABORATORY

The General Services Administration has authorized the use of this Commercial Item Description

This Commercial Item Description covers a disposable reagent system for collecting and diluting a small blood sample for purposes of a leukocyte (WBC) count.

Salient Characteristics:

Shall be a disposable blood specimen collection and dilution system for preparing a small sample of whole blood for leukocyte (WBC) count with a hemacytometer (not required to be supplied with this system)

Shall consist of the following:

A translucent plastic reservoir with flexible walls and containing 0.479 milliliters of a 3% acetic acid solution The reservoir shall have a neck end opening that is sealed with a thin plastic diaphragm

A pipet assembly consisting of a 25 microliter self-filling glass capillary pipet that is fitted to a rigid plastic holder that includes an overflow chamber The pipet shall be a straight and thin wall clear glass capillary tube with a uniform bore The proximal end of the holder shall include a protuberance that will make a snug fit when attached to the neck of the reservoir

A bullet shaped shield that fits over the capillary pipet and is also suitable to be used to puncture the diaphragm on the neck of the reservoir

An instruction sheet that includes information and proper procedures for using the blood specimen collection and dilution system

Shall have reservoir supplied in a plastic jar with sufficient moisture to maintain the volume stability of the acetic acid solution diluent. All reservoirs for the unit of issue shall be supplied in a single plastic jar

AMSC N/A FSC 6630

DISTRIBUTION STATEMENT A: Approved for public release distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Shall be designed to collect, dilute and dispense a blood sample in accordance with the following

With the bullet shaped shield attached to the capillary pipet on the pipet assembly, the diaphragm in the neck of the reservoir is penetrated in accordance with directions in the instruction sheet

With the shield removed from the pipet assembly, a whole blood sample is collected from a patient's pricked finger with the pipet in accordance with directions in the instruction sheet

With the pipet assembly seated and secured in the reservoir (with capillary tube penetrating the reservoir), a negative pressure shall be created, in accordance with the directions in the instruction sheet, to draw the blood sample into the diluent within the reservoir

The blood sample and diluent are mixed in accordance with directions in the instruction sheet

The pipet assembly is detached from the reservoir and reassembled to the reservoir in a reversed position to form a dropper assembly for dispensing the diluted blood through the bore of the capillary pipet

A hemacytometer chamber is filled with the diluted blood by squeezing the sides of the reservoir to dispense its contents

Workmanship. The collection and dilution system shall be free from defects which detract from its appearance or impair its serviceability

Unit. Package (PG) One package containing one hundred pipets, as specified, constitutes one unit

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

Inspections and testing. Inspections and testing shall be conducted in accordance with commercial practice on the collection and dilution system. Sampling for inspections and testing shall be in accordance with commercial practice.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Regulatory requirements

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Expiration dating period required, intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order

NOTE The following National Stock Number is covered by this document

NATIONAL STOCK NO.	ITEM IDENTIFICATION
6630-01-068-5256	PIPET-DILUENT, BLOOD, LABORATORY, Disposable, 25 Microliter, 100's

MILITARY INTERESTS

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project Number 6630-0505

Location: ENABLE/PIPETCID/S31/1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER
A-A-54126

2. DOCUMENT TITLE
Pipet-Diluent, Blood, Laboratory

3a. NAME OF SUBMITTING ORGANIZATION

b. ADDRESS (Street, City, State, ZIP Code)

4. TYPE OF ORGANIZATION (Mark one)

- ☐ VENDOR
- ☐ USER
- ☐ MANUFACTURER
- ☐ OTHER (Specify):

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER (Last, First, MI) – Optional

b. WORK TELEPHONE NUMBER (Include Area Code) – Optional

c. MAILING ADDRESS (Street, City, State, ZIP Code) – Optional

8. DATE OF SUBMISSION (YYMMDD)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.