A-A-54127
16 April 1990

COMMERCIAL ITEM DESCRIPTION

TUBE, DRYING, LABORATORY

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a straight form borosilicate glass drying tube for use in laboratory procedures

Salient Characteristics.

Shall be a 150 millimeter (mm) long straight form drying tube with a single bulb

Shall be suitable for use in laboratory procedures

Shall be fabricated from borosilicate glass conforming to Type I and Class A requirements in ASTM E-438 (Standard Specification for Glasses in Laboratory Apparatus)

Shall consist of a glass cylindrical neck tube and a glass stem tube with an intermediate expanded bulb of glass The neck tube, stem tube and expanded bulb shall have the same vertical axis The neck diameter shall be larger than the stem diameter

Shall have the top open end of the neck tube tapered to accommodate a size 0 rubber stopper (not required to be supplied with the drying tube).

Shall have the free end of the tube constricted or serrated or provided with a bulbous enlargement, to accommodate the insertion of a flexible rubber tubing (not required to be supplied with the drying tube)

Shall have the following dimensions

Overall length (exclusive of stem)	146 to 154 mm
Outside neck diameter	16 to 18 mm
Outside stem diameter	6 to 8-1/2 mm
Stem length	22 to 28 mm

Workmanship The drying tube shall be free from defects which detract from its appearance or impair its serviceability

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A Approved for public release, distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Unit Package (PG) One package containing twelve tubes, as specified constitutes one unit

Quality Assurance Provisions

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government s request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice

Inspections and tests Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements

Metric products Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable The contracting officer has the option of accepting or rejecting the product

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented be Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Other publications. The following documents form a part of this specification to the extent that is specified herein. Unless otherwise indicated, the effect on date of invitation for bids or request for proposals shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM E438 STANDARD SPECIFICATION FOR GLASSES IN LABORATORY APPARATUS

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

A-A-54127

(Application for copies should be addressed to American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103)

NOTE The following National Stock Number is covered by this document

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6640-01-308-5345	TUBE, DRYING, LABORATORY, 150 mm., 12's

MILITARY INTERESTS

Custodians

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent

DLA-IS

CIVIL AGENCY COORDINATING ACTIVITIES

VA-OSS
PHS
FDA-MPQAS

Project Number. 6640-1621

Location. ENABLE/CIDTUBE/S31/1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER
A-A-54127

2. DOCUMENT TITLE
Tube, Drying, Laboratory

3a. NAME OF SUBMITTING ORGANIZATION

3b. ADDRESS (Street, City, State, ZIP Code)

4. TYPE OF ORGANIZATION (Mark one)

- ☐ VENDOR
- ☐ USER
- ☐ MANUFACTURER
- ☐ OTHER (Specify):

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording.

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER (Last, First, MI) – Optional

7b. MAILING ADDRESS (Street, City, State, ZIP Code) – Optional

8. WORK TELEPHONE NUMBER (Include Area Code) – Optional

9. DATE OF SUBMISSION (YYMMDD)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil - Downloaded: 2025-10-26T09:29Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required [illegible], was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSIE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.