A-A-54128
16 May 1990

COMMERCIAL ITEM DESCRIPTION

DISPENSER, STRIP, DENTAL

The General Services Administration has authorized the use of this Commerical Item Description as a replacement for Federal Specification GG-D-00441C (DSA-DM) which is cancelled.

This commercial item description covers a plastic dispenser suitable for storing and dispensing into strip lengths, rolls of articulating paper or dental matrix material (see notes).

Salient characteristics:

Material. Dispensers shall be fabricated of molded general purpose polystyrene which is in accordance with ASTM Designation D 4549.

Style, design and dimensions. The dispenser shall conform to Figure 1 for style and design. The dispenser shall be of two-piece construction, consisting of a base and a cover which can be easily loaded, assembled and disassembled. The output channel portion of the dispenser cover shall have an opening or "cutout", (see Figure 1) through which the paper or matrix material may be unrolled and dispensed to the appropriate length. The strip shall be dispensed through the slot formed at the end of the output channel when the dispenser is assembled.

Construction shall be such that material being used in the dispenser shall pass along the outlet channel and through the dispensing end in a free and easy manner, and in the lengths desired. The scalloped outer surface of the dispenser cover, as shown in Figure 1, shall be suitable for hand gripping on the top and sides.

FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:29Z
Check the source to verify that this is the current version before use.

The dispenser shall have the following dimensions:

Characteristic	Dimension (inch)	Tolerance (inch)
Inside diameter of base	1-19/32	± 1/16
Inside diameter of cover	1-3/4	± 1/16
Wall thickness	1/16	± 1/64
Distance from center line of base to outside farthest point of channel	2-1/8	± 1/16
Distance from center line of base to inside farthest point of channel	7/8	± 1/32
Depth of base	2-1/32	± 1/32
Height of assembly	25/32	± 1/32
Width of channel opening at dispensing end	9/16	+ 1/8 -1/16

Finish. Dispensers shall have a smooth finish.

Identification marking. Each dispenser shall be permanently and legibly marked with the name or registered trademark of the manufacturer.

Workmanship. Dispensers shall be free from defects which may detract from their appearance or impair their serviceability.

Unit. Each (EA). One dispenser, as specified, constitutes one unit.

Unit package. Each dispenser shall be individually packaged in a box of appropriate size and design, with an adequate closure to prevent accidental opening during normal shipping and handling.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:29Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the used of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Sampling for inspection shall be performed in accordance with MIL-STD-105, except where otherwise indicated hereinafter.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 at inspection level II, with an AQL of 1.0 (percent defective), except dimensions shall be examined at inspection level S-2 with an AQL of 4.0 (percent defective). The unit of product for examination purposes shall be one dispenser. Dispensers shall be examined for defects including but not limited to those listed below.

Defects.

Dispenser not of two-piece construction.
Parts of dispenser cannot be easily taken apart or assembled.
Finish not as specified.
Style and design not as specified.
Manufacturer's name or registered trademark not present on dispenser.
Packaging, packing, and marking not as specified.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:29Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pounds units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and of regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

National Stock Number	Item Identification
6520-00-523-6050	DISPENSER, STRIP, DENTAL

NOTE 1: The following National Stock Numbers are referenced by this document.

National Stock Number	Item Identification
6520-00-542-3150	PAPER, ARTICULATING, DENTAL
6520-00-935-7172	STRIP, MATRIX, DENTAL

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

AGENT:


DLA-DM

CIVIL AGENCY COOPDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS


47935

PROJECT NO. 6520-2312
LOCATION: NCR - DISK 8


Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.




ASSEMBLY VIEW




BASE-TOP VIEW



COVER-TOP VIEW



BASE-SIDE VIEW



COVER-SIDE VIEW

FIGURE 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.