A-A-54131
17 April 1990

COMMERCIAL ITEM DESCRIPTION

RACK, HEMOSTATIC CLIP (McKenzie)

The General Services Administration has authorized the use of this commercial item description as a replacement for Military Specification MIL-R-36128A which is cancelled

This commercial item description covers a hemostatic clip holding rack suitable for use with McKenzie hemostatic clips and clip introducing forceps during surgical procedures

Salient characteristics:

General The rack shall be capable of accepting and dispensing McKenzie hemostatic clips specifically designed for this application and also be capable of accepting McKenzie style clip-applying forceps specifically designed to pick up and dispense McKenzie hemostatic clips (See Notes 2 and 3

Material The rack and base shall be fabricated from American Iron and Steel Institute (AISI), type 430 ferritic stainless steel The cover shall be fabricated from Class 3 austenitic stainless steel and the screws and machined parts shall be fabricated from Class 3 or Class 4 martensitic stainless steel in accordance with ASTM Designation F899

Hardness The base, rack, and cover shall have a Rockwell hardness B scale (HRB) of 75 minimum and the screws and machined parts shall have a Rockwell Hardness B scale (HRB) of 85 minimum when tested in accordance with ASTM Designation E18 or E92 whichever is applicable Conversion of hardness values shall be in accordance with ASTM Designation E140

Style design and dimensions The rack shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No 20536 for style design and dimensions

Finish The rack shall have a mirror, satin or matte finish The matte finish shall exhibit a uniform nondirectional surface texture equivalent to that produced by lightly blasting the prepared surface with fine abrasives under pressure

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A Approved for public release; distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Passivation. The rack shall be properly passivated

Corrosion resistance The rack shall show no evidence of corrosion or plating of copper when tested in accordance with ASTM Designation F1089

Identification marking and labeling The rack shall be legibly marked with the contractor's name or trademark and the word 'stainless' Foreign manufactured racks shall bear the name of the country of origin

Workmanship The rack shall be free from defects which detract from its appearance or impair its serviceability

Visual and dimensional examination Racks shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements

Materials The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements

Unit Each (EA) One rack as specified, constitutes one unit

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request at any time, or from time to time during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Sampling for examination Sampling for examination shall be conducted in accordance with MIL-STD-105, and as specified herein. Unit of product for examination purposes shall be one rack Sampling shall be inspection level II with an acceptable quality level (AQL) of 1.0

Examination Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one rack Sampling shall be inspection level S-2 with an AQL of 1.0.

Tests Tests shall be conducted to determine compliance with specification requirements Where feasible, the same sample shall be used for the determination of two or more test characteristics

Contractor Certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Metric products Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the lastest revision of Federal Standard No 376, and all other requirements of this document are met

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable

The contracting officer has the option of accepting or rejecting the product

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Regulatory requirements

Federal Food, Drug, and Cosmetic Act If the product covered by this document has been determined by the U S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 as amended, and as implemented by Federal Acquisition Regulation Subpart 23 4 to the maximum extent practical

Preservation packaging, packing, labeling, and marking Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and /or order

NOTE 1: The following National Stock Number is covered by this document

NSN	ITEM IDENTIFICATION
6515-00-356-6000	RAC> HEMOSTATIC CLIP, Mckenzie

NOTE 2 The McKenzie hemostatic clips are covered by National Stock Number (NSN, 6515-00-377-7380, Clip, Hemostatic, Mckenzie, Silver, 100s

NOTE 3 The McKenzie clip-applying forceps is covered by National Stock Number (NSN) 6515-00-321-6900 Forceps Hemostatic Clip Introducing Mckenzie However, this item is not managed by DPSC

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS·

Custodians

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent·

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES·

VA-OSS
PHS
FDA-MPQAS

Project No 6515-4786

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.