A-A-54132
18 April 1990

COMMERCIAL ITEM DESCRIPTION

CLAMP, PYLORUS (Payr)

The General Services Administration has authorized the use of this commercial item description as a replacement for Military Specification MIL-C-36726 which is cancelled

This commercial item description covers a pylorus clamp of the Payr design suitable for holding sections of the pylorus for suturing

Salient characteristics:

General The clamp shall be either 8 inches (Size I) or 11 inches (Size II) long

Material Other than the screws and nuts or pins, the clamp shall be Class 4 corrosion-resistant steel in accordance with ASTM Designation F899 The screws and nuts or pins shall be either class 3 austenitic stainless steel or class 4 steel See Note 2)

Hardness The clamp, other than the screws and nuts, or pins, shall have a Rockwell Hardness scale (HRC) of 40 to 47 When fabricated from Class 3 steel, the screws and nuts or pins shall have a Rockwell Hardness B scale (HRB) of 80 to 100 When fabricated from Class 4 steel, the hardness shall be 5 points on the Rockwell C scale less than for the rest of the clamp

Style, design and dimensions The clamp shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No 20515 for style, design and dimensions

Finish The clamp shall have a mirror satin or matte finish The matte finish shall exhibit a uniform nondirectional surface texture equivalent to that produced by lightly blasting the prepared surface with fine abrasives under pressure

Passivation The clamp shall be properly passivated

Corrosion resistance The clamp shall show no evidence of corrosion or plating of copper when tested in accordance with ASTM Designation F1089

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A Approved for public release, distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Performance The jaws of the clamp shall be uniformly serrated and shall mesh in the closed position The distal tip guide pin shall not bind when engaging the distal tip guide pin hole

Identification marking and labeling Clamps shall be legibly marked with the contractor's name or trademark and the word 'stainless'. Foreign manufactured clamps shall bear the name of the country of origin.

Workmanship The clamp shall be free from defects which detract from its appearance or impair its serviceability

Visual and dimensional examination Clamps shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements

Materials The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements

Unit Each (EA) One clamp of the size specified, constitutes one unit

QUALITY ASSURANCE PROVISIONS

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time or from time to time during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

Sampling for examination Sampling for examination shall be conducted in accordance with MIL-STD-105, and as specified herein Unit of product for examination purposes shall be one clamp Sampling shall be inspection level II with an acceptable quality level (AQL) of 1 0

Examination Examination shall be conducted to determine compliance with specification requirements

Sampling for tests Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein Unit of product for test purposes shall be one clamp Sampling shall be inspection level S-2 with an AQL of 1 0

Tests Tests shall be conducted to determine compliance with specification requirements Where feasible, the same sample shall be used for the determination of two or more test characteristics

Contract certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Metric products Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No 376 and all other requirements of this document are met

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units A request should be made to the contracting officer to determine if the product is acceptable

The contracting officer has the option of accepting or rejecting the product

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Regulatory requirements

Federal Food, Drug, and Cosmetic Act If the product covered by this document has been determined by the U S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23 4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking Unless otherwise specified, preservation packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and /or order

NOTE 1 The following National Stock Numbers are covered by this document

NSN	SIZE	Item Description
6515-00-320-3800	I	Clamp, Pylorus, Payr 8 Inch
6515-00-320-3840	II	Clamp, Pylorus, Payr, 11 Inch

NOTE 2 An acceptable alternate material for the screws and nuts or pins may be one with the following composition

Carbon	17%	max
Manganese	1 5%	max
Sulfur	45%	max
Chromium	18 0%	max
Molybdenum	3%	max optional
Silicon	1 0%	max

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order)

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES.

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4764

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-54132	CLAMP, PYLORUS (Payr)

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE, $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4960 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN. DPSC-RSTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.