
FED. SUP CLASS
6640

USER ACTIVITIES: ARMY - MD, DM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 82





PART NUMBER	DIMENSIONS (cm) INSIDE H	INSIDE W	INSIDE D	OUTSIDE H	OUTSIDE W	OUTSIDE D	VOLTAGE 50/60 Hz AC SINGLE PHASE	WATTAGE
MS36187-1	9.5	10.2	22.9	42	28	46	115	2240
MS36187-2	9.5	10.2	22.9	42	28	46	230	2240
MS36187-3	9.5	14.0	16.5	46	31	37	115	1680
MS36187-4	9.5	10.2	11.4	32	20	22	115	1050
MS36187-5	9.5	10.2	11.4	32	20	22	230	1050
MS36187-6	10.8	12.7	15.2	37	35	28	115	1510
MS36187-7	10.8	12.7	15.2	37	35	28	230	1510
MS36187-8	11.0	11.0	10.0	40	38	51	120	1200
MS36187-9	11.0	24.0	10.0	40	51	51	120	1800
MS36187-10	13.0	25.0	11.0	43	53	55	120	2100
MS36187-11	13.0	38.0	11.0	43	65	55	240	3000

MATERIAL: The furnace shall be constructed of welded steel and high temperature firebrick or alumina-silica composition insulation. The muffle may be steel or inconel; the heating elements shall be heavy gauge nickel-chromium wire.

DESIGN: The furnaces shall have a welded steel case, muffle with direct reading pyrometer with dual scale, continuously variable input adjuster for selection of operating temperature, and chromel-alumel thermocouple. Maximum operating temperature is 1093°C. A safety switch shall be provided to shut the power off upon door opening.

MS36187-1 thru -4 furnaces have a counterbalanced door which opens upward.

MS36187-5 thru -11 furnaces have a door that hinges downward and may serve as a shelf for removal of items from the chamber.

All electrical components shall conform to the standards of Underwriters Laboratories, Incorporated.

NOTE: Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

APPROVED 22 JUN 1983 REVISED

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		FURNACE, MUFFLE, LABORATORY	MS36187
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-1184
STANDARD DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:35Z
Check the source to verify that this is the current version before use.