
FED. SUP CLASS
6640



DIMENSIONS IN INCHES (mm)

PART NUMBER	CAPACITY mL	COMPONENT LANYARD	A DIM. INCHES (mm)
MS36564-1	20	Without	1/8 (3.0)
MS36564-2	20	With	1/8 (3.0)
MS36564-3	20	Without	3/16 (5.0)

MATERIAL: Bulb body shall be made of compound containing pure natural rubber.

Valve shall consist of a nickel-plated brass seat with a hardware-grade steel ball.

Lanyard for MS36564-2 shall be of 18 ply twine in accordance with Military Specification MIL-T-2520, Type I.

DESIGN: Bulb shall be ovoid shape with a check valve at one end and an inlet hole at the other.

Lanyard shall be 24.5±.5 inches (62.2±1.3 cm) long and securely attached to the neck of valve by means of two half hitches.

NOTES:

1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.
2. For design feature purposes, this standard takes precedence over procurement document referenced herein.
3. Referenced documents shall be of the issue in effect on date of invitations for bid.
4. Hole diameter in MS36564-1 and -2 shall provide an easily inserted and snug fit on a 4±.4 mm diameter tube. Hole diameter in MS36564-3 shall provide an easily inserted and snug fit on a 1/4 inch (6.4 mm) diameter detector tube.
5. FSCM: 40912.

(A) ENTIRE STANDARD REVISED

APPROVED 17 SEP 1968 REVISED (A) 17 FEB [illegible]

USER ACTIVITIES: ARMY - MD, DM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - AS
AIR FORCE - 82

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A.	EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust	MS 99		BULB, ASPIRATOR, LABORATORY	MS36564
Procurement Specifications	MIL-B-50015	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 72 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE PROJECT NUMBER 6640-1217

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:43Z
Check the source to verify that this is the current version before use.