RR-P-350B
April 29, 1975
SUPERSEDING
RR-P-00350A(DSA-DM)
May 20, 1968 and
RR-P-350
July 13, 1960

FEDERAL SPECIFICATION

PITCHER, CLINICAL SOLUTIONS

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE

1.1 This specification covers one type of 3 qt. aluminum, nesting pitcher, with a plastic handle.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

Federal Specifications:

QQ-A-250/2 - Aluminum Alloy 3003, Plate and Sheet.
PPP-B-566 - Boxes, Folding, Paperboard.
PPP-B-636 - Boxes, Shipping, Fiberboard.
PPP-B-676 - Boxes, Setup.

Federal Standards:

Fed. Std. No. 123 - Marking for Domestic Shipment (Civil Agencies).

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:09Z
Check the source to verify that this is the current version before use.

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402.

(Single copies of this specification and other Federal Specifications required by activities outside the Federal Government for bidding purposes are available without charge from Business Service Centers at the General Services Administration Regional Offices in Boston, New York, Washington, DC, Atlanta, Chicago, Kansas City, MO, Fort Worth, Denver, San Francisco, Los Angeles, and Seattle, WA.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

Military Specifications:

MIL-A-8625 - Anodic Coatings for Aluminum and Aluminum Alloys.

Military Standards:

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

Defense Personnel Support Center (DPSC) Drawing:

5039 - Pitcher, Clinical Solutions, Aluminum, Nesting, 3 Qt.

(Copies of Military Specifications, Standards, and Drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material.

3.1.1 Aluminum. Aluminum alloy used for fabrication of the pitchers shall be alloy 3003 conforming to QQ-A-250/2 for temper 0 annealed.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:09Z
Check the source to verify that this is the current version before use.

3.1.2 Plastic. Pitcher handle shall be molded from a high-impact type phenolic plastic, capable of withstanding sterilization without distortion, when tested as specified in 4.4.1 (See 6.4).

3.2 Construction. The pitchers shall be constructed in accordance with DPSC Drawing 5039 and as hereinafter specified.

3.2.1 Body. Pitcher body shall be seamlessly formed of a single sheet of 0.064-inch-thick (No. 14, Brown and Sharpe (B. & S.) gage) aluminum, which shall not be reduced below 0.049 inch thick at any point in the finished pitcher. The rim of the pitcher shall be formed with a spout, and the edge shall be rolled into a closed bead all around. Pitchers shall be capable of being nested into each other without jamming. The handle retainers shall be formed as shown on the drawing, of 0.091-inch-thick, (No. 11, B. & S. gage) aluminum. Each retainer shall be securely attached to the body with six spot-welds.

3.2.2 Handle. Handle shall be made of plastic as specified in 3.1.2, and shall conform in shape shown for a typical handle on DPSC Drawing 5039. Handles shall permit stacking of pitchers. Handles shall be secured to the pitcher with two No. 6-32 screwposts. Screwposts and screws shall be chromium-plated steel. As an alternate to that shown in Section A-A, a binding post 7/8 inch long with a 6-32 internal thread shall be acceptable.

3.2.2.1 Alternate plastic handle. Alternate plastic handles, which approximate the general shape and size of the standard handle, will be acceptable provided that plastic conforms to the requirements of 3.1.2. (see 6.4). Approval of handles must be secured from the contracting officer prior to the commencement of production.

3.3 Finish.

3.3.1 Polishing. Pitchers shall be finished to meet the maximum surface roughness values shown on DPSC Drawing 5039. Spotwelds shall be ground smooth.

3.3.2 Protective finish. Pitchers shall be alkaline cleaned and anodized in accordance with MIL-A-8625 for type II, sulfuric-acid coatings and hot-water-sealed.

3.4 Identification marking. The bottom outside surface of each pitcher, as shown on DPSC Drawing 5039, shall be permanently and legibly marked with the manufacturer's name or registered trademark. All markings shall be impressed at even pressure and shall not show on the inside surface of the pitcher.

Source: https://assist.dla.mil -- Downloaded: 2026l0-26T12:09Z
Check the source to verify that this is the currentversion before use.

3.5 Workmanship. Workmanship shall be first class throughout. Pitchers shall be free from defects which detract from their appearance or which may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, and auditory, investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the component or material may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and Table I.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:09Z
Check the source to verify that this is the current version before use.

Table I

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Total defects (Major and Minor combined)	II	4.0
For dimensional examination	S-2	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 for an inspection level of S-3 and an AQL of zero.

4.3 Examination. The pitchers shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with Table II.

Table II. Categories and defects*

Major	
101	Pitcher not complete with component parts.
102	Pitchers fail to nest into each other without jamming.
103	Pitcher body and handle retainers not free of cracks, tears, holes, ripples.
104	Beaded pitcher rim not properly formed and closed.
105	Edges and corners of handle retainers not free of burs and sharpness.
106	Spotwelding of handle retainers to pitcher body not secure.
107	Handle not free of cracks.
108	Handle fails to assemble properly into retainer.
109	Spotwelding not smooth, not free of cracks and burns.
110	Identification marking shows on inside of pitcher body.
Minor	
201	Number and approximate location of spotwelds, securing handle retainers to pitcher body, not as specified.
202	Identification marking incomplete, missing or illegible.
203	Finish not free of pitting and other surface imperfections.
204	Handle not free of chipped areas and other defects.
205	Complete pitcher not free from grease, dirt, and extraneous matter.

* Examination shall not be restricted to the classified possible defects listed above.

Source: https://assist.dla.mil -- Downloaded: 202610-26T12:09Z
Check the source to verify that this is the currentversion before use.

4.3.2 Dimensional examination. The pitchers shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted, as necessary, to determine compliance with specification requirements.

4.4.1 Sterilization test. Samples of the plastic handle shall be autoclaved at 20 pounds per square inch (p.s.i.) for 20 minutes.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C, as specified (See 6.2).

5.1.1 Unit of issue. One pitcher, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table III shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table III

Unit package	Intermediate package	Exterior container
1 unit	Not required	12 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Handle shall be detached. Handle and screws shall be packaged in a sealed envelope of appropriate size. The envelope shall be placed within the pitcher. The pitcher shall be packaged in a box of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676. Closure shall be secured to prevent spilling of contents under normal handling.

Source: https://assist.dla.mil -- Downloaded: 202610-26T12:09Z
Check the source to verify that this is the currentversion before use.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B, or C, as specified (See 6.2).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-636, type CF, class weather-resistant. Closure and strapping shall be in accordance with appendix of applicable box specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-636, type CF, class domestic. Method II closure as specified in PPP-B-636 shall be utilized.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Civilian agencies. In addition to any special marking required by contract or order, interior packages and shipping containers shall be marked in accordance with FED-STD-123.

5.3.2 Military requirements.

5.3.2.1 Unit packages. Each unit package shall be marked in accordance with MIL-STD-129. Contract or purchase order number and name of contractor shall be shown.

5.3.2.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026l0-26T12:09Z
Check the source to verify that this is the currentversion before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:09Z
Check the source to verify that this is the current version before use.