INTERIM COMMERCIAL ITEM DESCRIPTION

CASE, MEDICAL, SHIPPING AND STORAGE, Polyethylene, Rigid

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a waterproof, multi-use, reusable container suitable for use as a medical transit and storage case

Salient characteristics:

Shall be a medical transit and storage case suitable for protection of delicate and sensitive medical devices during transportation and extended storage.

The case shall be one of two sizes constructed in accordance with DPSC drawing number 24539 dated 21 October 1986.

1. Case Size 1 - Nominal inner dimensions shall be 30 inches (76.20 cm) long by 18 inches (45.70 cm) wide by 20 inches (50.80 cm) high.

2. Case Size 2 - Nominal inner dimensions shall be 30 inches (76.20 cm) long by 18 inches (45.70 cm) wide by 13 inches (33.02 cm) high.

Each case shall consist of the following parts:

1. A removable lid which shall be interchangeable with either size case.

2. A base (bottom).

3. An adjustable aluminum frame, affixed to the interior of the case.

AMSC N/A FSC 6545

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

4. Two (2) aluminum partitions that lock into the interior frame.

   a. Case Size 1 - NSN 6515-01-241-7021
   b. Case Size 2 - NSN 6515-01-241-7040

5. Two (2), 1/2 by 28 by 17-1/2 inch (1.27 by 71.12 by 44.45 cm) sheets of 2 pound density polyethylene foam cushioning, NSN 8135-01-251-5465.

6. 1 by 28 by 17-1/2 inch (2.54 by 71.12 by 44.45 cm) sheets of 2 pound density polyethylene foam cushioning, NSN 8135-01-250-0472.

   a. Case Size 1 - Ten (10) sheets.
   b. Case Size 2 - Eight (8) sheets.

7. Three (3), 1 by 28 by 17-1/2 inch (2.54 by 71.12 by 44.45 cm) sheets of 2 pound density convoluted polyurethane foam cushioning, NSN 8135-01-250-0471.

8. Two (2) Packing Procedure Aids.

   Note that a portion of the foam cushioning shall be used in covering the interior of the case as described in following salient characteristics. The remaining foam cushioning shall be shipped loose in the cases to be used in the custom packing of the medical devices.

Lid and base of the case shall be constructed of rotationally molded polyethylene which shall meet the requirements of FEDSPEC L-P-390C for type one, Class M, grade 2, general purpose polyethylene.

The outer case shall be Forest Green, No. 24087, which shall conform to Military Specification MIL-E-52798 (ME) and FED-STD-595A. The case shall have the color pigmentation compounded throughout the depth of the material. The foam shall be either dark gray or black conforming to FED-STD-595A.

Each case shall be designed and constructed to resist solvents. These are both petroleum solvents as well as chemical decontamination solvents and solution.

Shall be designed and constructed to be abrasion resistant. Scratches and scrapes shall not be readily discernible.

Shall be designed and constructed to be easily cleaned with solvents and mild detergent solutions without harmful effects to the case.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

The case shall be marked externally with one aluminum-metal, phototype, or equal label attached to the front right side of the case below the handle. The label shall include the Caduceus medical symbol; the item identification "CASE, MEDICAL, SHIPPING AND STORAGE", the National Stock Number; the nominal inside dimensions; manufacturers part number and manufacturers name.

The case shall have 4 black arrows placed on it to indicate its "up" position on both the front and back in 2 places each.

There shall be no nutrient materials used in the construction of the cases that will support fungal life.

Where unpainted aluminum is used, it shall be protected in accordance with MIL-A-8625, type II, anodic coating.

Cases shall have a tongue and groove closure at the parting line, and be so designed that structural loads shall not be transmitted to the closure gasket. Cases shall have double wall construction at the parting line for additional strength. Pre-formed gaskets that can be readily replaced without the use of special tools, and which are retained in place by the structure, shall be used. One piece gaskets are preferred but joints may be cemented or joined by vulcanizing provided the joints are the same dimensions as any other place on the gasket. Nominal gasket compression shall be 15% to 25%, and shall be sufficient to preclude case leakage, when pressurized 1/2 PSI above the valve pressure. Closure structure shall prevent the compression of the gasket beyond 50%.

Shall permit rapid packing and unpacking without the use of tools by personnel wearing gloves or artic mittens. Each closure fastener shall be 1/4 turn, heavy duty, fastener conforming to the provisions of Military Specification MIL-C-4150H. One each shall be mounted on each end, and four each mounted on each side for a total of ten each. Each device shall be recessed and shall not protrude beyond the outside edge of the case wall when in a locked position. Fasteners shall be constructed of stainless steel, and shall be mounted to the case with stainless steel rivets. Backplates shall be used to strengthen the mounting of the fasteners to the case walls. They shall not interfere with the cover removal when unlatched.

Provisions shall be made to permit field replacement of worn or damaged fasteners with the use of simple common hand tools, for example, drill, blind riveter, etc., without destroying the waterproof seal.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Handles shall conform to Military Standard MIL-STD-1472B and Military Specification MIL-C-4150H. Each case shall be equipped with 6 spring loaded, folding, recessed, stainless steel handles. They shall be mounted 1 at each end and 2 in each side of the base.

Handles shall be spring loaded so as to fold down against the side of the case automatically when not in use. They shall stop open at approximately 90 degrees when extended for use. The grip diameter shall be not less than 1/2 inch (1.27 cm). The clear inside dimension shall be not less than 4-1/2 inches (11.25 cm) in length and 2-1/8 inches (5.31 cm) in depth.

Handles shall be of heavy duty design and made of stainless steel. Handles shall be affixed to the case walls with stainless steel bolts and nuts, and backplates to strengthen attachment. Gasketing material shall insure water tightness.

Provision shall be made to permit field replacement of worn or damaged handles with the use of simple common hand tools: for example, wrenches, pliers, etc., without destroying the waterproof seal.

Each case shall have molded-in male and female positive lock stacking ribs. Four (4) male ribs shall extend above the top surface of the case lid, and four (4) female ribs shall be recessed into the bottom of the base. Size and location of the ribs shall be in strict accordance with the dimensions and tolerances shown on DPSC drawing 24539. The ribs shall provide a stable stacking configuration with a minimum of movement in any direction. Proof of stacking strength shall be determined by Test Method 5016.1 (Superimposed Load Test) Federal Standard 101 C. Cases must be able to support the weight of like cases to a height of 16 feet (4.88 meters).

Each case shall be equipped with a vacuum/pressure relief valve to equalize the air pressure inside the case with that of the outside air. The valve shall be in accordance with Military Specification MIL-V-27166, type 3, with reseal pressures of 5 psig pressure and 2 psig vacuum. The valve shall be installed in the recessed portion of the lid on one side. It shall not protrude beyond the outside edge of the case. The valve assembly shall be replaceable with the use of standard hand tools. The valve shall be of aluminum construction conforming to MIL-V-27166. Aluminum used shall conform to Federal Specification QQ-A-22516. Care shall be taken as not to cover the inside of the PRV with cushioning material which would render it ineffectual.

All fittings shall be self-draining, and shall be attached to the case so that there will be no ingress of water.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Each case shall be equipped with 1 inch (2.54 cm) layer of flexible, 2 pound density, polyethylene foam NSN 8135-01-250-0472, conforming to Federal Specification PPP-C-1752B. This layer of cushioning shall be affixed to the end, side, top and bottom interior surfaces of the case in order to fill all indentations and voids created by the ribs within the vertical and horizontal planes between the case interior surface and the interior framework. This cushioning shall be affixed as a single, squared-off piece to each void and indentation of the case walls with pressure sensitive adhesive in such a manner as to form a secure bond, but not so permanent as to preclude replacement.

Each case shall have an interior framework in accordance with DPSC drawing 24539, constructed of 3/16 inch (0.48 cm) and 1/8 inch (0.32 cm) American Standard Construction grade aluminum affixed to the interior of the case. The framework shall extend from the bottom interior surface to the base of the upper lip of the parting line (vertical dimensions $\pm$ 1/4 inch) and from the interior edge of each side and the interior edge of each end (horizontal dimensions). The framework shall be designed and constructed as to provide compartments in a 1 inch (2.54 cm) increment in the 30 inch (76.2 cm) dimension and will be adjustable upon demand without the use of tools. The interior compartment shall be able to accomodate 33 pounds (14.9 kg) of instrument and be self-supporting. Each compartment shall not have more than 1/16 inch (o.16 cm) change in dimension in any direction for 1 year.

Each case shall be provided with 2 rigid dividers which shall be constructed of 3/16 inch (o.48 cm) thick American Standard construction grade aluminum in accordance with DPSC drawing 24539. These dividers shall be able to be locked into the case interior framework without the use of tools. They shall be adjustable to any 1 inch (2.54 cm) increment in the same manner. The dividers shall be self-locking in all direction and will extend from the full depth of the case to 2 inches (5.08 cm) above the parting line $\pm$.25 inch (0.64 cm).

Each divider shall be equipped with a layer of 1 inch (2.54 cm) flexible, 2 pound density, polyethylene foam, NSN 8135-01-250-0472 affixed with pressure sensitive adhesive to each side of the divider to form a secure bond but not so permanent as to preclude replacement.

In addition to the layer of foam specified previously to fill the indentations and voids created by the ribs, each lid interior shall have the following layers of flexible, 2 pound density polyethylene foam, NSN 8135-01-250-0472, with pressure sensitive adhesive, that conforms to Federal Specification PPP-C-1752B: A 2 inch (5.08 cm)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

layer applied to the top, and a 1 inch (2.54 cm) layer applied to each end. Additionally each lid interior shall have a 1 inch (2.54 cm) layer of 2 pound density, convoluted polyurethane foam, NSN 8135-01-250-0471, that conforms to Military Specification MIL-P-26514E, affixed to the layers of cushioning required above. Mounting shall be accomplished through the use of pressure sensitive adhesive in a manner as to effect a secure bond, but not so permanent as to preclude replacement.

In addition to the layer of foam specified previously to fill the indentations and voids in the base, each base interior shall have a 1 inch (2.54 cm) layer of flexible, 2 pound density polyethylene foam, NSN 8135-01-250-0472, from the full depth of the case to the top of the frame insert at each end. Foam shall be affixed with pressure sensitive adhesive to form a secure bond, but not so permanent as to preclude replacement.

The additional loose sheets of foam cushioning shall be provided with peel-off, pressure sensitive adhesive backing. When applied, this adhesive shall form a secure bond, however, not so permanent as to preclude replacement.

Shall have packed within its interior two (2) packaging procedure aides (PPA). Each PPA shall be printed on 8 1/2" x 11" paper stock and shall be laminated on both sides. The PPA shall recommend procedures to follow to optimize the effectiveness of the case during packing.

Reference standard. Where standard is referenced herein, the issue of that standard in effect on the date of solicitation shall be used to determine compliance with stated requirements.

Workmanship. The medical case shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One medical case, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Responsibility for inspection cont'd. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one medical case.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6545-01-238-5597	Case Size 1 - 30 x 18 x 20 inch
6545-01-238-5598	Case Size 2 - 30 x 18 x 13 inch

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS
FDA-MPQAS

Project: 6545-0087

Location: i.cid.dc.lrb Disk # 4

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements All users of military standardization documents are invited to provide suggestions This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and filed In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-00 5185 8(TM)	CASE, MEDICAL, SHIPPING AND STORAGE

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street, City State ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendat on

6 REMARKS

7a NAME OF SUBMITTER *(Last First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:02Z
Check the source to verify that this is the current version before use.