A-A-54110
5 March 1990

COMMERCIAL ITEM DESCRIPTION

TACHOMETER, STROBOSCOPIC (Portable, Battery Powered)

The General Services Administration has authorized the use of this commercial item description.

his commercial item description covers a stroboscopic tachometer ith digital readout.

alient characteristics.

Shall be a portable photo electric type non-developmental (NDI) battery operated stroboscopic tachometer with digital readout.

Shall be supplied with a photo electric probe that is not an integral part of the stroboscopic tachometer housing. The probe shall be connected to the housing through a minimum 15 inch long retractable cord. The probe shall also be attachable to the stroboscopic tachometer housing to facilitate one hand operation.

Shall contain modularized solid state circuitry that is constructed and assembled in accordance with the following:

Adjustments, that are necessary, can be made without removing any component, printed circuit board, or module.

Allows for use of a circuit board/module extender for making adjustments that may be required.

Shall be designed to be a hand held photo electric type tachometer for measuring rotational parameters with probe located within a range that includes 1/4 to 12 inches from a rotating object.

Shall have design and operational features that include the following:

A measuring range that includes 100 to 100,000 revolutions per minute (rpm).

M' N/A FSC 6680

IS.RIBUTION STATEMENT A. Approved for public release; i.tribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

A liquid crystal display (LCD) readout with at least five (5) digits that are not less than 1/2 inch in height.

An accuracy of ± (.01% + 1 digit).

A crystal controlled time base.

Selector switches/terminals that include suitable markings for function, range and/or test.

All controls and LCD readout included on same front panel.

Shall operate by using any suitable combination of standard 1.5 volt battery cells, or 9 volt transistor type battery cell (s), or combination of both. Note that battery cell (s) are not required to be supplied with the stroboscopic tachometer.

Shall be capable of proper operation in an ambient range that includes 0 to 50 degrees Centigrade (C), and in a 95% relative humidity environment that is non-condensing.

Shall be supplied with a rigid protective carrying and storage case constructed of hard plastic, fiberglass, aluminum, or other similar type material. The case shall include a foam insert that is precut to the shape of the inserted stroboscopic tachometer in place. The overall dimensions of the protective case shall not exceed 10 high X 8 wide X 4 deep (inches). The weight of the empty protective case shall not exceed four (4) pounds.

Workmanship. The stroboscopic tachometer and storage case shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One tachometer, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and testing. Inspections and testing shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Sampling for inspections and testing shall be in accordance with commercial practice.

Tests. The stroboscopic tachometer shall also be tested to determine compliance with the operational and accuracy testing requirements.

Operational test. Each stroboscopic tachometer shall be tested for proper battery powered operation by measuring the speed of a steady rotating object whose speed has been established by suitable means. A tachometer that cannot properly provide a proper reading or distinct digital display shall be cause for rejection.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Accuracy test. A tachometer shall be used to measure the speed of rotating objects in accordance with the following:

a. One measurement taken for a known speed within a range of 100 to 1,000 RPM.
b. One measurement taken for a known speed within a range of 8,000 to 12,000 RPM.

The measured speeds and reference speeds of rotating objects shall be recorded and compared. A measured speed that differs from its referenced speed by more than ± (.01% + 1 digit) shall be cause for rejection. The sampling for testing shall be in accordance with MIL-STD-105, with an inspection level of S-1 and an acceptance number of zero.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6680-01-221-9044	TACHOMETER, STROBOSCOPIC Portable, Battery Powered

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

PREPARING ACTIVITY:

DoD - MB

Custodians:

Army - MD
Navy - MS
Air Force - 03

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

PHS

Project No. 6680-0217

Location: CID/TACHOMETER2/S31/12

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:29Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54110	TACHOMETER, STROBOSCOPIC (Portable, Battery Powered)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:29Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:29Z
Check the source to verify that this is the current version before use.