COMMERCIAL ITEM DESCRIPTION

FIXATION DEVICE, EXTERNAL, ORTHOPEDIC

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a disposable sterile external fixation system for the immediate stabilization and transportation of patients with open, compound bone fractures.

Salient characteristics:

General. The system shall consist of the following items:

QTY	ITEM
1	Fixation Rod, Radiolucent, Composite Material, 350 mm Length, with Protective End Caps
4 or 2	4 Adjustable, Single Pin-To-Rod Clamps or 2 Pivoting Double Pin-To-Rod Clamps
4	5.0 mm Threaded Pin, 165 mm Length, Self Tapping and Self Drilling
1	Combination Pin Driving/Clamp Adjusting Wrench Handle

Rod. Rod shall be fabricated of a radiolucent composite material. The composite shall be graphite/epoxy or carbon fiber/epoxy. The rod shall have a round cross-section and a length of 350 (+/-3)mm. When loaded in 3-point bending, the rod shall have a flexural rigidity of 26 Nm^2 minimum. The diameter of the rod shall be either 8 (+/-.5)mm or 11 (+/-.5)mm and shall be compatible with the adjustable clamps of the system. The surface texture of the rod shall enable the clamps to securely lock to it. The rod shall have protective end caps and shall be capable of being cut by a Gigli saw blade evenly and without frayed sharp edges.

DISTRIBUTION STATEMENT A. Approved for public release, distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Adjustable Single Pin-To-Rod Clamp. The adjustable clamp shall be fabricated of a lightweight composite material or anodized aluminum. The clamp shall be capable of securely locking onto the rod at all points along the length of the rod and at any angle. The clamp shall consist of two nuts, one rod-clamp body, and one pin-clamp body. One nut shall secure the rod-clamp body to the rod. The other nut shall secure the pin-clamp body to the pin. The pin-clamp body shall rotate 360° on the rod-clamp body, independent of the rod-clamp position. The clamp components shall be compatible with the rod and the other instruments.

The clamps shall be numbered and the nuts color coded to denote proper sequence of application as stated in the instructions. The two nuts on each clamp shall be a different color, one from the other. The threaded portion of the clamp shall be notched to prevent complete disengement of the nuts from the clamp body.

Adjustable Double Pin-to-Rod Clamp. The adjustable clamp shall be fabricated of a lightweight composite material or anodized aluminmum. The clamp shall be capable of securely locking onto the rod at all points along the length of the rod and at any angle. The clamp shall consist of two bolts, one rod-clamp body, and one pin-clamp body. One bolt shall secure the rod-clamp body to the rod. The other bolt shall secure the pin-clamp body to the pin. The pin-clamp body shall have a pivoting feature which shall allow independent rotation of each pin through an angle of +/- 19° as a minimum. The clamp components shall be compatible with the rod and the other instruments.

The threaded portion of the clamp shall be notched to prevent complete disengagement of the bolts from the clamp body.

Self-drilling, Self-tapping Pin. Pin shall be fabricated of stainless steel conforming to the composition specified in the latest revision of ASTM Designation F138, Grade 2. Pin shall have a 5.0mm diameter, 165 (+/- 17)mm length, with a maximum thread length of 50mm. Shall be self-tapping and have a drilling point. The flute shall be specifically designed to pull back bone debris, thereby decreasing heat production during insertion. Pin shall not require a predrilled hole for insertion.

Combination wrench. Wrench shall be fabricated of a suitable corrosion resistant metal or metal alloy. The wrench shall be suitable for loosening and tightening the clamp bolts and clamp nuts. Wrench shall also accommodate the 5.0 mm pins, without slipping.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Marking and Surface Preparation. The size of the drill bit, drill sleeve, 5.0 mm pin, and wrench shall be permanently etched on each instrument. Operative ends of the wrench shall be marked to delineate proper application. Additionally, the manufacturer's name (or registered trademark) and manufacturer's reference number shall be marked on drill bit, drill sleeve, pins, and wrench. Surface preparation and marking shall be in accordance with ASTM Designation F86.

Each system shall have instruction for application of the device printed on Tyvek^{tm} paper. The instructions shall include the following:

a) Description
b) Indications
c) Contents of the package
d) Application technique-Shall be step-by-step instructions with professional diagrams of each step in the application of the device

All of the components shall be disposable.

Sterilization. The system shall be capable of being sterilized by steam sterilization if the package is opened but not used.

The system shall be packaged sterile with a sterility assurance level of 10^{-6}. The sterilization date shall be visable on the outer surface of the package. The packaged system shall have a minimum storage life of five (5) years.

The packaged system shall be stackable and have a maximum weight of 25 ounces (710 grams).

Unit. Each (EA). One individually packaged, sterile, external fixation system with instructions for use, as specified, constitutes one unit. The individual package for each system shall be double packaged in a sealed, peel open type container capable of permitting aspetic delivery of the inner container onto a sterile surface.

Workmanship. The external fixation system shall be free from defects.

Visual and dimensional examination. Fixation system shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for Inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and as specified herein. Unit of product for examination purposes shall be one fixation system. Sampling shall be inspection level II with an acceptable quality level (AQL) of 1.0.

Examination. The drills shall be examined to determine compliance with all requirements contained in this specification.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and specified herein. Unit of product for test purposes shall be one fixation system. Sampling shall be inspection level S-2, AQL of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-325-3715	Fixation Device, External, Orthopedic

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No.: 6515-4725

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN DPSC-PSD
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-54296	FIXATION DEVICE, EXTERNAL, ORTHOPEDIC

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street City State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER (Last First MI) - Optional	b WORK TELEPHONE NUMBER (Include Area Code) - Optional
c MAILING ADDRESS (Street City State ZIP Code) - Optional	8 DATE OF SUBMISSION (YYMMDD)

DD FORM 1426 PREVIOUS EDITION IS OBSOLETE

(TO DETACH THIS FORM CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.