COMMERCIAL ITEM DESCRIPTION

AWL, ORAL SURGERY, Wire Passing

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers awls designed for use in passing wires for treatment of facial fractures and preprosthetic surgery.

Salient characteristics:

Classification: The awls covered by this specification shall be of the following types:

No. 17 - Maxillary
No. 18 - Mandibular
No. 19 - Zygomatic

Material: The No. 17 and No. 18 awls shall be fabricated from class 4, type 410x martensitic stainless steel in accordance with ASTM F899.

The No. 19 awl shall be of a two piece construction. The operative end shall be fabricated from class 4, type 440A martensitic stainless steel in accordance with ASTM F899. The handle shall be fabricated from chrome plated brass.

Hardness. The operative end of the awl shall have a Rockwell hardness of C44 to C48 when tested for hardness in accordance with ASTM E18 or E92 and E140.

Style, design and dimensions. The awl shall have a sharp, pointed tip designed for piercing tissue. The wire passing hole shall be located on the centerline of the operative end. The wire passing hole shall be free from burrs and sharp edges. The awl shall be designed for use in passing wires for treatment of facial fractures and prosthetic surgery.

The No. 17 awl shall be designed for use in passing maxillary wires. The No. 17 awl style, design and dimensions shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 22428.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

The No. 18 awl shall be designed for use in passing circum-mandibular wires. The style, design and dimensions of the No. 18 awl shall be in accordance with DPSC Drawing No. 22429.

The No. 19 awl shall be designed for use in passing circum-zygomatic wires. The style, design and dimensions of the No. 19 awl shall be in accordance with DPSC Drawing No. 22430.

Finish. The awl shall have a satin or mirror finish.

Passivation. The awl shall be properly passivated.

Marking. The awl shall be marked with the name or registered trademark of the manufacturer, the awl number, and the word "stainless". If the awl is of foreign manufacture, the instrument shall bear the name of the country of manufacture.

Corrosion resistance. The awl shall show no signs of corrosion when tested in accordance with the copper sulfate and boil test requirements of ASTM F1089.

Workmanship. The awl shall be free from defects which detract from its appearance or impair is serviceability.

Unit. Each (EA). One awl of the number specified, constitutes one unit.

Unit package. The awl shall be packaged in a container designed to prevent damage during shipment or storage.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

Certificates of quality. Certificates of quality, supplied by the manufacturer of the component or materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, inspection level of II, with an AQL of 2.5 (percent defective) except for dimensions which shall be conducted at inspection level S-1, with an AQL of 1.0 (percent defective). The unit of product for examination purposes shall be one awl as specified in this description.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, with an inspection level of S-2 and an AQL of 1.0 (percent defective). The unit of product for test purposes shall be one awl.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pounds units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be a specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document:

National Stock Number	Item Identification
6520-01-058-2534	AWL, ORAL SURGERY, Wire Passing, Maxillary, No. 17
6520-01-058-2535	AWL, ORAL SURGERY, Wire Passing, Mandibular, No 18
6520-01-058-2536	AWL, ORAL SURGERY, Wire Passing, Zygomatic, No. 19

Ordering data. Intermediate/exterior package quantities, labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6520-2380
LOCATION: ML WORD/AWLORAL

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC- RST
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54299	AWL, ORAL SURGERY, Wire Passing

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.