A-A-54312
21 January 1991

COMMERCIAL ITEM DESCRIPTION

CASE, DENTAL INSTRUMENTS (Sterilizing and Storage)

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for MIL-C-36260B which is cancelled.

This Commercial Item Description covers a compartmented dental instrument case suitable for holding and autoclaving endodontic instruments, and maintaining them under aseptic conditions.

Salient characteristics:

Material. The dental instrument case shall be fabricated of AISI type 304 stainless steel, except the hinges may be of type 302 or 304 stainless steel in accordance with ASTM F899 for chemical composition.

Style, design and dimensions. The dental case shall be divided into 68 small compartments and one large compartment. The cover shall be fastened to the body of the case by two piano hinges. When closed, the cover shall overlap the base on all four sides by approximately 3/8 inch.

The latched cover shall provide a baffle type closure to prevent exposure of the sterilized instruments to ambient air.

The cover shall be provided with a swiveled, "U" shaped Aberdeen barrel latch, designed such that it may be positioned to hold the cover partially open during autoclaving, to permit the flow of steam to all instruments. The hinges shall permit the cover and the base to mate with a snug fit which allows proper alignment of the latch for opening and closing. All components of the case shall fit into their respective positions without the use of excessive force, and shall be properly aligned and secured. Components such as hinges and latches shall operate smoothly. The cover shall have a flush type fit when closed. The style, design and dimensions of the dental instrument case shall be in accordance with Defense Personnel Support Center (DPSC) Drawing #23377.

Autoclavability. The dental instrument case, in its entirety, shall be properly passivated and shall show no signs of corrosion after being subjected to 10 autoclave cycles at 270° + 5°F and 27 + 1 psi for a minimum of 20 minutes each cycle.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Finish. The dental instrument case shall have a smooth satin finish. All finishes shall be uniform except for necessary sound weld marks on the bottom surface of the case. The rivets, hinges and latches shall have a uniform satin or polished finish.

Identification marking. Each dental instrument case shall be permanently and legibly stamped or etched with the name or registered trademark of the manufacturer.

Workmanship. All surfaces (including edges and corners) shall be free from sharp edges, distortion, or dents. The case shall be free of welding and solder marks. The case shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One dental instrument case, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:32Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, at inspection level S-2, with an AQL of 2.5 (percent defective). The unit of product for examination purposes shall be one instrument case as specified herein.

Examination. The complete dental case shall be examined to determine compliance with all requirements contained herein.

Sampling for tests. Sampling for tests (dimensional), shall be conducted in accordance with MIL-STD-105, at an inspection level of S-1 and the acceptance number shall be zero for all sample sizes. The unit of product for test purposes shall be one case.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and there-after as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered Materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

National Stock Number	Item Identification
6520-00-890-1883	CASE, DENTAL INSTRUMENTS, (Sterilizing and Storage)

Ordering data. Intermediate/extermediate package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:32Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project Number: 6520-2354
Location: WORD/CASE/DENTAL

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-MSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54312	CASE, DENTAL INSTRUMENTS (Sterilizing and Storage)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording.

b. Recommended Wording.

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.