A-A-54325
18 January 1991

COMMERCIAL ITEM DESCRIPTION

AIRWAY, PHARYNGEAL, RUBBER (Guedel)

The General Services Administration has authorized the use of this commercial item description as a replacement for Military Specification MIL-A-36163 (BUMED) which is cancelled.

This commercial item description covers a rubber pharyngeal airway of the Guedel type suitable for use in anesthesia procedures.

Salient characteristics:

General. The airway shall be a curved, clean, ready for use, reusable, rubber pharyngeal tube and shall conform to American Standards Institute (ANSI) Standard Z79.3 for Anesthetic Equipment-Oropharyngeal and Nasopharyngeal airways.

Material. The airway shall be fabricated from medical grade rubber in accordance with ANSI Z79.3. The insert shall be metal or nylon as specified.

Style, design, and dimensions. The airway shall be furnished in the following sizes:

Size 0 - Infant
Size 2 - Child
Size 3 - Small Adult
Size 5 - Large Adult

The airway shall be in accordance with Defense Personnel Support Center (DPSC) No. 20065.

Finish. The exterior and interior surfaces shall be smooth for an easy and non-grabbing insertion.

Toxicity and biocompatibility. Toxicity and biocompatibility shall be in accordance with ANSI Z79.3.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Certificates of biocompatibility. A certificate of compliance, stipulating that offered items meet the biocompatibility test, shall be acceptable, provided that the contractor has supplied acceptable items to the Government within the past three years and that there been no changes to material, formulation, and method of processing since the actual testing was performed.

The certificate shall include the date the testing was performed, the test method and results, the name and title of the person evaluating the results, the current contract number and NSN, a statement certifying that material, processing, and formulation have not changed since the time when biocompatibility was performed and the signature and title of the person making the certification.

Workmanship. The airway shall be free from defects which detract from its appearance or impair its serviceability.

PRODUCT QUALITY LEVEL.

Visual and dimensional examination. Shall be examined to verify that the design, construction and workmanship are in accordance with contractual requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Each (EA). One airway of the size specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specifications where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by of for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory applicants or procedures) and testing (determination by technical means and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105. Inspection shall be conducted at Inspection level II and an AQL (Acceptable Quality Level) of 1.0 percent. The unit of product for test purposes shall be one airway.

Contract certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings , specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the products is acceptable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply , and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and /or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	SIZE
6515-00-299-8749	AIRWAY, PHARYNGEAL, Rubber, Guedel, Infant	0
6515-00-300-2900	AIRWAY, PHARYNGEAL, Rubber, Guedel, Large Adult	5
6515-00-299-8748	AIRWAY, PHARYNGEAL, Rubber, Guedel, Child	2
6515-00-300-2910	AIRWAY, PHARYNGEAL, Rubber, Guedel, Small Adult	3

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4910

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.