A-A-54334
4 February 1991

COMMERCIAL ITEM DESCRIPTION

CELL WASHING SYSTEM

The General Services Administration has authorized the use of this Commercial Item Description

This Commercial Item Description covers a blood cell washing system for deglycerolizing thawed red blood cells.

Salient Characteristics:

Shall be a blood cell washing system for deglycerolizing thawed red blood cells by a sequential dilution method. Shall be a gravity fed continuous flow type red blood cell washing system suitable for high glycerol freezing protocols

Shall be suitable for operation from 50/60 Hz AC supply lines The applicable rated line voltage shall be as indicated in the contract and/or purchase order (see "Ordering data" paragraph).

Shall include a cabinet housing with a pole stand. The pole stand shall have a 45 inch minimum height and shall include four (4) height adjustable support hooks that are suitable for suspending plastic type IV solution bags.

Shall have the cabinet housing designed and constructed to include the following:

A fixed speed centrifuge designed to accommodate an inserted washbowl that is a component of the washbowl/waste bag set The centrifuge shall operate at a speed of 5850 rpm ± 125 rpm

A top mounted oscillating mixing platform with four (4) adjustable magnetic mounting posts that include mounting pins The mixing platform shall operate with a oscillation rate of 180 cycles/minute ± 3 cycles/minute The magnetic mounting posts shall be adjustable for securing a primary collection bag (i e unit containing glycerolized red blood cells) to the mixing platform

Electrical controls for activating and deactivating the mixing platform and centrifuge

AMSC N/A FSC 6540

DISTRIBUTION STATEMENT A. Approved for public release, distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

A cooling fan that operates whenever the mixing platform and centrifuge are in operation

A 30 minute mechanical timer with audio alarm that is graduated in one (1) minute intervals The timer shall be used for indicating when a preset time had been reached, without controlling any component of the blood cell washing system

Mounting pins that are secured to the side of the housing, for mounting a 6000 milliliter (ml) capacity waste bag that is a component of the washbowl/waste bag set.

Suitable fittings for supporting an inserted pole stand in an upright position.

Shall be suitable for use with a disposable plastic washbowl/waste bag set (not required to be supplied with the blood cell washer) with the following characteristics:

Consists of a centrifuge bowl with inlet and outlet port connectors, an integrally attached cell wash harness, and a detachable 6000 ml capacity waste bag

The cell wash harness consists of various lengths of plastic tubing, four (4) "Y" type connectors, and a four (4) color spike harness (i e with a yellow, a blue, and 2 red color coded spike terminations).

The yellow color spike intended for connection to a 3000 ml capacity container with a 0.9% saline/0.2% dextrose solution

The blue color spike intended for connection to a 150 ml capacity container with a 12% saline solution.

The red color spikes intended for connection to two (2) primary collection bags containing glycerolized red blood cells

The wash solution perfuses the red blood cells in the spinning washbowl, and then exits separately through the outlet port of the washbowl into the waste bag for disposal

The washed unit of red blood cells is siphoned from the washbowl into a blood recovery pack that is suspended from a support hook on the pole stand

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Workmanship. The blood cell washer shall be free from defects which detract from its appearance or impair its serviceability

Unit. Each (EA). One blood cell washer, as specified, constitutes one unit

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

Inspections and testing Inspections and testing shall be conducted on the blood cell washer in accordance with commercial practice Sampling for the inspections and testing shall be in accordance with commercial practice In addition, each blood cell washer shall be tested to demonstrate its capability for proper operation The testing shall check for the proper operation of the electric controls, centrifuge, mixing platform and cooling fan. Any component that does not operate properly shall be cause for rejection of the blood cell washer.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

NOTE The following National Stock Numbers are covered by this document

6640-01-235-6131

6640-01-315-8452

Ordering data (Applicable line voltage intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order)

MILITARY INTERESTS.

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY

DoD-MB

Agent.

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project Number 6640-1651

Location ENABLE/BLOODCID/S31

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4866 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54334	BLOOD CELL WASHER

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.