A-A-54335
4 February 1991

COMMERCIAL ITEM DESCRIPTION

PUMP, CIRCULATING

The General Services Administration has authorized the use of this Commercial Item Description in preference to Military Specification MIL-P-42063 (DM) which was cancelled.

This Commercial Item Description covers an electrical operated circulating impeller type pump for use in a water bath to thaw frozen blood units.

Salient Characteristics.

Shall be a single speed or a dual speed circulating impeller type water pump, for use in a temperature controlled water bath to thaw frozen blood units. Shall provide a horizontal water discharge.

Shall be suitable for electrical operation from 50/60 Hz AC supply lines. A 5/6 reduction in motor speed on a 50 Hz supply line from that on a 60 Hz supply line shall be acceptable. The applicable line voltage shall be as indicated in the contract and/or order (see "Ordering data" paragraph)

Shall be supplied with a clamp with a 1-1/4 inch wide jaw for attaching to the side of a water bath.

Shall have a nominal overall height of 14 inches with a submersible range level that includes 3 to 7 inches

Shall, if supplied with a single speed water pump, be rated for a delivery capacity of 10 liters per minute The single speed water pump shall be electrically operated by a thermally protected motor that is rated for 1550 revolutions per minute (rpm) at no load.

Shall, if supplied with a dual speed water pump, be rated for delivery capacities of 7 and 15 liters per minute The dual speed water pump shall be electrically operated by a thermally protected motor that is rated for 3000 to 3400 rpm at the 15 liters per minute pumping delivery mode

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Shall be constructed with a top mounted motor assembly and a bottom mounting terminating with a base that encloses impeller type blades. A vertical shaft shall connect the motor shaft to the impeller type blades, and shall be enclosed in a corrosion-resisting steel housing

Shall include a corrosion-resistant metal top mounted housing for the motor assembly and controls, and a corrosion-resisting steel housing for the submersible parts The top mounted housing shall include a non-detachable clamp for attaching subject circulating pump to the side of a water bath.

Shall at the option of the supplier be supplied with or without an integral electrical operated heater assembly.

Workmanship. The circulating pump shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA) One pump, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Inspections and testing. Inspections and testing shall be conducted in accordance with commercial practice. Sampling for the inspections and testing shall be in accordance with commercial practice. In addition, each circulating pump shall be tested to demonstrate its capability for proper operation.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order

NOTE: The following National Stock Numbers are covered by this document.

6640-01-235-1378

6640-01-301-8165

Ordering data. (Applicable AC line voltage intermediate/ exterior package quantities, labeling, and marking must be specified in the contract and/or order.)

MILITARY INTERESTS

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES·

VA-OSS
PHS
FDA-MPQAS

Project Number 6640-1663

Location· Enable/PUMPCID/S31

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4066 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnal Support Center
Attn: DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54335	PUMP, CIRCULATING

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION (Mark one)
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER (Specify): _____
b. ADDRESS (Street, City, State, ZIP Code)	

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER (Last, First, MI) – Optional	b. WORK TELEPHONE NUMBER (Include Area Code) – Optional
c. MAILING ADDRESS (Street, City, State, ZIP Code) – Optional	8. DATE OF SUBMISSION (YYMMDD)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.