A-A-54339
14 February 1991

COMMERCIAL ITEM DESCRIPTION

SURGICAL PACK, HEAD AND NECK, DISPOSABLE

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a sterile, disposable surgical pack containing the components necessary to drape a patient and the operative field for various types of head and neck operative procedures.

Salient characteristics:

The head and neck pack shall contain the following components with the dimensions indicated. The components shall be manufactured from a nonwoven fabric, which, except for the gown wrapper and towel, is fluid repellent (treated to be resistant to penetration from solutions, oil, alcohol and surfactants). The material shall meet the surface resistivity requirements of the National Fire Protective Association. The nonwoven fabric shall also resist tearing, pilling, and linting.

The components of the pack shall include the following:

Overwrap. The overwrap shall be a minimum of 44 by 85 inches, fan-folded for ease in opening, impervious to fluids and suitable for use as a cover for the surgical back table. A 26 inch wide absorbent nonwoven paper inner-liner shall be secured with adhesive to the full length of the overwrap.

Head Drape. The head drape shall be made of a soft material with minimal memory or spring-back qualities, measuring a minimum of 26 by 42 inches, designed with a 1 by 26 inch double-faced, pressure sensitive tape for securing the drape to the patient's forehead. The head drape shall have an impervious or very high barrier coated piece of nonwoven fabric measuring 10-1/2 by 33 inches located on the opposite side of the drape, with one long edge wrapped around 1/2 inch and caught under the pressure sensitive tape area, suitably

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

secured with adhesive at the two narrow ends and leaving a cuff 6 inches minimum deep to minimize contamination during placement. The entire head drape shall be resistant to penetration by fluids and dry particles. The reinforced area shall be impervious to fluids yet highly absorbent, eliminating strike-through while preventing pooling of fluids. The head drape shall have adequate directions for proper placement clearly imprinted thereon.

Split Sheet. The split sheet shall be made of a soft nonwoven fabric with draping qualities like linen, with minimal memory or spring-back qualities, measuring a minimum of 72 by 114 inches, designed with a "U" shaped double-faced pressure sensitive tape on the patient side of the sheet and the inner edges of the split. The split shall measure a minimum of 2-1/2 inches wide by 20 inches long and be located at one end of the sheet. The tape shall facilitate securing the split sheet to the patient. The split sheet shall have a minimum 18 by 20 inch impervious coated piece of nonwoven fabric reinforcement located around the pressure sensitive tape area, suitably secured with adhesive on the operative side of the sheet. The split sheet shall have a two-hole tubing tab located on the bottom portion of the reinforced area below the split. The tab shall allow securing of suction tubing, and active electrodes without the use of towel clips. The reinforced area shall be impervious to fluids, eliminating strike-through while preventing pooling of fluids. The split sheet shall be fan-folded to facilitate placement on the patient and shall have adequate directions for proper placement clearly imprinted thereon.

Gowns. The pack shall include two large size gowns made of a soft, comfortable, pliable, and quiet nonwoven fabric, resistant to fluid and dry particle penetration, with high breathability, facilitating comfort when worn for long periods of time. The gowns shall be large enough to cover the entire back of the wearer and contain stockinette cuffs conforming to the wrists, and two sets of a suitable tie closure that will not rip, or separate from the gown when being tied. One set shall be inside and one set outside the gown. The ties shall be designed to minimize the possibility of contamination when being tied. One of the gowns and one of the towels shall be individually packaged in a light plastic or polypropylene wrapper, with minimal memory for ease of opening, and be located in the top of the overwrap for use by the scrub nurse.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Mayo stand cover. The Mayo stand cover shall measure a minimum of 23 by 55 inches, and shall be made of a plastic material that is impervious to fluids. The cover shall be reinforced with a minimum 20 by 24 inch piece of nonwoven fabric pad, laminated to the area that will cover the Mayo stand tray. The pad shall be suitable to absorb the heat and moisture of the instruments.

Towels. The pack shall contain at least one towel measuring a minimum of 15 by 20 inches, and shall be highly absorbent (not fluid repellent), so that the towel will dry two arms following a surgical scrub.

Suture bag. Each pack shall contain a plastic, paper, or nonwoven fabric suture bag.

Pack. Each pack shall contain the components described above, individually wrapped in a suitable tear resistant, heat sealed package which will allow aseptic handling, and maintain sterility of the contents unless opened. Each pack may also be wrapped in an exterior dust cover with all components clearly listed, along with the sterility lot number, date of pack and the manufacturer's name. The pack assembly from the top down shall be: towel on top of overwrapped scrub nurse gown, Mayo stand cover and surgeon's gown, suture bag, head drape, split sheet. Each pack component shall be identified with the identification visible as the component is removed from the pack. Each component shall be clean and contain no foreign matter or tears.

Sterility. The surgical packs shall be sterile.

Workmanship. The surgical packs shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing five or six individually packaged sterile surgical packs, as specified, constitutes one unit. Each surgical pack shall be packaged sterile in a tear-resistant heat-sealed package which allows aseptic handling and maintains sterility of the contents until opened or damaged.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one surgical pack as specified in this description.

Sampling for tests. Sampling for tests (dimensional) shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one surgical pack.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include, but shall not be limited to the following:

Sterility test. The surgical packs shall be tested for sterility.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Surface resistivity or static decay test. Testing shall be conducted as described in NFPA-99 (National Fire Codes). A copy of which is available from the National Fire Protection Association, 470 Atlantic Avenue, Boston, MA 02110.

Water repellency test.

Principle of method: The fabric to be tested is used to seal an inverted vented mason jar containing a saline solution and the penetration of the solution through the fabric noted.

Apparatus:

a. 2-11/16 inch diameter circular cutting die.

b. Mason jars (1 qt.) with holes drilled in the bottom and metal rings with openings 2-3/16 inch in diameter and fitted with gaskets. Gasket size is outside diameter 2-3/4 inch, inside diameter 2-3/16 inch. The gaskets are to be cut from a coated fabric No. ES-555-B.

c. One glass plate.

d. One mirror.

e. A tensor high intensity lamp to place between the glass and mirror to shine on the sample during the test.

Reagents: 0.9 percent saline solution.

Sample preparation: Die cut 5 samples randomly across the width of the fabric or cut 5 - 5 inch squares randomly across the width of the fabric.

Procedure:

a. Hold a finger on the vent hole, pour enough saline solution into a jar to give a 4-1/2 inch head (approximately 600 mls.)

b. Place the test sample on the mouth of the jar and screw the metal ring and the gasket in place with sufficient torque to prevent leakage.

c. Invert the jar onto the glass plate which is mounted six to eight inches above the mirror.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

d. Start the timing of the test as soon as the jar is inverted onto the glass plate and terminate it at the first sign of penetration. Penetration is evidenced when saline penetrates the fabric and the glass plate. Occasionally some condensation (a light "fog") will be observed on the surface of the glass and care must be taken that this is not confused with a testing failure.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-01-074-6667

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6530-2197

Location: CID/SURG/PACK/H15-37

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54339	14 February 1991

3. DOCUMENT TITLE
SURGICAL PACK, HEAD AND NECK, DISPOSABLE

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)
	e. DATE SUBMITTED (YYMMDD)

7. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSTE	AC 215-737-2870	AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19101-8419	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.