[METRIC]
A-A-54341
12 February 1991

COMMERCIAL ITEM DESCRIPTION

BUR, DENTAL, Excavating Extra Long Shank (Steel)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers round, dental excavating, extra long shank steel burs suitable for use with latch-type angle handpieces for opening dental pulp chambers to gain access during endodontic procedures and for oral and maxillofacial surgical procedures.

Salient characteristics:

Classification. This specification covers round excavating burs of sizes No. 2, 4, No. 6 and No. 8, as specified (See Notes).

General. Except as noted below, the burs shall be in accordance with ANSI (American National Standards Institute) ADA (American Dental Association) Specification No. 23 for Type I, class 1a burs. In case of a conflict this document shall govern.

Material. The burs shall be fabricated from steel having a minimum tungsten content of 3.00 percent, and a minimum carbon content of 1.00 percent. As an alternate, the burs shall be fabricated from steel having a minimum tungsten content of 0.90 percent, a minimum vanadium content of 0.08 percent and a minimum carbon content of 1.00 percent.

Hardness. The hardness of the bur heads shall be not less than 750 on the Vickers scale (Diamond pyramid hardness) when tested in accordance with ASTM E92. The bur necks shall be properly tempered to resist fracture or bending during use.

Style and design. The bur head shall have a round shape. The bur shall be of one piece construction consisting of a shank, neck and head. The blades shall be uniformly cut and sharp. The overall length of the bur shall be 32 mm ± 0.5mm. The shank length shall be 22 mm ± 1 mm. The minimum number of blades, head length, head diameter and neck diameter shall be in accordance with ANSI/ADA No. 23.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Each package of six burs shall be packaged in a rigid container designed to prevent damage during shipment or storage.

Marking. Marking of burs shall be optional. The use of a trademark is optional. The type of marking or placement of the marking shall not adversely affect or interfere with the handpiece. The marking shall be distinguishable to the unaided eye.

Finish. The bur head shall have a finely machined finish free of untrimmed metal. The blades shall be properly sharpened. The shank and neck shall have a smooth finish. The latch end shall be cleanly cut and finished to interface with angle handpieces.

Workmanship. The bur shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing six burs of one size as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality for raw material, supplied by the manufacturer of the steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 at inspection level II, with an AQL of 2.5 (percent defective) except for dimensions which shall be conducted at inspection level S-1, AQL of 1.0 (percent defective). The unit of product for examination purposes shall be one package.

Examination. The examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 at an inspection level of S-1 with an AQL of 1.0 (percent defective). The unit of product for test purposes shall be one bur.

Microscopic inspection. Examination for bur defects not readily perceptible to the naked eye shall be made using a microscope of at least 10 power (10X) magnification.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterioe package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

National Stock Number	Item Identification
6520-01-283-1069	Bur, Dental, Excavating, Angle Handpiece, No. 2
6520-01-283-1070	Bur, Dental, Excavating, Angle Handpiece, No. 4
6520-01-283-1071	Bur, Dental, Excavating, Angle Handpiece, No. 6
6520-01-283-1433	Bur, Dental, Excavating, Angle Handpiece, No. 8

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians.

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD - MB

Agent:

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6520-2368
Location: Word/Bur/(Steel)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide inputs. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-MSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54341	BUR, DENTAL, Excavating Extra Long Shank (Steel)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording.

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.