A-A-54344
30 April 1991

COMMERCIAL ITEM DESCRIPTION

CURETTE, ENDOMETRIAL BIOPSY (Straight, Plastic)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers an endometrial biopsy curette suitable for microscopic examination or culturing of uterine menstrual contents.

Salient Characteristics:

General. The curette shall be a single use, self-contained, sterile, disposable curette capable of inducing suction with a cylinder-piston arrangement.

Material. The curette shall consist of a clear, flexible, uniform polypropylene sheath that is at least 23.5 cm in length with a minimum outside diameter of 3.0 mm. The sheath shall have at least three markings to include the depth of insertion of the sheath into the uterine cavity plus a proximal mark for orientation of lateral opening.

Style and design. The sheath shall contain an inner piston, with a stiff piston rod, capable of traveling the full length of the sheath when pulled or pushed by an external finger-control knob. The distal end of the sheath shall be closed, rounded and smooth to preclude perforation of the uterine cavity. The style and design shall be in accordance with figure 1.

The sheath shall have a piston stop at the proximal end of sheath. The distal tip of the sheath shall have a lateral opening of 2.4 mm diameter for sample extraction. The opening shall be rounded, smooth, and without cutting edges to avoid injuries. As an alternate, another lateral opening can be added diagonally at the distal end of the curette.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Sterility. The curette shall be sterile in accordance with the USP or AAMI requirements. Sterility Assurance Level shall be 0.000001.

Identification marking and labeling. Curettes shall be permanently marked with the contractor's name or trademark. Foreign manufactured curettes shall bear the name of the country of origin.

Workmanship. Curettes shall be processed in such a manner as to be uniform and free from cracks, pits, crevices, burrs, tool marks, and dirt or polishing materials.

PRODUCT QUALITY LEVEL.

Visual and dimensional examination. Curettes shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with contractural requirements.

Unit. Package (PG). One package containing fifty individually packaged sterile curettes, as specified, constitutes one unit. The individual package shall be a suitable container capable of maintaining sterility of contents unless package is damaged or opened. Complete and detailed instructions for use shall be supplied with each unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one curette. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers'own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the products is acceptable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and /or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-261-0489	Curette, Endometrial Biopsy, Straight Plastic, 24.13 cm long, 3.0 mm Diameter, Disposable, 50s.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:


DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4945

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.



SECTION A-A

* PRECISE SIZE SHALL BE DETERMINED BY FABRICATOR TO OBTAIN PROPER FUNCTIONING WITH MATING PARTS.

FIGURE 1 CURETTE, ENDOMETRIAL, SUCTION

NSN 6515-01-261-0489

UNLESS OTHERWISE SPECIFIED
DIMENSIONS ARE IN CENTIMETERS
TOLERANCES ON:

FRACTIONS DECIMALS ANGLES

±0.1 CM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T19:32Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54344	2. DOCUMENT DATE (YYMMDD) 30 April 1991

3. DOCUMENT TITLE
CURETTE, ENDOMETRIAL BIOPSY (Straight, Plastic)

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME Defense Personnel Support Center	b. TELEPHONE (Include Area Code) (1) Commercial: (215)737-2870 (2) AUTOVON: 444-2870
c. ADDRESS (Include Zip Code) ATTN: DPSC-MSTE 2800 S. 20th Street Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 *Previous editions are obsolete* 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.