A-A-54346
28 February 1991

COMMERCIAL ITEM DESCRIPTION

TUBING, GLASS (Laboratory Use)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers standard wall borosilicate glass tubing for laboratory use.

Salient Characteristics:

Shall be standard wall glass tubing for laboratory use. The cross section of the glass tubing shall be round

Shall be fabricated from borosilicate glass conforming to Type I and Class A requirements in ASTM E-438 (Standard Specification for Glasses in Laboratory Apparatus).

Shall be furnished in 48 inch lengths with tolerance of +/- 1/2 inch. The ends of the glass tubing shall be cut and may be furnished with or without the ends glazed or fire-polished.

Shall be furnished with the following millimeter (mm) size outside diameters and wall thicknesses (see "Ordering data" paragraph).

Outside diameter (mm)	Wall thickness (mm)
2	0.5
3	0.6
4	0.8
6	1.0
8	1.0
10	1.0
11	1.0
12	1.0
13	1.2
15	1.2
18	1.2
20	1.2
22	1.5
25	1.5

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A: Approved for public release:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Commercial tolerances shall apply to the dimensions for the outside diameters and wall thicknesses

Shall be supplied with a ring strain in the glass tubing (if present) that shall not exceed 750 psi (5.2 MPa) in accordance with ASTM E-671 (Standard Specification for Maximum Permissible Thermal Residual Stress in Annealed Glass Laboratory Apparatus).

Workmanship. The glass tubing shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing the required quantity of glass tubing, as specified in the contract and/or order, constitutes one unit (see "Ordering data" paragraph).

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted on the glass tubing in accordance with commercial practice to determine compliance with specification requirements. Sampling for inspections and tests shall be conducted in accordance with commercial practice.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Other publications. The following documents form a part of this specification to the extent that is specified herein Unless otherwise indicated, the effect on date of invitation for bids or request for proposals shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM E-438 STANDARD SPECIFICATION FOR GLASSES IN LABORATORY APPARATUS

ASTM E-671 STANDARD SPECIFICATION FOR MAXIMUM PERMISSIBLE THERMAL RESIDUAL STRESS IN ANNEALED GLASS LABORATORY APPARATUS.

(Application for copies should be addressed to American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103)

NOTE. The following National Stock Numbers are covered by this document:

National Stock Number	Item Identification	Outside Diameter
6640-01-324-1382	TUBING, GLASS, 15's	8 mm
6640-01-332-1273	TUBING, GLASS, 10's	20 mm
6640-01-332-1274	TUBING, GLASS, 5's	25 mm
6640-01-333-4152	TUBING, GLASS, 20's	6 mm
6640-01-333-4153	TUBING, GLASS, 12's	10 mm

Ordering data. (Outside diameter size, unit/intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order)

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project Number: 6640-1661
Location: ENABLE/TUBINGCD/S31

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1 2 3 and 8 In block 1 both the document number and revision letter should be given

2 The submitter of this form must complete blocks 4 5 6 and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

I. RECOMMEND A CHANGE:	1 DOCUMENT NUMBER A-A-54346	2 DOCUMENT DATE (YYMMDD) 28 February 1991

3 DOCUMENT TITLE
TUBING, GLASS (Laboratory Use)

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite if possible Attach extra sheets as needed)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a NAME (Last, First, Middle Initial)	b ORGANIZATION
c ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable) 7 DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME Defense Personnel Support Center	b TELEPHONE (Include Area Code) (1) Commercial (215)737-2870 (2) AUTOVON 444-2870
c ADDRESS (Include Zip Code) ATTN: DPSC-MSTE 2800 S. 20th Street Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403 Falls Church VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 Previous editions are obsolete *90/700

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.