A-A-54347
1 March 1991

COMMERCIAL ITEM DESCRIPTION

MONITOR, FETAL HEART

(With Recorder)

The General Services Administration has authorized the use of this Commercial Item Description

This Commercial Item Description covers a fetal heart monitor with recorder, capable of both antepartum and intrapartum monitoring

Salient Characteristics:

The fetal heart monitor with recorder shall be capable of antepartum and intrapartum monitoring

The system shall be capable of internal (direct) ECG heart rate monitoring The system shall be capable of monitoring fetal heart rate externally with pulsed doppler ultrasound incorporating autocorrelation signal processing technology

The monitor shall be capable of dual ultrasonic heart rate monitoring to allow for the simultaneous noninvasive monitoring of twins.

The monitor shall measure uterine pressure both internally (by intrauterine pressure transducer) and externally (by tocodynamometer) The system shall be capable of digitally displaying fetal heart rate and uterine pressure

The monitor shall have an operator controlled event marker for identification of the onset of a particular event on the strip-chart recorder

The fetal heart monitor shall incorporate an integral, dual channel, strip chart recorder capable of providing a continuous record of fetal heart rate and uterine activity as well as the operator controlled "event" marking The strip

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A Approved for public release distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

chart recorder shall provide for operator selected speeds of one (1) and three (3) cm/min The record shall be annotated with the time, date, chart speed, and mode of operation (ultra or ECG)

The system shall have the capability to annotate the strip chart recording with operator entered alpha/numeric patient indentification data

The monitor shall have the capability for use as either a stand-alone, bedside monitor or as a component of a central monitoring system. The compatibility is intended only for central systems of the offeror's manufacturer

The monitor shall interface with noninvasive blood pressure monitors and shall record the measured systolic, diastolic, and mean blood pressures, and pulse rate on the strip chart

Heart rate indicator shall function accurately (plus or minus 5 percent) over the range of 30-210 BPM (Beats Per Minute)

The uterine activity channel shall display zero to 100 millimeters of mercury of uterine pressure

The vertical scale on the monitor paper shall be one centimeter equals 30 beats per minute heart rate, and one centimeter equals 25 millimeters of mercury of pressure.

The system shall either have a paper collection tray or shall be provided with a standard monitor cart that otherwise provides for proper collection of the monitor paper

The monitor shall have the capability of recording and displaying both doppler and direct ECG signals simultaneously

The unit shall be of a rugged construction capable of withstanding hazards of normal use in a hospital setting

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

The fetal heart monitor shall be supplied with the following transducers.

a Ultrasonic transducer

b Tocodynamometer transducer

c Intrauterine pressure transducer

Electrical charactristics

a The system shall operate over the full tolerance range of one or more of the following line voltages and frequency combinations

* 115 VAC plus or minus 10 percent, 50 plus or minus 2 Hz
* 115 VAC plus or minus 10 percent, 60 plus or minus 2 Hz
* 230 VAC plus or minus 10 percent, 50 plus or minus 2 Hz
* 230 VAC plus or minus 10 percent, 60 plus or minus 2 Hz

Dual voltage and frequency operation is desirable but is not mandatory

b. Change over devices, switches, circuits, or transformers if used for the purpose of voltage or frequency conversion, shall be clearly marked and identified in the appropriate documentation which accompanies the unit. Mechanical or electrical adjustment to effect the conversion shall not require the use of specialized tools or skills

Operational and service data manuals

a Information content·

The contractor shall furnish manuals, handbooks and/or brochures containing complete operation, installation and service/maintenance instructions (including pictures of illustrations, as necessary) with complete schematics and wiring diagrams All manuals, handbooks, and/or brochures shall be written in the English language and all schematics and wiring diagrams shall use American electrical and electronics symbols The manuals shall include electrical data and connection diagrams for all applicable utilities The instructions shall also contain a complete list of all replaceable parts showing part number, name and quantity

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

required All service notes, service memos, etc , pertaining to the equipment and issued subsequent to the printing of the original instructions shall also be included These instructions shall be as a minimum, that furnished to service engineers (or servicemen) that normally install and service the equipment for the company or distributor When the system being procured includes, or will operate ancillary equipment, the service data shall include complete instructions and drawings which show interfacing of all system components

b Quantities, distribution and schedules

Two (2) complete sets of the above information shall be packaged with each unit of the required equipment. Where multiple quantities of the same unit are to be delivered to the same ultimate consignee, then only a total of 2 service manuals need to be supplied to each consignee However, 2 operation manuals still need to be supplied with each unit

Workmanship The fetal heart monitor shall be free from defects which detract from its appearance or impair it serviceability

Unit Each (EA) One fetal heart monitor, as specified, constitutes one unit.

Quality Assurance Provisions

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

Sampling for examination Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1 0 (percent defective) and an inspection level of II The unit of product for examination purposes shall be one fetal heart monitor

Sampling for tests Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.5 (percent defective) and an inspection level of S-2. The unit of product for test purposes shall be one fetal heart monitor

Tests Tests shall be conducted to determine compliance with specification requirements Where feasible, the same sample shall be used for the determination of two or more test characteristics Tests shall include, but shall not be limited to, the following:

1 Apply a minimum of four simulated ECG waveforms, varying the heart rate in the range 30 to 210 BPM. Displayed heart rate must be within 5 percent of the applied rate.

2 Verify pressure display has a range of 0 to 100 mmHg by applying a minimum of 3 static pressures in this range

3 Operate the strip chart recorder while applying simulated signals to verify the following:

a Operator selectable speeds of one and three cm/min.

b Automatic annotation of time, date, chart speed and mode of operation

c Operability of event marking system

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Metric products Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable

The contracting officer has the option of accepting or rejecting the product

Contractor certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Regulatory requirements

Federal Food, Drug and Cosmetic Act If the product covered by this document has been determined by the U.S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23 4 to the maximum extent practical

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order

Ordering data Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order

NOTE: The following National Stock Number is covered by this document.

6515-01-305-2706

MILITARY INTERESTS

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL ACENCY COORDINATING ACTIVITIES.

VA-OSS
USPHS
FDA-MPQAS

Project Number 6515-5008

Location CID/MONFECID 16D

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-54347	2. DOCUMENT DATE (YYMMDD): 1 March 1991

3. DOCUMENT TITLE: MONITOR, FETAL HEART (With Recorder)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME: DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSTE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code): 2800 SO. 20TH STREET PHILADELPHIA, PA 19101-8419	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403 Falls Church VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:32Z
Check the source to verify that this is the current version before use.