A-A-53546
28 June 1988

COMMERCIAL ITEM DESCRIPTION

EVACUATOR, SURGICAL WOUND

The General Services Administration has authorized the use of this commercial item description as a replacement for Military Specification MIL-E-37718 (DLA-DM), which is cancelled.

This commercial item description covers a sterile, disposable evacuator for surgical wounds.

Salient characteristics:

General. The evacuator shall be a sterile, disposable, plastic wound evacuator, compact size, and suitable for providing suction necessary to evacuate wounds before, during, and after surgical procedures. The evacuator shall be supplied with the following components:

1 Evacuator vessel
1 Perforated tube
1 Non-perforated tube
1 Y-connector
1 Needle

Material.

Evacuator vessel. Evacuator vessel shall be fabricated from opaque, flexible and transparent polyvinyl chloride (PVC) plastic which shall not cause trauma or irritate body tissue. Lead content shall not exceed 5 ppm. The springs shall be made of high carbon spring steel and be plated or coated with cadmium for corrosion resistance. The securing clip shall be made of rubber and rigid plastic.

Plastic tubing. All plastic tubing (perforated, non-perforated, Y-connector) shall be made of transparent, non-pyrogenic, flexible PVC plastic. The perforated tubing shall have a radiopaque line along its length and comply with current USP Class IV plastics requirements. Lead content shall not exceed 5 ppm for all plastic tubing.

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Needle. The needle shall be made of austenitic stainless steel and have a mirror finish.

Style, design and dimensions. The evacuator shall be a lightweight, compact size, cylindrically shaped, spring type vacuum evacuator having a 400 ml volume capacity. The top cover shall be 4-1/2 ± 1/2 inches in diameter, opaque and rigid and have a pre-positioned anti-reflux valve, pouring spout and pouring spout plug. The pouring spout shall have a minimum of 0.435 inch diameter for emptying fluid efficiently. The bottom of the evacuator shall also be 4-1/2 ± 1/2 inches in diameter and be opaque and rigid. The transparent side section shall have volume graduations from 0 to 400 ml. The bottom cover shall have imprinted instructions for use and a securing clip attached to it.

The perforated tube shall be 43 ± 2 inches long with a 12 ± 1 inch perforated section at the center of the tube. The inner and outer diameters shall be 1/16 ± 1/64 inches and 1/8 ± 1/64 inches respectively.

The unperforated tube shall be 31 ± 2 inches long. The inner and outer diameters shall be 1/8 ± 1/32 inches and 1/4 ± 1/32 inches respectively.

The Y-connector shall be designed to allocate one end for connection with the non-perforated tube and two ends for the perforated tube.

The tubes shall have the following minimum force security of attachments requirements when tested as specified herein:

Unperforated tube to vessel anti-reflux valve	7.0 lb.
Unperforated tube to Y-connector	7.0 lb.
Perforated tube to Y-connector	4.5 lb.
Perforated tube to needle	10.0 lb.

The needle shall be 6 ± 1/8 inches long with a 10° ± 2° bend 1-1/2 ± 1/8 inches from the tip of its trocar point. The connection between the needle and the perforated tube shall be smooth.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Performance. The evacuator vessel shall provide an initial suction of 50 mm Hg when tested as specified herein. Volume graduations of 200 ml and 400 ml shall be tested for volume accuracy and shall have no more than 10% error of the total volume. The evacuator shall be able to aspirate no less than 80% of its total volume. The evacuator shall not leak when tested as specified herein. The tubes shall demonstrate breaking strength and ultimate elongation of 8 lbs and 75% respectively when tested as specified herein.

Sterility. The evacuator and components shall be sterile in accordance with the USP for Medical Devices.

Workmanship. The evacuator shall be processed in such a manner as to be uniform and free from defects which affect life, performance utility, or appearance.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for assuring that materials are in accordance with requirements specified.

Unit. Each (EA). Each evacuator with components and operating instructions, as specified, constitutes one unit. Each evacuator shall be packaged in a sealed, peel open type container capable of maintaining sterility of contents unless package is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials, may be furnished in lieu of actual performance of composition testing by the contractor, provided lot identity had been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer/supplier, the NSN, the name of the component/material, the lot number, the lot size, the test method, and the test results.

Examination. The evacuator shall be examined to determine compliance with all requirements contained in this specification.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and inspection level of I. The unit of product for examination purposes shall be one complete wound evacuator. Defects shall be those listed below.

Defects

Perforated tube not smooth and not free from burrs.
Perforated tube-needle junction not smooth, i.e., lacking uniformity of diameters at junction.
Lack of concentricity of perforated tube and needle at junction.
Instructions missing or illegible.
Trocar point not sharp.
Printing or components flaking off.
Presence of foreign matter in or on perforated tube.

Sampling for tests. Sampling for tests shall be in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one complete wound evacuator. AQL shall be 1.0 (percent defective) with an inspection level of S-2.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Aspiration volume. The evacuator, non-perforated tube, Y-connector and perforated tube shall all be connected. A large pan of water shall be present with the water level equal to within one-half inch to that of the evacuator vessel in the fully primed state, i.e., the bottom of the evacuator vessel. Release the evacuator vessel from the fully primed state.

Volume accuracy. The volume shall be determined at the 100 cc and 400 cc levels by filling the evacuator to these levels and then emptying the evacuator contents into a graduate.

Initial suction. The evacuator shall be fully connected to a manometer, be fully primed by completely compressing, plugging the pouring spout and releasing plugged and the handle released. The negative pressure shall be read from the manometer.

Leakage of evacuator. The non-perforated tube shall be attached to the evacuator. The drain port shall be plugged. The end of the tube shall be connected to a vacuum pump and air aspirated to a vacuum of 160 mm Hg. The evacuator at the negative pressure of 160 mm Hg shall then be immersed in water for 10 seconds and observed for any sign of leakage.

Breaking strength and ultimate elongation. The breaking strength and ultimate elongation of the tubes shall be performed in accordance with ASTM Standard D-412 using straight specimens. Breaking strength is the observed force as defined in the standard.

Security of attachments. Weights as specified shall be applied gradually to test each of the attachments. The load shall be maintained for a minimum of ten seconds.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-149-0671	EVACUATOR, SURGICAL WOUND, Disposable, 400 ml

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-ME

Agent:

DLA-DH

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC-46648

Project No. 6515-4219

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
[illegible]-335-0	EVACUTOR, SURGICAL WOUND

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION (Mark one)
	☐ VENDOR
	☐ USER
b. ADDRESS (Street, City, State, ZIP Code)	☐ MANUFACTURER
	☐ OTHER (Specify)

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER (Last, First, MI) – Optional	b. WORK TELEPHONE NUMBER (Include Area Code) – Optional
c. MAILING ADDRESS (Street, City, State, ZIP Code) – Optional	8. DATE OF SUBMISSION (YYMMDD)

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better the DoD provides this form for use in submitting comments and suggestions for improvements All users of military standardization documents are invited to provide suggestions This form may be detached folded along the lines indicated taped along the loose edge (DO NOT STAPLE, and mailed In block 5 be as specific as possible about particular problem areas such as wording which required interpretation was too rigid restrictive loose ambiguous or was incompatible and give proposed wording changes which would alleviate the problems Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents nor to request waivers, deviations, or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
PHILADELPHIA,PENNSYLVANIA 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.