COMMERCIAL ITEM DESCRIPTION

# GLOVE, PATIENT EXAMINING AND TREATMENT (Plastic, Large Size, Sterile)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a sterile, disposable, impermeable, light-weight seamless plastic glove of one size which fits either hand. Glove shall be suitable for medical use in examinations, diagnostic and therapeutic procedures by physicians, nurses and paramedical personnel in clinic and nursing-unit care.

Salient characteristics:

Material. The glove shall be fabricated from virgin, translucent polyvinyl chloride.

Dusting powder. Should dusting powder be used, the gloves shall be uniformly prepowdered on the interior only with a minimum of Absorbable Dusting Powder, USP.

Style and design. The patient examining glove shall be constructed in one piece, free of seams and welts, and shall have a smooth finish. Shall be a five-fingered glove suitable for use on either hand. A small continuous bead shall be formed on the glove circumference at the cuff end to prevent tearing. The glove shall be supplied in a folded inner wrapper which in turn shall be enclosed in an outer wrapper as specified below. The glove shall be placed flat, prior to folding the inner wrapper over glove, with the cuff folded over for 1-1/2 to 3-1/2 inches. The inner wrapper shall be of sufficient size to permit such flat placement of the glove and shall completely cover and separate glove from outer wrapper. The inner wrapper shall be folded in such a manner as to permit easy removal of the glove. Inner wrapper shall have the "CUFF END" and "THUMB" clearly marked on it. The glove shall not slide in the package or adhere to the inner wrapper while in storage.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Dimensions. The glove shall have the dimensions (in inches) as indicated in Table 1.

Sterility. The glove shall be sterile and free from pyrogens in accordance with the USP. The glove shall have a minimum sterility assurance level of $1 \times 10^{-4}$.

Freedom from pinholes. The glove shall be supplied free of pinholes. The glove shall not leak when filled with 400 ml of water and suspended from the cuff end for 5 minutes.

Finish. The glove shall have a visually dull but smooth-to-the-touch exterior, and visually glossy and smooth-to-the-touch interior surface. The glove shall be free from holes, visible ingrained particles, tackiness, bloom, strings, bleeding, and other imperfections.

Low temperature flexibility. The glove shall not crack at -20°F after being held at that temperature for 1/2 hour and then flexed with a wooden rod.

High temperature exposure. The glove shall withstand exposure in a mechanical convection oven at 165°F for 72 hours without evidence of tackiness, plasticizer exudation, or other deterioration. It shall be examined after it is at a nominal 75°F.

Instructions. Each envelope shall be labeled with instructions for withdrawing and donning the glove without compromising its sterility.

Workmanship. Workmanship shall be first class throughout. Gloves shall be free from defects which detract from their appearance or may impair their serviceability.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

TABLE I.

| Characteristic | Dimensions |
|---|---|
| Overall length | 10 + 1, -1/4 |
| Flat width at wrist | 4-1/4 + 1/2, -1/4 |
| Flat width at palm at base of thumb | 4-1/4 ± 1/4 |
| Length of thumb | 2-3/8 ± 1/4 |
| Length of little finger | 2-1/4 ± 1/8 |
| Length of middle finger | 3 ± 1/4 |
| Length of index and ring fingers | 2-7/8 ± 3/8 |
| Flat width of thumb at base | 1-1/2, ± 1/8 |
| Flat width of thumb 3/4 inch from tip | 1-1/4, ± 1/4, -3/16 |
| Flat width of base of index, middle and ring fingers | 1-1/4 ± 1/8 |
| Same as line above, 3/4 inch from tips | 1-1/8 ± 1/8 |
| Flat width of base of little finger | 1-1/8 ± 1/8 |
| Same as line above, 3/4 inch from tip | 1-1/8 ± 1/8 |
| Thickness at finger tips | 0.0025 to 0.0045 |
| Thickness at remainder of glove (except for area adjacent to cuff edge) | 0.005 to 0.008 |

Outer wrapper (envelope). The outer wrapper (envelope) shall completely enclose the inner wrapper and glove, and shall be capable of maintaining sterility of the glove. Outer wrapper shall be so designed as to permit ready dispensing of the contents by sterile techniques. Closure of the outer wrapper shall be accomplished by heat or cold sealing. The outer wrapper shall be a sealed, peel-back outer wrapper. The peeled back design shall include two free tabs along the top edge of the wrapper. The tabs shall be capable of being readily peeled back along the sealed edges to expose the inner wrapper.

Physical requirements. The gloves shall comply with the physical requirements contained in Table II. Tests for physical requirements shall be conducted in accordance with ASTM D882.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

TABLE II. Physical Requirements.

| | |
|---|---|
| Tensile strength - pounds per square inch (average)* | 1500 min. |
| Tensile strength - pounds per square inch (single glove) | 1300 min. |
| Breaking strength - pounds per inch of width | 5 min. |
| Percentage elongation at break | 280 min. |

* The tensile strength shall be run on a minimum of 30 gloves and represents an average value.

Unit. Package (PG). One package containing 100 individually packaged sterile gloves, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be in accordance with MIL-STD-105 with and AQL of 1.0 (percent defective) and an Inspection Level of S-3. The unit of product for examination purposes shall be one glove.

Sampling for testing. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one glove. AQL shall be 2.5 with an inspection level of S-3.

Defects

Tears, large holes.
Blisters.
Imbedded grit-like particles.
Ingrained grit-like particles clearly noticeable without distending the glove.
Excessively thin spots (under 0.0025 inch).
Tackiness.
No inner wrapper included.
CUFF END not identified on inner wrapper.
Glove adhering to inner wrapper.
Gloves not having a small continuous bead located upon the cuff end.
Gloves powdered on exterior.
Foreign matter present.
Bloom, bleeding, strings, and other imperfections.
Fingertips thickness over 0.0045 inch.
Palm thickness over 0.008 inch.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

| NSN | ITEM IDENTIFICATION |
|---|---|
| 6515-00-226-7692 | GLOVE, PATIENT EXAMINING AND TREATMENT, Plastic, Large Size, Disposable, Sterile, 100's |

Ordering data (Intermediate exterior packaging quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4215



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

*(See Instructions – Reverse Side)*

| 1 DOCUMENT NUMBER | 2 DOCUMENT TITLE |
|---|---|
| A-A-53562 | GLOVE, PATIENT EXAMINING AND TREATMENT (Plastic, Large Size, Sterile) |

| 3a NAME OF SUBMITTING ORGANIZATION | 4 TYPE OF ORGANIZATION *(Mark one)* |
|---|---|
| | ☐ VENDOR |
| | ☐ USER |
| b ADDRESS *(Street, City, State, ZIP Code)* | ☐ MANUFACTURER |
| | ☐ OTHER *(Specify)* |

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

| 7a NAME OF SUBMITTER *(Last, First, MI)* – Optional | b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional |
|---|---|
| c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional | 8 DATE OF SUBMISSION *(YYMMDD)* |

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge (*DO NOT STAPLE*), mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

*NOTE* This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

*(Fold along this line)*

*(Fold along this line)*

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

**BUSINESS REPLY MAIL**
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN. DPSCRSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.