A-A-53582
29 July 1988

COMMERCIAL ITEM DESCRIPTION

MOUNT, RADIOGRAPHIC FILM, DENTAL

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers an open-window cardboard or plastic mount, suitable for use in filing dental x-ray films.

Salient characteristics:

The dental radiographic film mount shall be one of the following two types:

Type I - Capable of holding ten Posterior Number 3 Bite-Wing dental films - nominal film size 1-1/16 inch by 2-1/8 inch (2.7 cm by 5.4 cm).

Type II - Capable of holding eight number 2 standard Periapical dental films - nominal film size 1-1/4 inch by 1-5/8 inch (3.18 cm by 2.86 cm).

The mount shall be fabricated of either three-layer paper stock or two-layer paper stock glued together with a strong, durable, adhesive.

Paper stock used in the fabrication of the three - layer dental mounts shall have the following minimum values:

1. Thickness of 0.009 inch (.23 mm).
2. Basis weight (1000 sheets 22-1/2 by 28-1/2 inch (57.15 cm by 72.39 cm)) of 158 pounds.
3. Bursting strength of 60 pounds per square inch.

Paper stock used in the fabrication of the two-layer dental mounts shall have the following minimum values:

1. Thickness of 0.0135 inch (.34 mm).
2. Basis weight of 225 pounds.
3. Bursting strength of 100 pounds per square inch.

AMSC N/A FSC 6525

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

As an alternate, the mount may be of equivalent plastic construction.

The type I mount shall have nominal overall dimensions of 5 inches by 9-1/4 inches. The ten windows shall be aligned vertically and horizontally and evenly spaced in pairs with the long axis of the window parallel to the short axis of the mount. The mount shall permit insertion of film in the back panel.

The type II mount shall have nominal overall dimensions of 4-3/8 inches by 8-1/4 inches. The eight windows shall be aligned vertically and horizontally and evenly spaced in two rows of four windows each, with the long axis of the window parallel to the long axis of the mount. The mount shall permit insertion of film in the back panel.

The windows for the type I and type II mounts shall have rounded corners and shall be cut, free of ragged edges with no marginal separations. Windows shall be so constructed, in the back panel, to permit easy insertion and secure holding of the film.

The top margin of the mount front shall be imprinted with two blank lines and headings to permit insertion of the patient's name, identification number and date. In addition, date heading and blank lines shall be uniformly imprinted beneath each pair of windows across the width of the mount.

The mount shall be free of oil, grease, dirt, or other contamination.

Workmanship. The dental radiographic film mount shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One Package containing twenty-five mounts, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one dental radiographic film mount as specified in this description.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Defects.

Material not as specified.

Placement of windows not as specified.

Separation of paper layers because of insufficient glue.

Identification marking missing, illegible, or incomplete.

Windows do not have rounded corners and smooth edges.

Mount not free of tears or rough edges.

Mount not free of oil, grease, dirt, or other contamination.

Sampling for tests. Sampling for tests including dimensional test shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-3. The unit of product for test purposes shall be one dental radiographic film mount.

Tests

Film Insertion test. Films, of the sizes specified for type I and type II mounts, shall be tried for fit in the film mounts. For type I mounts, not less than four windows shall have the sample films inserted. For type II mounts, not less than three windows shall have the sample films inserted. Any film difficult to insert, or not securely held by mount, shall constitute a defect.

Thickness test. The paper to be tested shall be a single sheet of paper not less than 2 inches by 2 inches (51 mm by 51 mm) and free from wrinkles and creases.

After calibration of a suitable micrometer by means of an accepted method, the paper specimen shall be clamped between the faces of the micrometer under a steady pressure of 7.0 to 9.0 pounds per square inch (0.40 to 0.63 kg. per square cm). A thickness less than 0.009 inch (three-layer mount) or 0.0135 inch (two-layer mount) shall constitute a defect.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Basis weight test. The paper specimen to be tested shall consist of 10 sheets of 22-1/2 inches by 28-1/2 inches (57.15 cm by 72.39 cm) paper.

The weight of the 10 paper sheet specimen shall be determined using a suitable method. The weight of 1000 sheets shall then be calculated. A 1000 sheet weight less than 158 pounds (three-layer mount) or 225 pounds (two-layer mount) shall constitute a defect.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6525-00-142-8732 (Type I)
6525-01-030-9011 (Type II

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6525-1241

PGC 05108

Location: CID/MOUNT/DENTAL/H13-32

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge (DO NOT STAPLE), and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 1966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53582	MOUNT, RADIOGRAPHIC FILM, DENTAL

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER (Last, First, MI) – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2020-10-20T00:21Z
Check the source to verify that this is the current version before use.